

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL040825\
 Data File : VL042264.D
 Acq On : 08 Apr 2025 15:25
 Operator : SY/MD
 Sample : Q1728-11 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SSSG-B1-4-04012025

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 04/09/2025
 Supervised By :Semsettin Yesilyurt 04/10/2025

Quant Time: Apr 09 01:56:36 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_LFri Aug 2
 QLast Update : Fri Mar 28 02:00:52 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.803	49	167832	10.000	ppbv	0.02
33) 1,4-Difluorobenzene	3.981	114	455581	10.000	ppbv	0.03
55) Chlorobenzene-d5	8.901	117	396189	10.000	ppbv	0.03
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.393	95	304927	9.372	ppbv	0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	93.700%
Target Compounds						
					Qvalue	
13) Ethanol	1.729	45	2625m	2.393	ppbv	
15) Acetone	1.845	43	26507	1.129	ppbv	95
19) Isopropyl Alcohol	1.897	45	4088m	0.309	ppbv	
20) Methylene Chloride	2.075	84	1419	0.162	ppbv #	88
53) Toluene	6.959	91	8616	0.161	ppbv	96
59) m/p-Xylene	9.555	91	7675m	0.137	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL040825\
Data File : VL042264.D
Acq On : 08 Apr 2025 15:25
Operator : SY/MD
Sample : Q1728-11 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SSSG-B1-4-04012025

Quant Time: Apr 09 01:56:36 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Fri Mar 28 02:00:52 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 04/09/2025
Supervised By :Semsettin Yesilyurt 04/10/2025

