

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL040825\
 Data File : VL042271.D
 Acq On : 08 Apr 2025 19:03
 Operator : SY/MD
 Sample : Q1728-10 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SSSG-B14-4-04012025

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 04/09/2025
 Supervised By :Semsettin Yesilyurt 04/10/2025

Quant Time: Apr 09 01:59:38 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Fri Mar 28 02:00:52 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.793	49	156461	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.968	114	398650	10.000	ppbv	0.02
55) Chlorobenzene-d5	8.891	117	347563	10.000	ppbv	0.02
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.383	95	280057	9.812	ppbv	0.01
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.100%
Target Compounds						
					Qvalue	
9) Propene	1.489	41	1392	0.129	ppbv	89
15) Acetone	1.839	43	67031	3.063	ppbv	94
19) Isopropyl Alcohol	1.890	45	61799	5.005	ppbv #	34
20) Methylene Chloride	2.065	84	7654	0.939	ppbv	90
26) Hexane	2.832	57	3697	0.169	ppbv #	34
59) m/p-Xylene	9.545	91	8194m	0.166	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL040825\
Data File : VL042271.D
Acq On : 08 Apr 2025 19:03
Operator : SY/MD
Sample : Q1728-10 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SSSG-B14-4-04012025

Quant Time: Apr 09 01:59:38 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Fri Mar 28 02:00:52 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 04/09/2025
Supervised By :Semsettin Yesilyurt 04/10/2025

