

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL041019\
 Data File : VL033564.D
 Acq On : 10 Apr 2019 14:35
 Operator : SY/AP
 Sample : VL0410ABL01
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0410ABL01

Manual Integrations
 APPROVED

MMDadoda
 4/11/2019 1:01:00 PM

Quant Time: Apr 11 03:53:16 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL032519AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Mar 26 2019
 QLast Update : Tue Mar 26 16:00:50 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.76	49	1013126	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.29	114	2502132	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.24	117	2584459	10.00	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.57	95	2103392	10.38	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	103.80%
Target Compounds						
15) Acetone	3.91	43	47639	0.391	ppbv	91
20) Methylene Chloride	4.41	84	19275m	0.373	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL041019\
Data File : VL033564.D
Acq On : 10 Apr 2019 14:35
Operator : SY/AP
Sample : VL0410ABL01
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0410ABL01

Manual Integrations
APPROVED

MMDadoda
4/11/2019 1:01:00 PM

Quant Time: Apr 11 03:53:16 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL032519AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LTue Mar
QLast Update : Tue Mar 26 16:00:50 2019
Response via : Initial Calibration

