

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL041019\
Data File : VL033566.D
Acq On : 10 Apr 2019 16:17
Operator : SY/AP
Sample : QC040319 CAN 10284
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC040319 CAN 10284

Quant Time: Apr 11 04:00:00 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL032519AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Mar 26 16:00:50 2019
QLast Update : Tue Mar 26 16:00:50 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.76	49	931891	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.29	114	2324996	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.24	117	2262843	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.58	95	1773996	10.00	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.00%

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL041019\
Data File : VL033566.D
Acq On : 10 Apr 2019 16:17
Operator : SY/AP
Sample : OC040319 CAN 10284
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC040319 CAN 10284

Quant Time: Apr 11 04:00:00 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL032519AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Mar 26 16:00:50 2019
QLast Update : Tue Mar 26 16:00:50 2019
Response via : Initial Calibration

