

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL041019\
 Data File : VL033574.D
 Acq On : 10 Apr 2019 21:37
 Operator : SY/AP
 Sample : K2301-03DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 IA1DL

Manual Integrations
 APPROVED

MMDadoda
 4/11/2019 1:01:36 PM

Quant Time: Apr 11 04:51:09 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL032519AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Mar 2
 QLast Update : Tue Mar 26 16:00:50 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.76	49	901049	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.29	114	2131117	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.24	117	2155612	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.57	95	1698646	10.05	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.50%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
3) Chlorodifluoromethane	3.02	51	10375m	0.100	ppbv	
9) Propene	3.05	41	10194	0.234	ppbv	95
13) Ethanol	3.67	45	36340m	3.224	ppbv	
15) Acetone	3.91	43	59865	0.553	ppbv	95
19) Isopropyl Alcohol	4.04	45	8889m	0.130	ppbv	
20) Methylene Chloride	4.41	84	21460m	0.467	ppbv	
52) Tetrachloroethene	11.37	164	10325m	0.153	ppbv	
53) Toluene	9.95	91	19409m	0.093	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL041019\
Data File : VL033574.D
Acq On : 10 Apr 2019 21:37
Operator : SY/AP
Sample : K2301-03DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
IA1DL

Manual Integrations
APPROVED

MMDadoda
4/11/2019 1:01:36 PM

Quant Time: Apr 11 04:51:09 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL032519AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LTue Mar
QLast Update : Tue Mar 26 16:00:50 2019
Response via : Initial Calibration

