

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL041019\
 Data File : VL033576.D
 Acq On : 10 Apr 2019 22:56
 Operator : SY/AP
 Sample : K2312-01DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 A-6DL

Manual Integrations
 APPROVED

MMDadoda
 4/11/2019 1:01:49 PM

Quant Time: Apr 11 04:58:50 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL032519AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Mar 26 2019
 QLast Update : Tue Mar 26 16:00:50 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.76	49	876330	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.29	114	2022161	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.24	117	2135284	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.57	95	1698671	10.15	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.50%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
10) Heptane	8.25	43	43626	0.385	ppbv	98
13) Ethanol	3.65	45	130005	11.858	ppbv	64
15) Acetone	3.91	43	63033	0.598	ppbv	93
19) Isopropyl Alcohol	4.04	45	63980	0.960	ppbv #	92
20) Methylene Chloride	4.41	84	37871m	0.847	ppbv	
26) Hexane	5.78	57	31010	0.357	ppbv #	45

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL041019\
Data File : VL033576.D
Acq On : 10 Apr 2019 22:56
Operator : SY/AP
Sample : K2312-01DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
A-6DL

Manual Integrations
APPROVED

MMDadoda
4/11/2019 1:01:49 PM

Quant Time: Apr 11 04:58:50 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL032519AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Mar
QLast Update : Tue Mar 26 16:00:50 2019
Response via : Initial Calibration

