

Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL041019\
 Data File : VL033563.D
 Acq On : 10 Apr 2019 11:06
 Operator : SY/AP
 Sample : VSTDCCC010
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Apr 11 03:50:56 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL032519AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LTue Mar 26 16:00:50 2019
 QLast Update : Tue Mar 26 16:00:50 2019
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	125	0.00
2 T	Dichlorodifluoromethane	1.458	1.513	-3.8	155#	0.00
3	Chlorodifluoromethane	1.151	0.929	19.3	109	0.00
4	Chloromethane	0.609	0.516	15.3	112	0.00
5 T	Vinyl Chloride	0.599	0.508	15.2	111	0.00
6 T	Bromomethane	0.385	0.332	13.8	114	0.00
7	Chloroethane	0.223	0.190	14.8	111	0.00
8 T	Dichlorotetrafluoroethane	1.554	1.271	18.2	112	0.00
9 T	Propene	0.483	0.436	9.7	117	0.00
10 T	Heptane	1.292	1.113	13.9	113	0.00
11 T	Trichlorofluoromethane	1.797	1.394	22.4	105	0.00
12 T	1,1,2-Trichlorotrifluoroeth	1.207	0.974	19.3	109	0.00
13	Ethanol	0.125	0.115	8.0	114	0.00
14 T	Bromoethene	0.482	0.425	11.8	113	0.00
15 T	Acetone	1.202	0.906	24.6	110	0.00
16 T	1,3-Butadiene	0.477	0.408	14.5	98	0.00
17	tert-Butyl alcohol	0.221	0.183	17.2	107	0.00
18 T	1,1-Dichloroethene	0.519	0.450	13.3	114	0.00
19 T	Isopropyl Alcohol	0.761	0.669	12.1	115	0.00
20 T	Methylene Chloride	0.510	0.372	27.1	110	0.00
21 T	Allyl Chloride	0.760	0.662	12.9	109	0.00
22 T	trans-1,2-Dichloroethene	0.541	0.456	15.7	112	0.00
23 T	Vinyl Acetate	1.711	1.565	8.5	113	0.00
24 T	1,1-Dichloroethane	1.300	1.088	16.3	111	0.00
25 T	Ethyl Acetate	2.231	1.838	17.6	109	0.00
26 T	Hexane	0.992	0.840	15.3	114	0.00
27 T	Carbon Disulfide	1.180	1.081	8.4	109	0.00
28 T	Methyl tert-Butyl Ether	1.429	1.346	5.8	116	0.00
29 T	Chloroform	1.700	1.404	17.4	111	0.00
30 T	Cyclohexane	0.785	0.752	4.2	120	0.00
31 T	cis-1,2-Dichloroethene	0.986	0.874	11.4	111	0.00
32 T	1,1,1-Trichloroethane	1.619	1.410	12.9	110	0.00
33 I	1,4-Difluorobenzene	1.000	1.000	0.0	139	0.00
34 T	2-Butanone	0.601	0.471	21.6	112	0.00
35 T	Carbon Tetrachloride	0.708	0.560	20.9	108	0.00
36 T	Benzene	0.839	0.704	16.1	119	0.00
37 T	1,2-Dichloroethane	0.562	0.395	29.7	103	0.00
38 T	Trichloroethene	0.316	0.283	10.4	119	0.00
39 T	1,2-Dichloropropane	0.329	0.263	20.1	115	0.00
40 T	1,4-Dioxane	0.117	0.104	11.1	125	0.00
41 T	Tetrahydrofuran	0.330	0.272	17.6	113	0.00
42 T	Bromodichloromethane	0.766	0.598	21.9	110	0.00
43	Methyl Methacrylate	0.341	0.289	15.2	116	0.00
44 T	2,2,4-Trimethylpentane	1.466	1.144	22.0	113	0.00
45 T	t-1,3-Dichloropropene	0.418	0.392	6.2	112	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL041019\
 Data File : VL033563.D
 Acq On : 10 Apr 2019 11:06
 Operator : SY/AP
 Sample : VSTDCCC010
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Apr 11 03:50:56 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL032519AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LTue Mar 26 16:00:50 2019
 QLast Update : Tue Mar 26 16:00:50 2019
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	cis-1,3-Dichloropropene	0.488	0.436	10.7	113	0.00
47 T	1,1,2-Trichloroethane	0.373	0.290	22.3	113	0.00
48 T	Dibromochloromethane	0.630	0.522	17.1	110	0.00
49 T	Bromoform	0.601	0.479	20.3	83	0.00
50 T	4-Methyl-2-Pentanone	0.880	0.673	23.5	107	0.00
51 T	2-Hexanone	0.656	0.549	16.3	110	0.00
52 T	Tetrachloroethene	0.317	0.273	13.9	115	0.00
53 T	Toluene	0.975	0.846	13.2	116	0.00
54 T	1,2-Dibromoethane	0.584	0.473	19.0	113	0.00
55 I	Chlorobenzene-d5	1.000	1.000	0.0	132	0.00
56	1,1,1,2-Tetrachloroethane	0.458	0.357	22.1	110	0.00
57 T	Chlorobenzene	0.776	0.606	21.9	112	0.00
58 T	Ethyl Benzene	1.335	1.115	16.5	111	0.00
59 T	m/p-Xylene	1.193	0.930	22.0	98	0.00
60 T	o-Xylene	1.153	0.908	21.2	108	0.00
61 T	Styrene	0.792	0.706	10.9	112	0.00
62	Isopropylbenzene	1.633	1.304	20.1	109	0.00
63 T	1,1,2,2-Tetrachloroethane	0.849	0.627	26.1	107	0.00
64	n-propylbenzene	0.414	0.340	17.9	109	0.00
65	tert-Butylbenzene	1.439	1.173	18.5	111	0.00
66 T	Benzyl Chloride	0.669	0.678	-1.3	112	0.00
67	sec-Butylbenzene	2.047	1.635	20.1	109	0.00
68 S	1-Bromo-4-Fluorobenzene	0.784	0.743	5.2	127	0.00
69	p-Isopropyltoluene	1.667	1.362	18.3	108	0.00
70	n-Butylbenzene	1.766	1.420	19.6	107	0.00
71	2-Chlorotoluene	1.242	1.000	19.5	106	0.00
72 T	4-Ethyltoluene	1.297	1.097	15.4	111	0.00
73 T	1,3,5-Trimethylbenzene	1.241	1.043	16.0	112	0.00
74 T	1,2,4-Trimethylbenzene	1.330	1.066	19.8	110	0.00
75 T	1,3-Dichlorobenzene	0.822	0.638	22.4	111	0.00
76 T	1,4-Dichlorobenzene	0.801	0.643	19.7	111	0.00
77 T	1,2-Dichlorobenzene	0.780	0.624	20.0	109	0.00
78 T	Hexachloro-1,3-Butadiene	0.516	0.481	6.8	153#	0.00
79 T	Naphthalene	1.138	1.105	2.9	113	0.00
80 T	Naphthalene,2-methyl-	0.382	0.469	-22.8	109	0.00
81 T	1,2,4-Trichlorobenzene	0.631	0.564	10.6	113	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0