

Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL041019\
 Data File : VL033585.D
 Acq On : 11 Apr 2019 4:54
 Operator : SY/AP
 Sample : VSTDCCC010
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Apr 11 10:06:12 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL032519AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Mar 26 16:00:50 2019
 QLast Update : Tue Mar 26 16:00:50 2019
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	106	0.00
2 T	Dichlorodifluoromethane	1.458	1.160	20.4	101	0.00
3	Chlorodifluoromethane	1.151	0.976	15.2	97	0.00
4	Chloromethane	0.609	0.520	14.6	96	0.00
5 T	Vinyl Chloride	0.599	0.518	13.5	97	0.00
6 T	Bromomethane	0.385	0.338	12.2	99	0.00
7	Chloroethane	0.223	0.196	12.1	98	0.00
8 T	Dichlorotetrafluoroethane	1.554	1.348	13.3	101	0.00
9 T	Propene	0.483	0.405	16.1	93	0.00
10 T	Heptane	1.292	1.121	13.2	97	0.00
11 T	Trichlorofluoromethane	1.797	1.587	11.7	102	0.00
12 T	1,1,2-Trichlorotrifluoroeth	1.207	1.058	12.3	101	0.00
13	Ethanol	0.125	0.119	4.8	100	0.00
14 T	Bromoethene	0.482	0.422	12.4	96	0.00
15 T	Acetone	1.202	0.978	18.6	102	0.00
16 T	1,3-Butadiene	0.477	0.439	8.0	90	0.00
17	tert-Butyl alcohol	0.221	0.191	13.6	95	0.00
18 T	1,1-Dichloroethene	0.519	0.470	9.4	101	0.00
19 T	Isopropyl Alcohol	0.761	0.701	7.9	103	0.00
20 T	Methylene Chloride	0.510	0.398	22.0	100	0.00
21 T	Allyl Chloride	0.760	0.684	10.0	96	0.00
22 T	trans-1,2-Dichloroethene	0.541	0.473	12.6	99	0.00
23 T	Vinyl Acetate	1.711	1.476	13.7	91	0.00
24 T	1,1-Dichloroethane	1.300	1.060	18.5	92	0.00
25 T	Ethyl Acetate	2.231	1.852	17.0	94	0.00
26 T	Hexane	0.992	0.825	16.8	95	0.00
27 T	Carbon Disulfide	1.180	1.146	2.9	98	0.00
28 T	Methyl tert-Butyl Ether	1.429	1.169	18.2	86	0.00
29 T	Chloroform	1.700	1.438	15.4	97	0.00
30 T	Cyclohexane	0.785	0.713	9.2	97	0.00
31 T	cis-1,2-Dichloroethene	0.986	0.853	13.5	92	0.00
32 T	1,1,1-Trichloroethane	1.619	1.403	13.3	94	0.00
33 I	1,4-Difluorobenzene	1.000	1.000	0.0	111	0.00
34 T	2-Butanone	0.601	0.485	19.3	92	0.00
35 T	Carbon Tetrachloride	0.708	0.645	8.9	99	0.00
36 T	Benzene	0.839	0.706	15.9	95	0.00
37 T	1,2-Dichloroethane	0.562	0.448	20.3	94	0.00
38 T	Trichloroethene	0.316	0.295	6.6	99	0.00
39 T	1,2-Dichloropropane	0.329	0.265	19.5	92	0.00
40 T	1,4-Dioxane	0.117	0.094	19.7	90	0.00
41 T	Tetrahydrofuran	0.330	0.277	16.1	92	0.00
42 T	Bromodichloromethane	0.766	0.660	13.8	96	0.00
43	Methyl Methacrylate	0.341	0.297	12.9	95	0.00
44 T	2,2,4-Trimethylpentane	1.466	1.231	16.0	97	0.00
45 T	t-1,3-Dichloropropene	0.418	0.408	2.4	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL041019\
 Data File : VL033585.D
 Acq On : 11 Apr 2019 4:54
 Operator : SY/AP
 Sample : VSTDCCC010
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Apr 11 10:06:12 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL032519AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LTue Mar 26 16:00:50 2019
 QLast Update : Tue Mar 26 16:00:50 2019
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	cis-1,3-Dichloropropene	0.488	0.457	6.4	95	0.00
47 T	1,1,2-Trichloroethane	0.373	0.313	16.1	97	0.00
48 T	Dibromochloromethane	0.630	0.576	8.6	97	0.00
49 T	Bromoform	0.601	0.533	11.3	73	0.00
50 T	4-Methyl-2-Pentanone	0.880	0.765	13.1	97	0.00
51 T	2-Hexanone	0.656	0.592	9.8	94	0.00
52 T	Tetrachloroethene	0.317	0.285	10.1	95	0.00
53 T	Toluene	0.975	0.861	11.7	94	0.00
54 T	1,2-Dibromoethane	0.584	0.500	14.4	95	0.00
55 I	Chlorobenzene-d5	1.000	1.000	0.0	109	0.00
56	1,1,1,2-Tetrachloroethane	0.458	0.378	17.5	96	0.00
57 T	Chlorobenzene	0.776	0.635	18.2	97	0.00
58 T	Ethyl Benzene	1.335	1.168	12.5	96	0.00
59 T	m/p-Xylene	1.193	1.004	15.8	87	0.00
60 T	o-Xylene	1.153	0.991	14.1	97	0.00
61 T	Styrene	0.792	0.723	8.7	95	0.00
62	Isopropylbenzene	1.633	1.397	14.5	96	0.00
63 T	1,1,2,2-Tetrachloroethane	0.849	0.693	18.4	97	0.00
64	n-propylbenzene	0.414	0.361	12.8	95	0.00
65	tert-Butylbenzene	1.439	1.279	11.1	100	0.00
66 T	Benzyl Chloride	0.669	0.712	-6.4	97	0.00
67	sec-Butylbenzene	2.047	1.810	11.6	99	0.00
68 S	1-Bromo-4-Fluorobenzene	0.784	0.780	0.5	110	0.00
69	p-Isopropyltoluene	1.667	1.509	9.5	98	0.00
70	n-Butylbenzene	1.766	1.577	10.7	98	0.00
71	2-Chlorotoluene	1.242	1.073	13.6	94	0.00
72 T	4-Ethyltoluene	1.297	1.156	10.9	97	0.00
73 T	1,3,5-Trimethylbenzene	1.241	1.126	9.3	99	0.00
74 T	1,2,4-Trimethylbenzene	1.330	1.177	11.5	100	0.00
75 T	1,3-Dichlorobenzene	0.822	0.697	15.2	100	0.00
76 T	1,4-Dichlorobenzene	0.801	0.698	12.9	99	0.00
77 T	1,2-Dichlorobenzene	0.780	0.682	12.6	99	0.00
78 T	Hexachloro-1,3-Butadiene	0.516	0.514	0.4	135	0.00
79 T	Naphthalene	1.138	1.162	-2.1	98	0.00
80 T	Naphthalene,2-methyl-	0.382	0.529	-38.5#	101	0.00
81 T	1,2,4-Trichlorobenzene	0.631	0.605	4.1	100	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0