

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL041023\
 Data File : VL040331.D
 Acq On : 10 Apr 2023 18:55
 Operator : SY/MD
 Sample : 02213-14 LOQ 0.03
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 LOQ-AIR-02-QT2-2023

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 04/11/2023
 Supervised By :Mahesh Dadoda 04/11/2023

Quant Time: Apr 11 02:08:39 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL031523AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Wed Mar 15 16:14:43 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.166	49	878810	10.000	ppbv	-0.02
33) 1,4-Difluorobenzene	6.626	114	2165863	10.000	ppbv	-0.03
55) Chlorobenzene-d5	11.446	117	1963097	10.000	ppbv	-0.03
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.757	95	1637983	9.964	ppbv	-0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.600%
Target Compounds						
32) 1,1,1-Trichloroethane	5.986	97	4272m	0.031	ppbv	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL041023\
Data File : VL040331.D
Acq On : 10 Apr 2023 18:55
Operator : SY/MD
Sample : 02213-14 LOQ 0.03
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
LOQ-AIR-02-QT2-2023

Quant Time: Apr 11 02:08:39 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL031523AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Wed Mar 15 16:14:43 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 04/11/2023
Supervised By : Mahesh Dadoda 04/11/2023

