

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL041122\
 Data File : VL038887.D
 Acq On : 12 Apr 2022 9:56
 Operator : SY/AP
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Apr 13 06:44:20 2022
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL040522AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Apr 06 06:12:53 2022
 QLast Update : Wed Apr 06 06:12:53 2022
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	72	0.00
2 T	Dichlorodifluoromethane	1.418	1.857	-31.0#	104	0.00
3	Chlorodifluoromethane	1.220	1.171	4.0	74	0.00
4	Chloromethane	0.632	0.603	4.6	72	0.00
5 T	Vinyl Chloride	0.584	0.559	4.3	72	0.00
6 T	Bromomethane	0.283	0.270	4.6	72	0.00
7	Chloroethane	0.195	0.184	5.6	68	0.00
8 T	Dichlorotetrafluoroethane	1.325	1.291	2.6	75	0.00
9 T	Propene	0.646	0.617	4.5	72	0.00
10 T	Heptane	1.626	1.513	6.9	72	0.00
11 T	Trichlorofluoromethane	1.140	1.211	-6.2	79	0.00
12 T	1,1,2-Trichlorotrifluoroeth	0.873	0.911	-4.4	79	0.00
13	Ethanol	0.039	0.037#	5.1	78	0.00
14 T	Bromoethene	0.384	0.383	0.3	78	0.00
15 T	Acetone	0.996	0.993	0.3	81	0.00
16 T	1,3-Butadiene	0.569	0.552	3.0	73	0.00
17	tert-Butyl alcohol	0.893	0.648	27.4	52	0.00
18 T	1,1-Dichloroethene	0.398	0.417	-4.8	81	0.00
19 T	Isopropyl Alcohol	0.509	0.419	17.7	60	0.00
20 T	Methylene Chloride	0.361	0.330	8.6	72	0.00
21 T	Allyl Chloride	0.745	0.705	5.4	73	0.00
22 T	trans-1,2-Dichloroethene	0.403	0.428	-6.2	75	0.00
23 T	Vinyl Acetate	0.970	1.009	-4.0	77	0.00
24 T	1,1-Dichloroethane	0.876	0.868	0.9	72	0.00
25 T	Ethyl Acetate	2.526	2.460	2.6	72	0.00
26 T	Hexane	1.150	1.082	5.9	73	0.00
27 T	Carbon Disulfide	1.054	1.090	-3.4	74	0.00
28 T	Methyl tert-Butyl Ether	0.888	0.818	7.9	68	0.00
29 T	Chloroform	1.609	1.610	-0.1	76	0.00
30 T	Cyclohexane	1.013	0.939	7.3	72	0.00
31 T	cis-1,2-Dichloroethene	1.057	1.056	0.1	75	0.00
32 T	1,1,1-Trichloroethane	1.621	1.579	2.6	74	0.00
33 I	1,4-Difluorobenzene	1.000	1.000	0.0	71	0.00
34 T	2-Butanone	0.395	0.385	2.5	71	0.00
35 T	Carbon Tetrachloride	0.600	0.590	1.7	74	0.00
36 T	Benzene	0.910	0.850	6.6	72	0.00
37 T	1,2-Dichloroethane	0.416	0.434	-4.3	76	0.00
38 T	Trichloroethene	0.373	0.336	9.9	71	0.00
39 T	1,2-Dichloropropane	0.348	0.326	6.3	71	0.00
40 T	1,4-Dioxane	0.118	0.108	8.5	77	0.00
41 T	Tetrahydrofuran	0.375	0.359	4.3	72	0.00
42 T	Bromodichloromethane	0.657	0.658	-0.2	75	0.00
43	Methyl Methacrylate	0.348	0.357	-2.6	73	0.00
44 T	2,2,4-Trimethylpentane	1.540	1.408	8.6	73	0.00
45 T	t-1,3-Dichloropropene	0.364	0.357	1.9	65	0.00
46 T	cis-1,3-Dichloropropene	0.480	0.461	4.0	67	0.00
47 T	1,1,2-Trichloroethane	0.357	0.346	3.1	74	0.00
48 T	Dibromochloromethane	0.559	0.561	-0.4	72	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL041122\
 Data File : VL038887.D
 Acq On : 12 Apr 2022 9:56
 Operator : SY/AP
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Apr 13 06:44:20 2022
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL040522AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Apr 06 06:12:53 2022
 QLast Update : Wed Apr 06 06:12:53 2022
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	Bromoform	0.443	0.455	-2.7	74	0.00
50 T	4-Methyl-2-Pentanone	0.889	0.886	0.3	72	0.00
51 T	2-Hexanone	0.670	0.671	-0.1	70	0.00
52 T	Tetrachloroethene	0.318	0.285	10.4	71	0.00
53 T	Toluene	1.050	0.991	5.6	73	0.00
54 T	1,2-Dibromoethane	0.495	0.498	-0.6	74	0.00
55 I	Chlorobenzene-d5	1.000	1.000	0.0	77	0.00
56	1,1,1,2-Tetrachloroethane	0.438	0.397	9.4	73	0.00
57 T	Chlorobenzene	0.819	0.709	13.4	71	0.00
58 T	Ethyl Benzene	1.494	1.360	9.0	75	0.00
59 T	m/p-Xylene	1.255	1.124	10.4	76	0.00
60 T	o-Xylene	1.175	1.059	9.9	76	0.00
61 T	Styrene	0.716	0.665	7.1	72	0.00
62	Isopropylbenzene	1.763	1.553	11.9	74	0.00
63 T	1,1,2,2-Tetrachloroethane	0.744	0.659	11.4	74	0.00
64	n-propylbenzene	0.458	0.403	12.0	71	0.00
65	tert-Butylbenzene	1.564	1.322	15.5	73	0.00
66 T	Benzyl Chloride	0.318	0.193	39.3#	38	0.00
67	sec-Butylbenzene	2.195	1.936	11.8	75	0.00
68 S	1-Bromo-4-Fluorobenzene	0.759	0.781	-2.9	79	0.00
69	p-Isopropyltoluene	1.802	1.603	11.0	73	0.00
70	n-Butylbenzene	1.816	1.717	5.5	77	0.00
71	2-Chlorotoluene	1.326	1.216	8.3	76	0.00
72 T	4-Ethyltoluene	1.403	1.290	8.1	73	0.00
73 T	1,3,5-Trimethylbenzene	1.260	1.157	8.2	74	0.00
74 T	1,2,4-Trimethylbenzene	1.332	1.195	10.3	75	0.00
75 T	1,3-Dichlorobenzene	0.773	0.696	10.0	73	0.00
76 T	1,4-Dichlorobenzene	0.762	0.674	11.5	72	0.00
77 T	1,2-Dichlorobenzene	0.762	0.676	11.3	72	0.00
78 T	Hexachloro-1,3-Butadiene	0.602	0.468	22.3	71	0.00
79 T	Naphthalene	0.996	0.936	6.0	67	0.00
80 T	Naphthalene,2-methyl-	0.246	0.193	21.5	50	0.00
81 T	1,2,4-Trichlorobenzene	0.583	0.522	10.5	70	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0