

Data Path : W:\HPCHEM1\MSVOA L\DATA\VL041218\
 Data File : VL031855.D
 Acq On : 12 Apr 2018 2:50
 Operator : SY/AP
 Sample : VSTDIC0.1
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC0.1

Manual Integrations
 APPROVED

MMDadoda
 4/15/2018 3:06:10 PM

Quant Time: Apr 12 04:44:53 2018
 Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL041218AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Apr 12 04:44:53 2018
 QLast Update : Thu Apr 12 04:17:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.89	49	1311486	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.43	114	2762476	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.39	117	2677059	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.73	95	2204485	10.96	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	109.60%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
5) Vinyl Chloride	3.39	62	8819m	0.104	ppbv	
32) 1,1,1-Trichloroethane	6.74	97	24134m	0.092	ppbv	
35) Carbon Tetrachloride	7.27	117	22955m	0.111	ppbv	
38) Trichloroethene	8.10	130	8874m	0.084	ppbv	
52) Tetrachloroethene	11.53	164	9540m	0.091	ppbv	
63) 1,1,2,2-Tetrachloroethane	13.99	83	25053m	0.108	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA L\DATA\VL041218\
Data File : VL031855.D
Acq On : 12 Apr 2018 2:50
Operator : SY/AP
Sample : VSTDICC0.1
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.1

Manual Integrations
APPROVED

MMDadoda
4/15/2018 3:06:10 PM

Quant Time: Apr 12 04:44:53 2018
Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL041218AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Apr
QLast Update : Thu Apr 12 04:17:02 2018
Response via : Initial Calibration

