

Data Path : W:\HPCHEM1\MSVOA L\DATA\VL041218\
 Data File : VL031861.D
 Acq On : 12 Apr 2018 10:58
 Operator : SY/AP
 Sample : QC032918 CAN 10442
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC032918 CAN 10442

Quant Time: Apr 14 01:53:14 2018
 Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL041218AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Apr 13 16:50:02 2018
 QLast Update : Fri Apr 13 16:50:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.88	49	1362644	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.42	114	2705276	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.39	117	2595499	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.72	95	2093625	10.03	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.30%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
20) Methylene Chloride	4.52	84	42502	0.556	ppbv	# 78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA L\DATA\VL041218\
Data File : VL031861.D
Acq On : 12 Apr 2018 10:58
Operator : SY/AP
Sample : OC032918 CAN 10442
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC032918 CAN 10442

Quant Time: Apr 14 01:53:14 2018
Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL041218AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Apr 13 16:50:02 2018
QLast Update : Fri Apr 13 16:50:02 2018
Response via : Initial Calibration

