

Data Path : W:\HPCHEM1\MSVOA L\DATA\VL041218\
 Data File : VL031864.D
 Acq On : 12 Apr 2018 12:59
 Operator : SY/AP
 Sample : J2191-07DL 4X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 GAC-EFF-040218DL

Manual Integrations
 APPROVED

MMDadoda
 4/15/2018 3:06:48 PM

Quant Time: Apr 14 02:04:38 2018
 Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL041218AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Apr 13 16:50:02 2018
 QLast Update : Fri Apr 13 16:50:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.89	49	1236870	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.43	114	2525624	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.39	117	2471394	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.73	95	1988055	10.00	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.00%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
2) Dichlorodifluoromethane	3.16	85	143513	0.563	ppbv	96
9) Propene	3.13	41	43989m	0.884	ppbv	
11) Trichlorofluoromethane	4.11	101	123588	0.494	ppbv	87
13) Ethanol	3.75	45	285508	16.913	ppbv	96
15) Acetone	4.02	43	90614m	0.476	ppbv	
20) Methylene Chloride	4.52	84	99240	1.431	ppbv	87
24) 1,1-Dichloroethane	5.21	63	96714	0.503	ppbv	99
31) cis-1,2-Dichloroethene	5.77	61	39910m	0.318	ppbv	
32) 1,1,1-Trichloroethane	6.76	97	199873	0.875	ppbv	99
38) Trichloroethene	8.10	130	327992	3.557	ppbv	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA L\DATA\VL041218\
Data File : VL031864.D
Acq On : 12 Apr 2018 12:59
Operator : SY/AP
Sample : J2191-07DL 4X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
GAC-EFF-040218DL

Manual Integrations
APPROVED

MMDadoda
4/15/2018 3:06:48 PM

Quant Time: Apr 14 02:04:38 2018
Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL041218AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Apr
QLast Update : Fri Apr 13 16:50:02 2018
Response via : Initial Calibration

