

Data Path : W:\HPCHEM1\MSVOA L\DATA\VL041218\
 Data File : VL031872.D
 Acq On : 12 Apr 2018 19:25
 Operator : SY/AP
 Sample : J2191-03DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SSDS-INF-040218DL

Manual Integrations
 APPROVED

MMDadoda
 4/15/2018 3:07:45 PM

Quant Time: Apr 14 03:45:32 2018
 Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL041218AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Apr 13 2018
 QLast Update : Fri Apr 13 16:50:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.89	49	1272606	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.43	114	2728150	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.39	117	2584470	10.00	ppbv	0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.73	95	2078009	10.00	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.00%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
2) Dichlorodifluoromethane	3.17	85	143198	0.546	ppbv	94
9) Propene	3.14	41	43290m	0.846	ppbv	
10) Heptane	8.39	43	28221m	0.170	ppbv	
11) Trichlorofluoromethane	4.11	101	254654	0.990	ppbv	97
13) Ethanol	3.75	45	337752	19.446	ppbv	99
15) Acetone	4.01	43	432591	2.209	ppbv	100
20) Methylene Chloride	4.52	84	84725	1.188	ppbv	96
26) Hexane	5.91	57	51272	0.360	ppbv #	45
29) Chloroform	5.98	83	32001m	0.133	ppbv	
30) Cyclohexane	7.40	84	42966m	0.422	ppbv	
32) 1,1,1-Trichloroethane	6.76	97	10311m	0.044	ppbv	
34) 2-Butanone	5.47	43	22103m	0.117	ppbv	
53) Toluene	10.10	91	423192	1.546	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA L\DATA\VL041218\
Data File : VL031872.D
Acq On : 12 Apr 2018 19:25
Operator : SY/AP
Sample : J2191-03DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SSDS-INF-040218DL

Manual Integrations
APPROVED

MMDadoda
4/15/2018 3:07:45 PM

Quant Time: Apr 14 03:45:32 2018
Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL041218AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Apr
QLast Update : Fri Apr 13 16:50:02 2018
Response via : Initial Calibration

