

Data Path : W:\HPCHEM1\MSVOA L\DATA\VL041218\
 Data File : VL031874.D
 Acq On : 12 Apr 2018 20:44
 Operator : SY/AP
 Sample : J2191-04DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SVE-INF3-040218DL

Manual Integrations
 APPROVED

MMDadoda
 4/15/2018 3:07:55 PM

Quant Time: Apr 14 03:53:52 2018
 Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL041218AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Apr 13 16:50:02 2018
 QLast Update : Fri Apr 13 16:50:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.89	49	1293057	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.43	114	2674927	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.39	117	2589278	10.00	ppbv	0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.73	95	2079858	9.99	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.90%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
11) Trichlorofluoromethane	4.12	101	106209	0.406	ppbv	84
13) Ethanol	3.77	45	14034m	0.795	ppbv	
15) Acetone	4.02	43	112264	0.564	ppbv	100
20) Methylene Chloride	4.52	84	34060	0.470	ppbv #	85
24) 1,1-Dichloroethane	5.21	63	260562	1.296	ppbv	98
31) cis-1,2-Dichloroethene	5.77	61	158330	1.208	ppbv	94
32) 1,1,1-Trichloroethane	6.76	97	462300	1.935	ppbv	99
38) Trichloroethene	8.10	130	894781	9.162	ppbv	95
52) Tetrachloroethene	11.53	164	7258m	0.077	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA L\DATA\VL041218\
Data File : VL031874.D
Acq On : 12 Apr 2018 20:44
Operator : SY/AP
Sample : J2191-04DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SVE-INF3-040218DL

Manual Integrations
APPROVED

MMDadoda
4/15/2018 3:07:55 PM

Quant Time: Apr 14 03:53:52 2018
Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL041218AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Apr
QLast Update : Fri Apr 13 16:50:02 2018
Response via : Initial Calibration

