

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL041223\
 Data File : VL040336.D
 Acq On : 12 Apr 2023 10:43
 Operator : SY/MD
 Sample : VL0412ABL01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0412ABL01

Quant Time: Apr 13 05:36:30 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL031523AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Tue Apr 11 04:15:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.189	49	1100072	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.652	114	2723206	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.475	117	2416838	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	13.786	95	1936100	9.566	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	95.700%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL041223\
Data File : VL040336.D
Acq On : 12 Apr 2023 10:43
Operator : SY/MD
Sample : VL0412ABL01
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0412ABL01

Quant Time: Apr 13 05:36:30 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL031523AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Tue Apr 11 04:15:53 2023
Response via : Initial Calibration

