

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL041223\
 Data File : VL040345.D
 Acq On : 12 Apr 2023 19:10
 Operator : SY/MD
 Sample : 02311-01DL 20X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SV1DL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 04/13/2023
 Supervised By :Mahesh Dadoda 04/13/2023

Quant Time: Apr 13 05:40:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL031523AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_LFri Aug 2
 QLast Update : Tue Apr 11 04:15:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.195	49	1146915	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.658	114	3008705	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.481	117	2744053	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.793	95	2132747	9.281	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	92.800%
Target Compounds						
					Qvalue	
9) Propene	2.693	41	109394m	1.980	ppbv	
10) Heptane	7.590	43	243566	1.833	ppbv	96
13) Ethanol	3.256	45	7939m	1.959	ppbv	
20) Methylene Chloride	3.931	84	20036	0.341	ppbv #	87
26) Hexane	5.214	57	374232	3.488	ppbv #	35
52) Tetrachloroethene	10.632	164	5939m	0.086	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL041223\
Data File : VL040345.D
Acq On : 12 Apr 2023 19:10
Operator : SY/MD
Sample : 02311-01DL 20X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SV1DL

Quant Time: Apr 13 05:40:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL031523AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Tue Apr 11 04:15:53 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 04/13/2023
Supervised By :Mahesh Dadoda 04/13/2023

