

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL041223\
 Data File : VL040352.D
 Acq On : 13 Apr 2023 1:21
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Apr 13 08:16:55 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL031523AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Tue Apr 11 04:15:53 2023
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	93	0.00
2 T	Dichlorodifluoromethane	1.962	1.370	30.2#	98	0.00
3	Chlorodifluoromethane	1.257	1.181	6.0	98	0.00
4	Chloromethane	0.776	0.756	2.6	99	0.00
5 T	Vinyl Chloride	0.629	0.655	-4.1	101	0.00
6 T	Bromomethane	0.218	0.213	2.3	97	0.00
7	Chloroethane	0.241	0.244	-1.2	99	0.00
8 T	Dichlorotetrafluoroethane	1.679	1.596	4.9	97	0.00
9 T	Propene	0.482	0.474	1.7	103	0.00
10 T	Heptane	1.159	1.181	-1.9	94	0.00
11 T	Trichlorofluoromethane	1.778	1.647	7.4	92	0.00
12 T	1,1,2-Trichlorotrifluoroeth	1.335	1.149	13.9	85	0.00
13	Ethanol	0.035	0.029#	17.1	89	0.00
14 T	Bromoethene	0.498	0.508	-2.0	96	0.00
15 T	Acetone	1.217	1.092	10.3	93	0.00
16 T	1,3-Butadiene	0.553	0.569	-2.9	98	0.00
17	tert-Butyl alcohol	1.157	1.176	-1.6	92	0.00
18 T	1,1-Dichloroethene	0.573	0.532	7.2	90	0.00
19 T	Isopropyl Alcohol	0.583	0.547	6.2	92	0.00
20 T	Methylene Chloride	0.512	0.436	14.8	89	0.00
21 T	Allyl Chloride	0.789	0.764	3.2	91	0.00
22 T	trans-1,2-Dichloroethene	0.583	0.518	11.1	84	0.00
23 T	Vinyl Acetate	0.339	0.334	1.5	93	0.00
24 T	1,1-Dichloroethane	1.208	1.103	8.7	90	0.00
25 T	Ethyl Acetate	2.279	2.215	2.8	99	0.00
26 T	Hexane	0.936	0.949	-1.4	103	0.00
27 T	Carbon Disulfide	1.307	1.300	0.5	88	0.00
28 T	Methyl tert-Butyl Ether	1.074	0.993	7.5	91	0.00
29 T	Chloroform	1.652	1.630	1.3	109	0.00
30 T	Cyclohexane	0.737	0.802	-8.8	98	0.00
31 T	cis-1,2-Dichloroethene	0.911	0.983	-7.9	105	0.00
32 T	1,1,1-Trichloroethane	1.583	1.650	-4.2	97	0.00
33 I	1,4-Difluorobenzene	1.000	1.000	0.0	89	0.00
34 T	2-Butanone	0.494	0.510	-3.2	96	0.00
35 T	Carbon Tetrachloride	0.628	0.669	-6.5	92	0.00
36 T	Benzene	0.762	0.787	-3.3	95	0.00
37 T	1,2-Dichloroethane	0.496	0.503	-1.4	95	0.00
38 T	Trichloroethene	0.416	0.470	-13.0	92	0.00
39 T	1,2-Dichloropropane	0.299	0.303	-1.3	95	0.00
40 T	1,4-Dioxane	0.118	0.095	19.5	70	0.00
41 T	Tetrahydrofuran	0.285	0.297	-4.2	95	0.00
42 T	Bromodichloromethane	0.678	0.705	-4.0	93	0.00
43	Methyl Methacrylate	0.283	0.311	-9.9	93	0.00
44 T	2,2,4-Trimethylpentane	1.271	1.312	-3.2	95	0.00
45 T	t-1,3-Dichloropropene	0.247	0.297	-20.2	91	0.00
46 T	cis-1,3-Dichloropropene	0.351	0.408	-16.2	94	0.00
47 T	1,1,2-Trichloroethane	0.319	0.323	-1.3	93	0.00
48 T	Dibromochloromethane	0.525	0.539	-2.7	88	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL041223\
 Data File : VL040352.D
 Acq On : 13 Apr 2023 1:21
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Apr 13 08:16:55 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL031523AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Tue Apr 11 04:15:53 2023
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	Bromoform	0.394	0.400	-1.5	83	0.00
50 T	4-Methyl-2-Pentanone	0.764	0.779	-2.0	90	0.00
51 T	2-Hexanone	0.592	0.570	3.7	83	0.00
52 T	Tetrachloroethene	0.231	0.255	-10.4	90	0.00
53 T	Toluene	0.799	0.881	-10.3	93	0.00
54 T	1,2-Dibromoethane	0.328	0.350	-6.7	89	0.00
55 I	Chlorobenzene-d5	1.000	1.000	0.0	86	0.00
56	1,1,1,2-Tetrachloroethane	0.421	0.390	7.4	88	0.00
57 T	Chlorobenzene	0.711	0.654	8.0	88	0.00
58 T	Ethyl Benzene	1.139	1.185	-4.0	91	0.00
59 T	m/p-Xylene	1.043	1.020	2.2	89	0.00
60 T	o-Xylene	0.997	0.980	1.7	88	0.00
61 T	Styrene	0.506	0.551	-8.9	88	0.00
62	Isopropylbenzene	1.420	1.375	3.2	89	0.00
63 T	1,1,2,2-Tetrachloroethane	0.351	0.305	13.1	88	0.00
64	n-propylbenzene	0.347	0.347	0.0	90	0.00
65	tert-Butylbenzene	1.224	1.205	1.6	91	0.00
66 T	Benzyl Chloride	0.080	0.080	0.0	89	0.00
67	sec-Butylbenzene	1.750	1.755	-0.3	94	0.00
68 S	1-Bromo-4-Fluorobenzene	0.837	0.811	3.1	86	0.00
69	p-Isopropyltoluene	1.363	1.430	-4.9	95	0.00
70	n-Butylbenzene	1.501	1.602	-6.7	98	0.00
71	2-Chlorotoluene	1.081	1.052	2.7	89	0.00
72 T	4-Ethyltoluene	1.090	1.131	-3.8	91	0.00
73 T	1,3,5-Trimethylbenzene	0.988	1.006	-1.8	92	0.00
74 T	1,2,4-Trimethylbenzene	1.126	1.127	-0.1	94	0.00
75 T	1,3-Dichlorobenzene	0.610	0.621	-1.8	97	0.00
76 T	1,4-Dichlorobenzene	0.582	0.614	-5.5	98	0.00
77 T	1,2-Dichlorobenzene	0.586	0.615	-4.9	102	0.00
78 T	Hexachloro-1,3-Butadiene	0.430	0.431	-0.2	99	0.00
79 T	Naphthalene	0.666	0.876	-31.5#	95	0.00
80 T	Naphthalene,2-methyl-	0.169	0.128	24.3	48	0.00
81 T	1,2,4-Trichlorobenzene	0.410	0.470	-14.6	96	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0