

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL041321\
 Data File : VL036806.D
 Acq On : 13 Apr 2021 16:55
 Operator : SY/AP
 Sample : VIBLK
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VIBLK

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 04/14/2021
 Supervised By :Mahesh Dadoda 04/14/2021

Quant Time: Apr 14 07:20:03 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL032521A
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
 QLast Update : Thu Mar 25 09:42:40 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.66	49	1365489	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.18	114	3241299	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.09	117	2431832	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.42	95	1584060	8.17	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	81.70%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
15) Acetone	3.87	43	46551m	0.346	ppbv	
20) Methylene Chloride	4.34	84	8409m	0.134	ppbv	
30) Cyclohexane	7.16	84	26951m	0.224	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL041321\
Data File : VL036806.D
Acq On : 13 Apr 2021 16:55
Operator : SY/AP
Sample : VIBLK
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Manual Integrations
APPROVED

Quant Time: Apr 14 07:20:03 2021

Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL032521

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:

QLast Update : Thu Mar 25 09:42:40 2021

Response via : Initial Calibration

Reviewed By :Semsettin Yesilyurt 04/14/2021
Supervised By :Mahesh Dadoda 04/14/2021

