

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL041422\
Data File : VL038938.D
Acq On : 15 Apr 2022 8:00
Operator : SY/AP
Sample : VIBLK
Misc : 400mL/MSVOA L
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Quant Time: Apr 15 09:58:02 2022
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL040522AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Apr 06 06:12:53 2022
QLast Update : Wed Apr 06 06:12:53 2022
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	1168595	10.00	ppbv	-0.01
33) 1,4-Difluorobenzene	7.16	114	3078833	10.00	ppbv	-0.01
55) Chlorobenzene-d5	12.07	117	2772160	10.00	ppbv	-0.01

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.39	95	2067090	9.83	ppbv	-0.01
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.30%

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL041422\
Data File : VL038938.D
Acq On : 15 Apr 2022 8:00
Operator : SY/AP
Sample : VIBLK
Misc : 400mL/MSVOA L
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Quant Time: Apr 15 09:58:02 2022
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL040522AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Apr 06 06:12:53 2022
QLast Update : Wed Apr 06 06:12:53 2022
Response via : Initial Calibration

