

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL041422\
 Data File : VL038928.D
 Acq On : 14 Apr 2022 22:04
 Operator : SY/AP
 Sample : N2259-02
 Misc : 400mL/MSVOA L
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA L
 ClientSampleId :
 1165-19-SG-2

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 04/15/2022
 Supervised By : Mahesh Dadoda 04/18/2022

Quant Time: Apr 15 10:12:53 2022

Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL040522A

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:

QLast Update : Wed Apr 06 06:12:53 2022

Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.67	49	1206982	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.18	114	2936460	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.09	117	2660218	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.41	95	2162217	10.72	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	107.20%

Target Compounds

						Ovalue
2) Dichlorodifluoromethane	3.05	85	110780	0.647	ppbv	97
4) Chloromethane	3.15	50	38757	0.508	ppbv	96
9) Propene	3.03	41	399731m	5.129	ppbv	
10) Heptane	8.12	43	2671392	13.608	ppbv	98
11) Trichlorofluoromethane	3.95	101	440754	3.204	ppbv	99
13) Ethanol	3.61	45	640343	135.202	ppbv #	100
15) Acetone	3.85	43	24270483	201.975	ppbv #	74
19) Isopropyl Alcohol	3.97	45	372790	6.070	ppbv #	100
20) Methylene Chloride	4.36	84	187929	4.312	ppbv	90
25) Ethyl Acetate	5.68	43	360748	1.183	ppbv	99
26) Hexane	5.69	57	183787	1.324	ppbv	89
30) Cyclohexane	7.13	84	89833	0.734	ppbv #	63
34) 2-Butanone	5.25	43	200203	1.725	ppbv	98
35) Carbon Tetrachloride	7.02	117	11737m	0.067	ppbv	
36) Benzene	6.89	78	250481	0.938	ppbv	97
44) 2,2,4-Trimethylpentane	7.87	57	212767	0.470	ppbv #	1
50) 4-Methyl-2-Pentanone	8.75	43	244225	0.936	ppbv	94
53) Toluene	9.81	91	14757662	47.856	ppbv	90
58) Ethyl Benzene	12.71	91	655578	1.649	ppbv	91
59) m/p-Xylene	12.97	91	1770707	5.302	ppbv	98
60) o-Xylene	13.69	91	693736	2.220	ppbv	94
61) Styrene	13.53	104	35069	0.184	ppbv	92
62) Isopropylbenzene	14.66	105	91909m	0.196	ppbv	
64) n-propylbenzene	15.56	120	71093m	0.584	ppbv	
65) tert-Butylbenzene	16.76	119	123607	0.297	ppbv #	56
67) sec-Butylbenzene	17.29	105	83333	0.143	ppbv	99
69) p-Isopropyltoluene	17.64	119	45685	0.095	ppbv #	86
70) n-Butylbenzene	18.55	91	185670	0.384	ppbv #	57
72) 4-Ethyltoluene	15.83	105	385246	1.032	ppbv	98
73) 1,3,5-Trimethylbenzene	16.00	105	339917	1.014	ppbv	92
74) 1,2,4-Trimethylbenzene	16.76	105	1197452	3.379	ppbv #	69
79) Naphthalene	22.20	128	91838m	0.346	ppbv	
80) Naphthalene,2-methyl-	24.97	142	23916	0.366	ppbv #	58

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL041422\
Data File : VL038928.D
Acq On : 14 Apr 2022 22:04
Operator : SY/AP
Sample : N2259-02
Misc : 400mL/MSVOA L
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
1165-19-SG-2

Manual Integrations
APPROVED

Quant Time: Apr 15 10:12:53 2022


Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL040522


Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:

QLast Update : Wed Apr 06 06:12:53 2022

Response via : Initial Calibration

Reviewed By :Semsettin Yesilyurt 04/15/2022
Supervised By :Mahesh Dadoda 04/18/2022





#1

Bromochloromethane

Concen: 10.000 ppbv

RT: 5.67 Instrument: MSVOA_1

Delta R.T. 0.00

Lab File: ClientSampleId:

Acq: 14 Apr 2022 22:04

Tot Ion:

Ion Ratio

49 100

128 46.


130 58.

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 04/15/2022




Supervised By :Mahesh Dadoda 04/18/2022



Abundance Ion 49.10 (48.80 to 49.80): VL038812.D (-880) (-)

Ion 128.00 (127.70 to 128.70): VL038812.D (-880) (-)

Ion 130.00 (129.70 to 130.70): VL038812.D (-880) (-)



#2

Dichlorodifluoromethane

Concen: 0.647 ppbv

RT: 3.05 min Scan# 83

Delta R.T. 0.00 min

Lab File: VL038928.D


Acq: 14 Apr 2022 22:04

Tgt Ion: 85 Resp: 110780

Ion Ratio Lower Upper



85 100


87 34.1 25.9 38.9



Abundance Ion 85.00 (84.70 to 85.70): VL038812.D (-71) (-)

Ion 87.00 (86.70 to 87.70): VL038812.D (-71) (-)





#4

Chloromethane

Concen: 0.508 ppbv

RT: 3.15 Instrument: MSVOA_1

Delta R.T.

Lab File: ClientSampleId:

Acq: 14 Apr 2022 22:04

Tot Ion:

Ion Ratio

50 100





52 32.

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 04/15/2022

Supervised By :Mahesh Dadoda 04/18/2022



#9

Propene

Concen: 5.129 ppbv m

RT: 3.03 min Scan# 75

Delta R.T. 0.02 min

Lab File: VL038928.D




Acq: 14 Apr 2022 22:04


Tgt Ion: 41 Resp: 399731

Ion Ratio Lower Upper

41 100

42 28.9 54.2 81.4#









#10
 Heptane
 Concen: 13.608 ppbv
 RT: 8.12
 Instrument: MSVOA_1
 Delta R.T.:
 Lab File: ClientSampleId:
 1165-19-SG-2
 Acq: 14 Apr 2022 22:04

Tot Ion:
 Ion Ratio
 43 100
 57 50.




Manual Integrations
 APPROVED


Reviewed By :Semsettin Yesilyurt 04/15/2022
 Supervised By :Mahesh Dadoda 04/18/2022



#11
 Trichlorofluoromethane
 Concen: 3.204 ppbv
 RT: 3.95 min Scan# 363
 Delta R.T.: 0.00 min
 Lab File: VL038928.D
 Acq: 14 Apr 2022 22:04

Tgt Ion:101 Resp: 440754
 Ion Ratio Lower Upper
 101 100
 103 63.3 51.4 77.0









#13
Ethanol
Concen: 135.202 ppbv
RT: 3.61
Delta R.T. MSVOA_L
Lab File: ClientSampleId :
Acq: 14 Apr 2022 22:04

Tot Ion:
Ion Ratio
45 100
46 37.




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 04/15/2022
Supervised By :Mahesh Dadoda 04/18/2022



#15
Acetone
Concen: 201.975 ppbv
RT: 3.85 min Scan# 331
Delta R.T. 0.01 min
Lab File: VL038928.D
Acq: 14 Apr 2022 22:04

Tgt Ion: 43 Resp:24270483
Ion Ratio Lower Upper
43 100
58 41.8 22.5 33.7#





#19

Isopropyl Alcohol

Concen: 6.070 ppbv

RT: 3.97 Instrument:

Delta R.T. MSVOA_1

Lab File: ClientSampleId:

Acq: 14 Apr 2022 22:04

Tot Ion:

Ion Ratio

45 100





58 0.

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 04/15/2022

Supervised By :Mahesh Dadoda 04/18/2022



#20

Methylene Chloride

Concen: 4.312 ppbv

RT: 4.36 min Scan# 488

Delta R.T. 0.01 min

Lab File: VL038928.D

Acq: 14 Apr 2022 22:04

Tgt Ion: 84 Resp: 187929




Ion Ratio Lower Upper




84 100

49 148.1 129.4 194.0

51 42.6 37.4 56.2

86 56.1 51.4 77.2









#25
Ethyl Acetate
Concen: 1.183 ppbv
RT: 5.68
Instrument: MSVOA_1
Delta R.T.:
Lab File: ClientSampleId : 1165-19-SG-2
Acq: 14 Apr 2022 22:04

Tot Ion:
Ion Ratio
43 100
45 10.
61 8.
70 6.9 5.6 8.4


Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 04/15/2022
Supervised By :Mahesh Dadoda 04/18/2022



#26
Hexane
Concen: 1.324 ppbv
RT: 5.69 min Scan# 903
Delta R.T. 0.01 min
Lab File: VL038928.D
Acq: 14 Apr 2022 22:04

Tgt Ion: 57 Resp: 183787
Ion Ratio Lower Upper
57 100
41 93.0 72.0 108.0
43 206.3 186.6 280.0
56 55.0 42.8 64.2









#30
Cvclohexane
Concen: 0.734 ppbv
RT: 7.13 min
Instrument: MSVOA_1
Delta R.T.:
Lab File: ClientSampleId:
Acq: 14 Apr 2022 22:04

Tot Ion:
Ion Ratio
84 100
56 121.
41 150.




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 04/15/2022
Supervised By :Mahesh Dadoda 04/18/2022



#33
1,4-Difluorobenzene
Concen: 10.000 ppbv
RT: 7.18 min Scan# 1366
Delta R.T. 0.01 min
Lab File: VL038928.D
Acq: 14 Apr 2022 22:04

Tgt Ion:114 Resp: 2936460
Ion Ratio Lower Upper
114 100
63 26.5 19.8 29.8
88 18.8 14.2 21.4









#34
2-Butanone
Concen: 1.725 ppbv
RT: 5.25
Delta R.T. MSVOA_L
Lab File: ClientSampleId :
Acq: 14 Apr 2022 22:04

Tot Ion:
Ion Ratio
43 100
72 21.




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 04/15/2022
Supervised By :Mahesh Dadoda 04/18/2022



#35
Carbon Tetrachloride
Concen: 0.067 ppbv m
RT: 7.02 min Scan# 1316
Delta R.T. 0.00 min
Lab File: VL038928.D
Acq: 14 Apr 2022 22:04

Tgt Ion:117 Resp: 11737
Ion Ratio Lower Upper
117 100
119 109.5 77.7 116.5
121 40.1 24.5 36.7#





#36

Benzene

Concen: 0.938 ppbv

RT: 6.89 Instrument :

Delta R.T. MSVOA_1





Lab File: ClientSampleId :

Acq: 14 Apr 2022 22:04

Tot Ion:
Ion Ratio
78 100
77 20.
50 17.

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 04/15/2022
Supervised By :Mahesh Dadoda 04/18/2022



#44

2,2,4-Trimethylpentane

Concen: 0.470 ppbv




RT: 7.87 min Scan# 1581


Delta R.T. 0.00 min

Lab File: VL038928.D

Acq: 14 Apr 2022 22:04

Tgt Ion: 57 Resp: 212767
Ion Ratio Lower Upper
57 100
41 109.5 25.0 37.4#
56 135.9 25.3 37.9#









#50
4-Methyl-2-pentanone
Concen: 0.936 ppbv
RT: 8.75
Delta R.T. MSVOA_L
Lab File: ClientSampleId :
Acq: 14 Apr 2022 22:04

Tot Ion:
Ion Ratio
43 100
58 40.




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 04/15/2022
Supervised By :Mahesh Dadoda 04/18/2022



#53
Toluene
Concen: 47.856 ppbv
RT: 9.81 min Scan# 2185
Delta R.T. 0.01 min
Lab File: VL038928.D
Acq: 14 Apr 2022 22:04

Tgt Ion: 91 Resp:14757662
Ion Ratio Lower Upper
91 100
92 69.4 49.2 73.8





#55

Chlorobenzene-d5

Concen: 10.000 ppbv

RT: 12.09

Delta R.T. MSVOA_L

Lab File: ClientSampleId:

Acq: 14 Apr 2022 22:04

Tot Ion:1

Ion Rati

117 100

82 67.





119 32.

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 04/15/2022

Supervised By :Mahesh Dadoda 04/18/2022



#58

Ethyl Benzene

Concen: 1.649 ppbv

RT: 12.71 min Scan# 3086

Delta R.T. 0.01 min

Lab File: VL038928.D




Acq: 14 Apr 2022 22:04


Tgt Ion: 91 Resp: 655578

Ion Ratio Lower Upper

91 100

106 28.3 26.7 40.1









#59
m/p-Xylene
Concen: 5.302 ppbv
RT: 12.97
Instrument: MSVOA_1
Delta R.T.:
Lab File: ClientSampled:
Acq: 14 Apr 2022 22:04

Tot Ion:
Ion Ratio
91 100
106 45.




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 04/15/2022
Supervised By :Mahesh Dadoda 04/18/2022



#60
o-Xylene
Concen: 2.220 ppbv
RT: 13.69 min Scan# 3391
Delta R.T. 0.00 min
Lab File: VL038928.D
Acq: 14 Apr 2022 22:04

Tgt Ion: 91 Resp: 693736
Ion Ratio Lower Upper
91 100
106 43.6 23.8 71.4









#61
Stvrene
Concen: 0.184 ppbv
RT: 13.53
Delta R.T. MSVOA_L
Lab File: ClientSampled :
Acq: 14 Apr 2022 22:04

Tot Ion:1
Ion Rati
104 100
78 55.
103 54.




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 04/15/2022
Supervised By :Mahesh Dadoda 04/18/2022



#62
Isopropylbenzene
Concen: 0.196 ppbv m
RT: 14.66 min Scan# 3694
Delta R.T. 0.00 min
Lab File: VL038928.D
Acq: 14 Apr 2022 22:04

Tgt Ion:105 Resp: 91909
Ion Ratio Lower Upper
105 100
120 26.3 21.8 32.8





#64

n-propylbenzene

Concen: 0.584 ppbv m

RT: 15.56

Delta R.T. MSVOA_L

Lab File: ClientSampled :

Acq: 14 Apr 2022 22:04

Tot Ion:1

Ion Rati

120 100





91 525.

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 04/15/2022

Supervised By :Mahesh Dadoda 04/18/2022



#65

tert-Butylbenzene

Concen: 0.297 ppbv

RT: 16.76 min Scan# 4346

Delta R.T. 0.03 min

Lab File: VL038928.D

Acq: 14 Apr 2022 22:04




Tgt Ion:119 Resp: 123607


Ion Ratio Lower Upper

119 100

91 116.8 62.8 94.2#

134 0.1 16.2 24.4#





#67

sec-Butylbenzene

Concen: 0.143 ppbv

RT: 17.29

Delta R.T. MSVOA_L

Lab File: ClientSampleId :

Acq: 14 Apr 2022 22:04

Tot Ion:1

Ion Ratio

105 100





134 18.

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 04/15/2022

Supervised By :Mahesh Dadoda 04/18/2022



#68

1-Bromo-4-Fluorobenzene

Concen: 10.715 ppbv

RT: 14.41 min Scan# 3615

Delta R.T. 0.00 min

Lab File: VL038928.D

Acq: 14 Apr 2022 22:04




Tgt Ion: 95 Resp: 2162217


Ion Ratio Lower Upper

95 100

174 62.3 52.0 78.0

175 4.6 4.0 6.0









#69
 p-Isopropyltoluene
 Concen: 0.095 ppbv
 RT: 17.64 min
 Delta R.T. MSVOA_L
 Lab File: ClientSampleId :
 1165-19-SG-2
 Acq: 14 Apr 2022 22:04

Tot Ion: 1
 Ion Ratio
 119 100
 134 30.
 91 38.




Manual Integrations
 APPROVED


Reviewed By : Semsettin Yesilyurt 04/15/2022
 Supervised By : Mahesh Dadoda 04/18/2022



#70
 n-Butylbenzene
 Concen: 0.384 ppbv
 RT: 18.55 min Scan# 4902
 Delta R.T. 0.01 min
 Lab File: VL038928.D
 Acq: 14 Apr 2022 22:04

Tgt Ion: 91 Resp: 185670
 Ion Ratio Lower Upper
 91 100
 92 40.4 44.6 66.8#
 134 70.3 19.2 28.8#





#72

4-Ethyltoluene

Concen: 1.032 ppbv

RT: 15.83

Delta R.T. MSVOA_L

Lab File: ClientSampled :

Acq: 14 Apr 2022 22:04

Tot Ion:1

Ion Rati

105 100

120 29.

91 13.



77 13.9

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 04/15/2022

Supervised By :Mahesh Dadoda 04/18/2022




Abundance Ion 105.10 (104.80 to 105.80): V

Ion 120.20 (119.90 to 120.90): V

Ion 91.10 (90.80 to 91.80): VL0

Ion 77.10 (76.80 to 77.80): VL0

Time-->



#73

1,3,5-Trimethylbenzene

Concen: 1.014 ppbv

RT: 16.00 min Scan# 4108

Delta R.T. 0.01 min

Lab File: VL038928.D



Acq: 14 Apr 2022 22:04

Tgt Ion:105 Resp: 339917

Ion Ratio Lower Upper

105 100


120 43.2 38.7 58.1





Abundance Ion 105.10 (104.80 to 105.80): V

Ion 120.20 (119.90 to 120.90): V

Time-->





#74
1,2,4-Trimethylbenzene
Concen: 3.379 ppbv
RT: 16.76
Delta R.T. MSVOA_L
Lab File: ClientSampleId :
Acq: 14 Apr 2022 22:04





Tot Ion:1
Ion Ratio
105 100
120 39.
91 11.
77 14.5 17.3 25.9#

Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 04/15/2022
Supervised By :Mahesh Dadoda 04/18/2022




#79
Naphthalene
Concen: 0.346 ppbv m
RT: 22.20 min Scan# 6037
Delta R.T. 0.02 min
Lab File: VL038928.D
Acq: 14 Apr 2022 22:04



Tgt Ion:128 Resp: 91838
Ion Ratio Lower Upper
128 100
127 9.9 10.6 15.8#
129 16.2 9.4 14.0#





#80
 Naphthalene, 2-methyl-
 Concen: 0.366 ppbv
 RT: 24.97
 Instrument : MSVOA_1
 Delta R.T :
 Lab File: ClientSampleId :
 1165-19-SG-2
 Acq: 14 Apr 2022 22:04

Tot Ion:1
 Ion Rati
 142 100
 141 78.
 115 110.

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 04/15/2022
 Supervised By :Mahesh Dadoda 04/18/2022

