

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL041422\
 Data File : VL038930.D
 Acq On : 14 Apr 2022 23:25
 Operator : SY/AP
 Sample : N2259-04
 Misc : 400mL/MSVOA L
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA L
 ClientSampleId :
 1165-19-SG-4

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 04/15/2022
 Supervised By : Mahesh Dadoda 04/18/2022

Quant Time: Apr 15 10:23:25 2022

Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL040522A

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:

QLast Update : Wed Apr 06 06:12:53 2022

Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.67	49	1219953	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.18	114	3009538	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.09	117	2636717	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.41	95	2181406	10.91	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	109.10%

Target Compounds

						Ovalue
2) Dichlorodifluoromethane	3.05	85	94236	0.545	ppbv	98
4) Chloromethane	3.14	50	38513	0.500	ppbv	93
9) Propene	3.02	41	354099m	4.495	ppbv	
10) Heptane	8.12	43	3762424	18.962	ppbv	98
11) Trichlorofluoromethane	3.95	101	186426	1.341	ppbv	96
13) Ethanol	3.61	45	897841	187.555	ppbv #	100
15) Acetone	3.85	43	27874327	229.499	ppbv #	67
19) Isopropyl Alcohol	3.97	45	351004	5.654	ppbv #	100
20) Methylene Chloride	4.35	84	367116	8.334	ppbv	98
25) Ethyl Acetate	5.68	43	550603	1.787	ppbv	97
26) Hexane	5.69	57	311762	2.222	ppbv #	82
30) Cyclohexane	7.13	84	121183	0.980	ppbv #	66
34) 2-Butanone	5.25	43	247834	2.084	ppbv #	91
36) Benzene	6.89	78	290333	1.060	ppbv	99
39) 1,2-Dichloropropane	7.18	63	753194	7.193	ppbv #	26
44) 2,2,4-Trimethylpentane	7.87	57	293365m	0.633	ppbv	
50) 4-Methyl-2-Pentanone	8.74	43	1081775	4.044	ppbv	99
53) Toluene	9.81	91	17380569	54.993	ppbv	87
58) Ethyl Benzene	12.71	91	965836	2.451	ppbv	93
59) m/p-Xylene	12.96	91	2480374	7.493	ppbv	100
60) o-Xylene	13.69	91	988313	3.191	ppbv	95
61) Styrene	13.52	104	41090m	0.218	ppbv	
62) Isopropylbenzene	14.66	105	133111	0.286	ppbv	95
64) n-propylbenzene	15.56	120	109156	0.904	ppbv	75
65) tert-Butylbenzene	16.76	119	184555	0.447	ppbv #	62
67) sec-Butylbenzene	17.28	105	139780	0.241	ppbv	96
69) p-Isopropyltoluene	17.64	119	76549m	0.161	ppbv	
70) n-Butylbenzene	18.55	91	293441	0.613	ppbv #	57
72) 4-Ethyltoluene	15.83	105	573428	1.550	ppbv	98
73) 1,3,5-Trimethylbenzene	15.99	105	490748	1.477	ppbv	97
74) 1,2,4-Trimethylbenzene	16.76	105	1746142	4.971	ppbv #	71
79) Naphthalene	22.18	128	121444m	0.462	ppbv	
80) Naphthalene,2-methyl-	24.97	142	33088	0.511	ppbv #	64

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL041422\
Data File : VL038930.D
Acq On : 14 Apr 2022 23:25
Operator : SY/AP
Sample : N2259-04
Misc : 400mL/MSVOA L
ALS Vial : 9 Sample Multiplier: 1


Instrument :
MSVOA_L
ClientSampleId :
1165-19-SG-4

Quant Time: Apr 15 10:23:25 2022
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL040522
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
QLast Update : Wed Apr 06 06:12:53 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 04/15/2022
Supervised By :Mahesh Dadoda 04/18/2022





#1

Bromochloromethane

Concen: 10.000 ppbv

RT: 5.67 Instrument: MSVOA_1

Delta R.T. 0.00

Lab File: ClientSampleId: 1165-19-SG-4

Acq: 14 Apr 2022 23:25

Tot Ion:

Ion Ratio

49 100

128 46.





130 60.

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 04/15/2022

Supervised By :Mahesh Dadoda 04/18/2022



#2

Dichlorodifluoromethane

Concen: 0.545 ppbv

RT: 3.05 min Scan# 83

Delta R.T. 0.00 min

Lab File: VL038930.D




Acq: 14 Apr 2022 23:25


Tgt Ion: 85 Resp: 94236

Ion Ratio Lower Upper

85 100

87 31.3 25.9 38.9





#4

Chloromethane

Concen: 0.500 ppbv

RT: 3.14 Instrument: MSVOA_1

Delta R.T.

Lab File: ClientSampleId:

Acq: 14 Apr 2022 23:25

Tot Ion:

Ion Ratio

50 100





52 34.

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 04/15/2022

Supervised By :Mahesh Dadoda 04/18/2022



#9

Propene

Concen: 4.495 ppbv m

RT: 3.02 min Scan# 74

Delta R.T. 0.01 min

Lab File: VL038930.D




Acq: 14 Apr 2022 23:25


Tgt Ion: 41 Resp: 354099

Ion Ratio Lower Upper

41 100

42 34.7 54.2 81.4#









#10
 Heptane
 Concen: 18.962 ppbv
 RT: 8.12
 Instrument: MSVOA_L
 Delta R.T.:
 Lab File: ClientSampleId:
 Acq: 14 Apr 2022 23:25

Tot Ion:
 Ion Ratio
 43 100
 57 50.




Manual Integrations
 APPROVED


Reviewed By :Semsettin Yesilyurt 04/15/2022
 Supervised By :Mahesh Dadoda 04/18/2022



#11
 Trichlorofluoromethane
 Concen: 1.341 ppbv
 RT: 3.95 min Scan# 363
 Delta R.T.: 0.00 min
 Lab File: VL038930.D
 Acq: 14 Apr 2022 23:25

Tgt Ion:101 Resp: 186426
 Ion Ratio Lower Upper
 101 100
 103 61.4 51.4 77.0









#13
Ethanol
Concen: 187.555 ppbv
RT: 3.61
Instrument : MSVOA_1
Delta R.T. :
Lab File: ClientSampleId :
Acq: 14 Apr 2022 23:25

Tot Ion:
Ion Ratio
45 100
46 35.




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 04/15/2022
Supervised By :Mahesh Dadoda 04/18/2022



#15
Acetone
Concen: 229.499 ppbv
RT: 3.85 min Scan# 331
Delta R.T. 0.01 min
Lab File: VL038930.D
Acq: 14 Apr 2022 23:25

Tgt Ion: 43 Resp:27874327
Ion Ratio Lower Upper
43 100
58 45.4 22.5 33.7#









#19
Isopropyl Alcohol
Concen: 5.654 ppbv
RT: 3.97
Delta R.T. MSVOA.L
Lab File: ClientSampleId :
Acq: 14 Apr 2022 23:25

Tot Ion:
Ion Ratio
45 100
58 0.




Manual Integrations
APPROVED




Reviewed By :Semsettin Yesilyurt 04/15/2022
Supervised By :Mahesh Dadoda 04/18/2022



#20
Methylene Chloride
Concen: 8.334 ppbv
RT: 4.35 min Scan# 487
Delta R.T. 0.01 min
Lab File: VL038930.D
Acq: 14 Apr 2022 23:25

Tgt Ion: 84 Resp: 367116
Ion Ratio Lower Upper
84 100
49 158.2 129.4 194.0
51 46.1 37.4 56.2
86 66.3 51.4 77.2









#25
Ethyl Acetate
Concen: 1.787 ppbv
RT: 5.68
Instrument: MSVOA_1
Delta R.T.:
Lab File: ClientSampleId : 1165-19-SG-4
Acq: 14 Apr 2022 23:25

Tot Ion:
Ion Ratio
43 100
45 9.
61 8.
70 6.7 5.6 8.4


Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 04/15/2022
Supervised By :Mahesh Dadoda 04/18/2022



#26
Hexane
Concen: 2.222 ppbv
RT: 5.69 min Scan# 903
Delta R.T. 0.01 min
Lab File: VL038930.D
Acq: 14 Apr 2022 23:25

Tgt Ion: 57 Resp: 311762
Ion Ratio Lower Upper
57 100
41 91.8 72.0 108.0
43 186.4 186.6 280.0#
56 52.8 42.8 64.2









#30
Cvclohexane
Concen: 0.980 ppbv
RT: 7.13 min
Instrument: MSVOA_1
Delta R.T.:
Lab File: ClientSampleId :
Acq: 14 Apr 2022 23:25

Tot Ion:
Ion Ratio
84 100
56 139.
41 164.




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 04/15/2022
Supervised By :Mahesh Dadoda 04/18/2022



#33
1,4-Difluorobenzene
Concen: 10.000 ppbv
RT: 7.18 min Scan# 1367
Delta R.T.: 0.01 min
Lab File: VL038930.D
Acq: 14 Apr 2022 23:25

Tgt Ion:114 Resp: 3009538
Ion Ratio Lower Upper
114 100
63 25.7 19.8 29.8
88 18.1 14.2 21.4









#34
2-Butanone
Concen: 2.084 ppbv
RT: 5.25
Instrument : MSVOA_1
Delta R.T.:
Lab File: ClientSampleId :
Acq: 14 Apr 2022 23:25

Tot Ion:
Ion Ratio
43 100
72 25.




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 04/15/2022
Supervised By :Mahesh Dadoda 04/18/2022



#36
Benzene
Concen: 1.060 ppbv
RT: 6.89 min Scan# 1277
Delta R.T. 0.01 min
Lab File: VL038930.D
Acq: 14 Apr 2022 23:25

Tgt Ion: 78 Resp: 290333
Ion Ratio Lower Upper
78 100
77 23.0 18.2 27.2
50 16.8 14.0 21.0









#39
 1,2-Dichloropropane
 Concen: 7.193 ppbv
 RT: 7.18
 Instrument: MSVOA_1
 Delta R.T.:
 Lab File: ClientSampled:
 Acq: 14 Apr 2022 23:25

Tot Ion:
 Ion Ratio
 63 100
 41 0.
 62 16.




Manual Integrations
 APPROVED


Reviewed By :Semsettin Yesilyurt 04/15/2022
 Supervised By :Mahesh Dadoda 04/18/2022



#44
 2,2,4-Trimethylpentane
 Concen: 0.633 ppbv m
 RT: 7.87 min Scan# 1582
 Delta R.T. 0.01 min
 Lab File: VL038930.D
 Acq: 14 Apr 2022 23:25

Tgt Ion: 57 Resp: 293365
 Ion Ratio Lower Upper
 57 100
 41 109.6 25.0 37.4#
 56 139.6 25.3 37.9#









#50
 4-Methyl-2-Pentanone
 Concen: 4.044 ppbv
 RT: 8.74
 Delta R.T. MSVOA_L
 Lab File: ClientSampleId :
 1165-19-SG-4
 Acq: 14 Apr 2022 23:25

Tot Ion:
 Ion Ratio
 43 100
 58 37.




Manual Integrations
 APPROVED


Reviewed By :Semsettin Yesilyurt 04/15/2022
 Supervised By :Mahesh Dadoda 04/18/2022



#53
 Toluene
 Concen: 54.993 ppbv
 RT: 9.81 min Scan# 2185
 Delta R.T. 0.01 min
 Lab File: VL038930.D
 Acq: 14 Apr 2022 23:25

Tgt Ion: 91 Resp:17380569
 Ion Ratio Lower Upper
 91 100
 92 71.3 49.2 73.8





#55

Chlorobenzene-d5

Concen: 10.000 ppbv

RT: 12.09

Delta R.T. MSVOA_L

Lab File: ClientSampled :

Acq: 14 Apr 2022 23:25

Tot Ion:1

Ion Rati

117 100

82 65.


119 33.

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 04/15/2022




Supervised By :Mahesh Dadoda 04/18/2022



Abundance Ion 117.10 (116.80 to 117.80): V

1500000 Ion 82.10 (81.80 to 82.80): VL0

Ion 119.10 (118.80 to 119.80): V



#58

Ethyl Benzene

Concen: 2.451 ppbv

RT: 12.71 min Scan# 3085

Delta R.T. 0.00 min

Lab File: VL038930.D


Acq: 14 Apr 2022 23:25

Tgt Ion: 91 Resp: 965836

Ion Ratio Lower Upper



91 100


106 29.6 26.7 40.1



Abundance Ion 91.10 (90.80 to 91.80): VL0

500000 Ion 106.10 (105.80 to 106.80): V









#59
m/p-Xylene
Concen: 7.493 ppbv
RT: 12.96
Delta R.T. MSVOA_L
Lab File: ClientSampleId :
Acq: 14 Apr 2022 23:25

Tot Ion:
Ion Ratio
91 100
106 46.




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 04/15/2022
Supervised By :Mahesh Dadoda 04/18/2022



#60
o-Xylene
Concen: 3.191 ppbv
RT: 13.69 min Scan# 3390
Delta R.T. -0.00 min
Lab File: VL038930.D
Acq: 14 Apr 2022 23:25

Tgt Ion: 91 Resp: 988313
Ion Ratio Lower Upper
91 100
106 43.9 23.8 71.4









#61
Stvrene
Concen: 0.218 ppbv m
RT: 13.52
Instrument: MSVOA_1
Delta R.T.:
Lab File: ClientSampleId :
Acq: 14 Apr 2022 23:25

Tot Ion:1
Ion Rati
104 100
78 55.
103 0.




Manual Integrations
APPROVED




Reviewed By :Semsettin Yesilyurt 04/15/2022
Supervised By :Mahesh Dadoda 04/18/2022



#62
Isopropylbenzene
Concen: 0.286 ppbv
RT: 14.66 min Scan# 3693
Delta R.T. -0.00 min
Lab File: VL038930.D
Acq: 14 Apr 2022 23:25

Tgt Ion:105 Resp: 133111
Ion Ratio Lower Upper
105 100
120 24.8 21.8 32.8









#64
n-propylbenzene
Concen: 0.904 ppbv
RT: 15.56 min
Instrument: MSVOA_1
Delta R.T.: 0.02 min
Lab File: ClientSampleId : 1165-19-SG-4
Acq: 14 Apr 2022 23:25

Tot Ion:1
Ion Rati
120 100
91 515.


Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 04/15/2022
Supervised By :Mahesh Dadoda 04/18/2022



#65
tert-Butylbenzene
Concen: 0.447 ppbv
RT: 16.76 min Scan# 4345
Delta R.T.: 0.02 min
Lab File: VL038930.D
Acq: 14 Apr 2022 23:25

Tgt Ion:119 Resp: 184555
Ion Ratio Lower Upper
119 100
91 110.2 62.8 94.2#
134 0.2 16.2 24.4#





#67

sec-Butylbenzene

Concen: 0.241 ppbv

RT: 17.28

Delta R.T. MSVOA_L

Lab File: ClientSampleId :

Acq: 14 Apr 2022 23:25

Tot Ion:1

Ion Ratio

105 100





134 17.

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 04/15/2022

Supervised By :Mahesh Dadoda 04/18/2022



#68

1-Bromo-4-Fluorobenzene

Concen: 10.907 ppbv

RT: 14.41 min Scan# 3616

Delta R.T. 0.01 min

Lab File: VL038930.D

Acq: 14 Apr 2022 23:25




Tgt Ion: 95 Resp: 2181406


Ion Ratio Lower Upper

95 100

174 62.3 52.0 78.0

175 4.6 4.0 6.0









#69
 p-Isopropyltoluene
 Concen: 0.161 ppbv m
 RT: 17.64 min
 Instrument: MSVOA_1
 Delta R.T.:
 Lab File: ClientSampleId :
 1165-19-SG-4
 Acq: 14 Apr 2022 23:25

Tot Ion: 1
 Ion Ratio
 119 100
 134 46.
 91 54.




Manual Integrations
 APPROVED


Reviewed By : Semsettin Yesilyurt 04/15/2022
 Supervised By : Mahesh Dadoda 04/18/2022



#70
 n-Butylbenzene
 Concen: 0.613 ppbv
 RT: 18.55 min Scan# 4902
 Delta R.T.: 0.01 min
 Lab File: VL038930.D
 Acq: 14 Apr 2022 23:25

Tgt Ion: 91 Resp: 293441
 Ion Ratio Lower Upper
 91 100
 92 39.9 44.6 66.8#
 134 69.0 19.2 28.8#









#72
4-Ethyltoluene
Concen: 1.550 ppbv
RT: 15.83 min
Instrument: MSVOA_1
Delta R.T.:
Lab File: ClientSampleId : 1165-19-SG-4
Acq: 14 Apr 2022 23:25

Tot Ion: 1
Ion Ratio
105 100
120 28.
91 11.
77 12.2 10.6 15.8




Manual Integrations
APPROVED


Reviewed By : Semsettin Yesilyurt 04/15/2022
Supervised By : Mahesh Dadoda 04/18/2022



#73
1,3,5-Trimethylbenzene
Concen: 1.477 ppbv
RT: 15.99 min Scan# 4107
Delta R.T.: 0.00 min
Lab File: VL038930.D
Acq: 14 Apr 2022 23:25

Tgt Ion: 105 Resp: 490748
Ion Ratio Lower Upper
105 100
120 46.5 38.7 58.1









#74
 1,2,4-Trimethylbenzene
 Concen: 4.971 ppbv
 RT: 16.76
 Delta R.T. MSVOA_L
 Lab File: ClientSampleId :
 1165-19-SG-4
 Acq: 14 Apr 2022 23:25

Tot Ion:1
 Ion Ratio
 105 100
 120 42.
 91 12.
 77 14.3 17.3 25.9#




Manual Integrations
 APPROVED


Reviewed By :Semsettin Yesilyurt 04/15/2022
 Supervised By :Mahesh Dadoda 04/18/2022



#79
 Naphthalene
 Concen: 0.462 ppbv m
 RT: 22.18 min Scan# 6033
 Delta R.T. 0.00 min
 Lab File: VL038930.D
 Acq: 14 Apr 2022 23:25

Tgt Ion:128 Resp: 121444
 Ion Ratio Lower Upper
 128 100
 127 13.7 10.6 15.8
 129 14.4 9.4 14.0#





#80
 Naphthalene, 2-methyl-
 Concen: 0.511 ppbv
 RT: 24.97
 Instrument : MSVOA_1
 Delta R.T
 Lab File: ClientSampleId :
 1165-19-SG-4
 Acq: 14 Apr 2022 23:23

Tot Ion:1
 Ion Rati
 142 100
 141 82.
 115 104.

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 04/15/2022
 Supervised By :Mahesh Dadoda 04/18/2022

