

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL041422\
 Data File : VL038934.D
 Acq On : 15 Apr 2022 2:07
 Operator : SY/AP
 Sample : N2259-08
 Misc : 400mL/MSVOA L
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA L
 ClientSampleId :
 1165-19-IA-3

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 04/15/2022
 Supervised By : Mahesh Dadoda 04/18/2022

Quant Time: Apr 15 10:15:34 2022

Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL040522A

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:

QLast Update : Wed Apr 06 06:12:53 2022

Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.67	49	1190431	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.18	114	3104358	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.08	117	2754205	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.40	95	2190411	10.48	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	104.80%

Target Compounds

	R.T.	QIon	Response	Conc	Units	Ovalue
2) Dichlorodifluoromethane	3.05	85	103891	0.615	ppbv	89
4) Chloromethane	3.14	50	33948	0.451	ppbv	91
9) Propene	3.02	41	298437m	3.883	ppbv	
10) Heptane	8.12	43	3145471	16.246	ppbv	99
11) Trichlorofluoromethane	3.95	101	213592	1.574	ppbv	98
13) Ethanol	3.61	45	643487	137.755	ppbv #	100
15) Acetone	3.85	43	25963967	219.072	ppbv #	70
19) Isopropyl Alcohol	3.97	45	555443	9.170	ppbv #	100
20) Methylene Chloride	4.35	84	251991	5.863	ppbv	94
25) Ethyl Acetate	5.68	43	449977	1.496	ppbv	99
26) Hexane	5.69	57	201654	1.473	ppbv	99
30) Cyclohexane	7.13	84	119567	0.991	ppbv	99
34) 2-Butanone	5.25	43	238099	1.941	ppbv	92
35) Carbon Tetrachloride	7.01	117	10699m	0.057	ppbv	
36) Benzene	6.89	78	468124	1.658	ppbv	97
44) 2,2,4-Trimethylpentane	7.87	57	220651	0.461	ppbv #	1
50) 4-Methyl-2-Pentanone	8.75	43	175619	0.636	ppbv	95
52) Tetrachloroethene	11.22	164	15310m	0.155	ppbv	
53) Toluene	9.81	91	17294083	53.048	ppbv	87
58) Ethyl Benzene	12.70	91	1088129	2.644	ppbv	92
59) m/p-Xylene	12.96	91	2704819	7.823	ppbv	100
60) o-Xylene	13.68	91	1094716	3.384	ppbv	94
61) Styrene	13.52	104	48078m	0.244	ppbv	
62) Isopropylbenzene	14.66	105	151966m	0.313	ppbv	
64) n-propylbenzene	15.56	120	126371	1.002	ppbv	84
67) sec-Butylbenzene	17.28	105	143234	0.237	ppbv	96
69) p-Isopropyltoluene	17.64	119	71855	0.145	ppbv	96
70) n-Butylbenzene	18.55	91	306726	0.613	ppbv #	56
72) 4-Ethyltoluene	15.83	105	613211	1.587	ppbv	98
73) 1,3,5-Trimethylbenzene	15.99	105	542820	1.564	ppbv	97
74) 1,2,4-Trimethylbenzene	16.75	105	1884631	5.137	ppbv #	68
79) Naphthalene	22.18	128	162942m	0.594	ppbv	
80) Naphthalene,2-methyl-	24.96	142	49456	0.731	ppbv #	64

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL041422\
Data File : VL038934.D
Acq On : 15 Apr 2022 2:07
Operator : SY/AP
Sample : N2259-08
Misc : 400mL/MSVOA L
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
1165-19-IA-3

Manual Integrations
APPROVED

Quant Time: Apr 15 10:15:34 2022


Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL040522


Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:

QLast Update : Wed Apr 06 06:12:53 2022

Response via : Initial Calibration

Reviewed By :Semsettin Yesilyurt 04/15/2022
Supervised By :Mahesh Dadoda 04/18/2022





#1

Bromochloromethane

Concen: 10.000 ppbv

RT: 5.67 Instrument: MSVOA_1

Delta R.T. ClientSampleId:

Lab File: 1165-19-1A-3

Acq: 15 Apr 2022 2:07

Tot Ion:

Ion Ratio

49 100

128 48.





130 62.

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 04/15/2022

Supervised By :Mahesh Dadoda 04/18/2022



#2

Dichlorodifluoromethane

Concen: 0.615 ppbv

RT: 3.05 min Scan# 82

Delta R.T. -0.00 min

Lab File: VL038934.D




Acq: 15 Apr 2022 2:07


Tgt Ion: 85 Resp: 103891

Ion Ratio Lower Upper

85 100

87 26.5 25.9 38.9





#4

Chloromethane

Concen: 0.451 ppbv

RT: 3.14 Instrument:

Delta R.T. MSVOA_L

Lab File: ClientSampleId:

Acq: 15 Apr 2022 2:07

Tot Ion:

Ion Ratio

50 100





52 25.

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 04/15/2022

Supervised By :Mahesh Dadoda 04/18/2022



#9

Propene

Concen: 3.883 ppbv m

RT: 3.02 min Scan# 73

Delta R.T. 0.01 min

Lab File: VL038934.D




Acq: 15 Apr 2022 2:07


Tgt Ion: 41 Resp: 298437

Ion Ratio Lower Upper

41 100

42 39.7 54.2 81.4#









#10
 Heptane
 Concen: 16.246 ppbv
 RT: 8.12
 Instrument: MSVOA_1
 Delta R.T.:
 Lab File: ClientSampleId:
 Acq: 15 Apr 2022 2:07

Tot Ion:
 Ion Ratio
 43 100
 57 49.




Manual Integrations
 APPROVED


Reviewed By :Semsettin Yesilyurt 04/15/2022
 Supervised By :Mahesh Dadoda 04/18/2022



#11
 Trichlorofluoromethane
 Concen: 1.574 ppbv
 RT: 3.95 min Scan# 362
 Delta R.T. -0.00 min
 Lab File: VL038934.D
 Acq: 15 Apr 2022 2:07

Tgt Ion:101 Resp: 213592
 Ion Ratio Lower Upper
 101 100
 103 63.0 51.4 77.0









#13
Ethanol
Concen: 137.755 ppbv
RT: 3.61
Instrument : MSVOA_1
Delta R.T. :
Lab File: ClientSampleId :
Acq: 15 Apr 2022 2:07

Tot Ion:
Ion Ratio
45 100
46 37.




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 04/15/2022
Supervised By :Mahesh Dadoda 04/18/2022



#15
Acetone
Concen: 219.072 ppbv
RT: 3.85 min Scan# 330
Delta R.T. 0.00 min
Lab File: VL038934.D
Acq: 15 Apr 2022 2:07

Tgt Ion: 43 Resp:25963967
Ion Ratio Lower Upper
43 100
58 44.0 22.5 33.7#





#19

Isopropyl Alcohol

Concen: 9.170 ppbv

RT: 3.97 Instrument:

Delta R.T. MSVOA_1

Lab File: ClientSampleId:

Acq: 15 Apr 2022 2:07

Tot Ion:

Ion Ratio

45 100





58 0.

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 04/15/2022

Supervised By :Mahesh Dadoda 04/18/2022



#20

Methylene Chloride

Concen: 5.863 ppbv

RT: 4.35 min Scan# 486

Delta R.T. 0.01 min

Lab File: VL038934.D

Acq: 15 Apr 2022 2:07

Tgt Ion: 84 Resp: 251991




Ion Ratio Lower Upper




84 100

49 149.7 129.4 194.0

51 47.8 37.4 56.2

86 65.7 51.4 77.2









#25
Ethyl Acetate
Concen: 1.496 ppbv
RT: 5.68
Instrument: MSVOA_1
Delta R.T.:
Lab File: ClientSampleId : 1165-19-IA-3
Acq: 15 Apr 2022 2:07

Tot Ion:
Ion Ratio
43 100
45 11.
61 9.
70 7.6 5.6 8.4


Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 04/15/2022
Supervised By :Mahesh Dadoda 04/18/2022



#26
Hexane
Concen: 1.473 ppbv
RT: 5.69 min Scan# 902
Delta R.T.: 0.00 min
Lab File: VL038934.D
Acq: 15 Apr 2022 2:07

Tgt Ion: 57 Resp: 201654
Ion Ratio Lower Upper
57 100
41 90.9 72.0 108.0
43 232.4 186.6 280.0
56 56.4 42.8 64.2









#30
Cvclohexane
Concen: 0.991 ppbv
RT: 7.13 min
Instrument: MSVOA_1
Delta R.T.:
Lab File: ClientSampleId:
Acq: 15 Apr 2022 2:07

Tot Ion:
Ion Ratio
84 100
56 141.
41 85.




Manual Integrations
APPROVED




Reviewed By :Semsettin Yesilyurt 04/15/2022
Supervised By :Mahesh Dadoda 04/18/2022



#33
1,4-Difluorobenzene
Concen: 10.000 ppbv
RT: 7.18 min Scan# 1365
Delta R.T.: 0.00 min
Lab File: VL038934.D
Acq: 15 Apr 2022 2:07

Tgt Ion:114 Resp: 3104358
Ion Ratio Lower Upper
114 100
63 24.9 19.8 29.8
88 17.8 14.2 21.4









#34
2-Butanone
Concen: 1.941 ppbv
RT: 5.25
Instrument: MSVOA_1
Delta R.T.:
Lab File: ClientSampleId :
Acq: 15 Apr 2022 2:07

Tot Ion:
Ion Ratio
43 100
72 25.


Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 04/15/2022
Supervised By :Mahesh Dadoda 04/18/2022



#35
Carbon Tetrachloride
Concen: 0.057 ppbv m
RT: 7.01 min Scan# 1313
Delta R.T. -0.01 min
Lab File: VL038934.D
Acq: 15 Apr 2022 2:07

Tgt Ion:117 Resp: 10699
Ion Ratio Lower Upper
117 100
119 60.5 77.7 116.5#
121 18.6 24.5 36.7#









#36
Benzene
Concen: 1.658 ppbv
RT: 6.89
Delta R.T. MSVOA_L
Lab File: ClientSampleId :
Acq: 15 Apr 2022 2:07

Tot Ion:
Ion Ratio
78 100
77 22.
50 15.




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 04/15/2022
Supervised By :Mahesh Dadoda 04/18/2022



#44
2,2,4-Trimethylpentane
Concen: 0.461 ppbv
RT: 7.87 min Scan# 1580
Delta R.T. -0.00 min
Lab File: VL038934.D
Acq: 15 Apr 2022 2:07

Tgt Ion: 57 Resp: 220651
Ion Ratio Lower Upper
57 100
41 120.2 25.0 37.4#
56 154.2 25.3 37.9#









#50
4-Methyl-2-Pentanone
Concen: 0.636 ppbv
RT: 8.75
Delta R.T. MSVOA_L
Lab File: ClientSampleId :
Acq: 15 Apr 2022 2:07

Tot Ion:
Ion Ratio
43 100
58 39.




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 04/15/2022
Supervised By :Mahesh Dadoda 04/18/2022



#52
Tetrachloroethene
Concen: 0.155 ppbv m
RT: 11.22 min Scan# 2624
Delta R.T. 0.01 min
Lab File: VL038934.D
Acq: 15 Apr 2022 2:07

Tgt Ion:164 Resp: 15310
Ion Ratio Lower Upper
164 100
166 116.4 101.0 151.6
129 79.4 88.7 133.1#
131 80.7 81.8 122.6#





#53

Toluene

Concen: 53.048 ppbv

RT: 9.81 Instrument: MSVOA_1

Delta R.T. -0.00 min

Lab File: ClientSampled: 1165-19-1A-3

Acq: 15 Apr 2022 2:07

Tot Ion:

Ion Ratio

91 100





92 71.

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 04/15/2022

Supervised By :Mahesh Dadoda 04/18/2022



#55

Chlorobenzene-d5

Concen: 10.000 ppbv

RT: 12.08 min Scan# 2890

Delta R.T. -0.00 min

Lab File: VL038934.D

Acq: 15 Apr 2022 2:07




Tgt Ion:117 Resp: 2754205


Ion Ratio Lower Upper

117 100

82 66.1 53.0 79.6

119 33.4 24.9 37.3









#58
Ethyl Benzene
Concen: 2.644 ppbv
RT: 12.70 min
Instrument: MSVOA_1
Delta R.T.:
Lab File: ClientSampleId :
Acq: 15 Apr 2022 2:07

Tot Ion:
Ion Ratio
91 100
106 28.




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 04/15/2022
Supervised By :Mahesh Dadoda 04/18/2022



#59
m/p-Xylene
Concen: 7.823 ppbv
RT: 12.96 min Scan# 3163
Delta R.T.: -0.03 min
Lab File: VL038934.D
Acq: 15 Apr 2022 2:07

Tgt Ion: 91 Resp: 2704819
Ion Ratio Lower Upper
91 100
106 46.2 37.1 55.7









#60
o-Xylene
Concen: 3.384 ppbv
RT: 13.68
Delta R.T.: MSVOA_L
Lab File: ClientSampled :
Acq: 15 Apr 2022 2:07

Tot Ion:
Ion Ratio
91 100
106 43.




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 04/15/2022
Supervised By :Mahesh Dadoda 04/18/2022



#61
Styrene
Concen: 0.244 ppbv m
RT: 13.52 min Scan# 3339
Delta R.T.: 0.00 min
Lab File: VL038934.D
Acq: 15 Apr 2022 2:07

Tgt Ion:104 Resp: 48078
Ion Ratio Lower Upper
104 100
78 53.6 40.5 60.7
103 49.9 39.0 58.4





#62

Isopropylbenzene

Concen: 0.313 ppbv m

RT: 14.66

Delta R.T. MSVOA_L

Lab File: ClientSampleId :

Acq: 15 Apr 2022 2:07

Tot Ion:1

Ion Ratio

105 100





120 24.

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 04/15/2022

Supervised By :Mahesh Dadoda 04/18/2022



#64

n-propylbenzene

Concen: 1.002 ppbv

RT: 15.56 min Scan# 3972

Delta R.T. 0.01 min

Lab File: VL038934.D




Acq: 15 Apr 2022 2:07


Tgt Ion:120 Resp: 126371

Ion Ratio Lower Upper

120 100

91 491.5 360.6 540.8





#67

sec-Butylbenzene

Concen: 0.237 ppbv

RT: 17.28

Delta R.T. MSVOA_L

Lab File:

Acq: 15 Apr 2022 2:07

ClientSampleId :

1165-19-IA-3

Tot Ion:1

Ion Ratio

105 100





134 17.

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 04/15/2022

Supervised By :Mahesh Dadoda 04/18/2022



#68

1-Bromo-4-Fluorobenzene

Concen: 10.485 ppbv

RT: 14.40 min Scan# 3613

Delta R.T. -0.00 min

Lab File: VL038934.D

Acq: 15 Apr 2022 2:07




Tgt Ion: 95 Resp: 2190411


Ion Ratio Lower Upper

95 100

174 62.0 52.0 78.0

175 4.6 4.0 6.0









#69
 p-Isopropyltoluene
 Concen: 0.145 ppbv
 RT: 17.64 min
 Delta R.T. 0.00 min
 Lab File: VL038812.D
 Acq: 15 Apr 2022 2:07

Manual Integrations
 APPROVED




Reviewed By :Semsettin Yesilyurt 04/15/2022
 Supervised By :Mahesh Dadoda 04/18/2022


Tot Ion:1
 Ion Ratio
 119 100
 134 28.
 91 30.



#70
 n-Butylbenzene
 Concen: 0.613 ppbv
 RT: 18.55 min Scan# 4901
 Delta R.T. 0.00 min
 Lab File: VL038934.D
 Acq: 15 Apr 2022 2:07

Tgt Ion: 91 Resp: 306726
 Ion Ratio Lower Upper
 91 100
 92 39.3 44.6 66.8#
 134 70.8 19.2 28.8#





#72

4-Ethyltoluene

Concen: 1.587 ppbv

RT: 15.83

Delta R.T. MSVOA_L





Lab File: ClientSampleId :

Acq: 15 Apr 2022 2:07

Tot Ion:1
Ion Rati
105 100
120 30.
91 13.
77 14.1 10.6 15.8

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 04/15/2022
Supervised By :Mahesh Dadoda 04/18/2022



#73

1,3,5-Trimethylbenzene

Concen: 1.564 ppbv




RT: 15.99 min Scan# 4106


Delta R.T. -0.00 min

Lab File: VL038934.D

Acq: 15 Apr 2022 2:07

Tgt Ion:105 Resp: 542820
Ion Ratio Lower Upper
105 100
120 46.2 38.7 58.1









#74
1,2,4-Trimethylbenzene
Concen: 5.137 ppbv
RT: 16.75
Delta R.T.:
Lab File: 1165-19-IA-3
Acq: 15 Apr 2022 2:07

Manual Integrations
APPROVED




Reviewed By :Semsettin Yesilyurt 04/15/2022
Supervised By :Mahesh Dadoda 04/18/2022


Tot Ion:1
Ion Ratio
105 100
120 39.
91 10.
77 13.5 17.3 25.9#



#79
Naphthalene
Concen: 0.594 ppbv m
RT: 22.18 min Scan# 6033
Delta R.T.: 0.00 min
Lab File: VL038934.D
Acq: 15 Apr 2022 2:07

Tgt Ion:128 Resp: 162942
Ion Ratio Lower Upper
128 100
127 15.6 10.6 15.8
129 14.8 9.4 14.0#





#80
 Naphthalene, 2-methyl-
 Concen: 0.731 ppbv
 RT: 24.96
 Delta R.T: MSVOA_L
 Lab File: ClientSampleId :
 1165-19-IA-3
 Acq: 15 Apr 2022 2:07

Tot Ion:1
 Ion Rati
 142 100
 141 95.
 115 96.

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 04/15/2022
 Supervised By :Mahesh Dadoda 04/18/2022

