

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL041719\
 Data File : VL033607.D
 Acq On : 17 Apr 2019 18:09
 Operator : SY/AP
 Sample : K2241-15 0.03 PPB
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 MDL-MDL-AIR-03-QT2-2019

Quant Time: Apr 18 06:35:20 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL032519AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Mar 26 16:00:50 2019
 QLast Update : Tue Mar 26 16:00:50 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.76	49	719621	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.29	114	1874783	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.24	117	1957509	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.57	95	1558767	10.16	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.60%

Target Compounds					Ovalue
5) Vinyl Chloride	3.29	62	2180m	0.051	ppbv
32) 1,1,1-Trichloroethane	6.62	97	4869m	0.042	ppbv
35) Carbon Tetrachloride	7.14	117	5538m	0.042	ppbv
38) Trichloroethene	7.95	130	1941m	0.033	ppbv
63) 1,1,2,2-Tetrachloroethane	13.82	83	5424m	0.033	ppbv

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL041719\
Data File : VL033607.D
Acq On : 17 Apr 2019 18:09
Operator : SY/AP
Sample : K2241-15 0.03 PPB
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
MDL-MDL-AIR-03-QT2-2019

Quant Time: Apr 18 06:35:20 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL032519AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Mar 26 16:00:50 2019
QLast Update : Tue Mar 26 16:00:50 2019
Response via : Initial Calibration

