

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL041719\
 Data File : VL033602.D
 Acq On : 17 Apr 2019 14:59
 Operator : SY/AP
 Sample : VL0417ABSD01
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0417ABSD01

Quant Time: Apr 18 05:57:45 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL032519AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Mar 26 16:00:50 2019
 QLast Update : Tue Mar 26 16:00:50 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.76	49	852486	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.29	114	2169687	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.24	117	2298507m	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.58	95	1735160	9.63	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	96.30%

Target Compounds					Ovalue
2) Dichlorodifluoromethane	3.08	85	910949	7.327	ppbv 99
3) Chlorodifluoromethane	3.01	51	892723	9.096	ppbv 99
4) Chloromethane	3.17	50	496323	9.557	ppbv 94
5) Vinyl Chloride	3.29	62	489574	9.595	ppbv 100
6) Bromomethane	3.52	94	316223	9.643	ppbv 92
7) Chloroethane	3.60	64	180397	9.499	ppbv 100
8) Dichlorotetrafluoroethane	3.22	85	1201434	9.069	ppbv 99
9) Propene	3.04	41	398626	9.672	ppbv 99
10) Heptane	8.25	43	1000522	9.082	ppbv 99
11) Trichlorofluoromethane	4.00	101	1404573	9.166	ppbv 100
12) 1,1,2-Trichlorotrifluoroet	4.55	101	958593	9.314	ppbv 99
13) Ethanol	3.65	45	118311	11.093	ppbv 67
14) Bromoethene	3.79	108	397487	9.683	ppbv 99
15) Acetone	3.90	43	904583	8.825	ppbv 98
16) 1,3-Butadiene	3.36	39	399185	9.818	ppbv 90
17) tert-Butyl alcohol	4.35	59	199534	10.587	ppbv 99
18) 1,1-Dichloroethene	4.35	96	427853	9.678	ppbv 98
19) Isopropyl Alcohol	4.02	45	697685	10.761	ppbv 100
20) Methylene Chloride	4.41	84	383433	8.812	ppbv 93
21) Allyl Chloride	4.48	41	628850	9.705	ppbv 99
22) trans-1,2-Dichloroethene	4.96	96	445408	9.657	ppbv 95
23) Vinyl Acetate	5.16	43	1433449	9.825	ppbv 99
24) 1,1-Dichloroethane	5.09	63	999150	9.016	ppbv 100
25) Ethyl Acetate	5.77	43	1669523	8.779	ppbv 99
26) Hexane	5.78	57	765991	9.059	ppbv 96
27) Carbon Disulfide	4.62	76	1072429	10.665	ppbv 99
28) Methyl tert-Butyl Ether	5.11	73	1216017	9.984	ppbv 100
29) Chloroform	5.85	83	1266586	8.738	ppbv 97
30) Cyclohexane	7.25	84	665878	9.954	ppbv 98
31) cis-1,2-Dichloroethene	5.64	61	777254	9.245	ppbv 96
32) 1,1,1-Trichloroethane	6.62	97	1297669	9.401	ppbv 99
34) 2-Butanone	5.33	43	1113643	8.543	ppbv 99
35) Carbon Tetrachloride	7.14	117	1397830	9.096	ppbv 97
36) Benzene	7.01	78	1655565	9.094	ppbv 98
37) 1,2-Dichloroethane	6.42	62	980012	8.031	ppbv 99
38) Trichloroethene	7.96	130	657909	9.603	ppbv 98
39) 1,2-Dichloropropane	7.74	63	608243	8.529	ppbv 97
40) 1,4-Dioxane	7.95	88	239496	9.416	ppbv 87
41) Tetrahydrofuran	6.15	42	639234	8.915	ppbv 97
42) Bromodichloromethane	7.92	83	1429174	8.599	ppbv 98
43) Methyl Methacrylate	8.14	69	656427	8.880	ppbv 100

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL041719\
 Data File : VL033602.D
 Acq On : 17 Apr 2019 14:59
 Operator : SY/AP
 Sample : VL0417ABSD01
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0417ABSD01

Quant Time: Apr 18 05:57:45 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL032519AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Mar 26 16:00:50 2019
 QLast Update : Tue Mar 26 16:00:50 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)

44) 2,2,4-Trimethylpentane	7.99	57	2727097	8.572	ppbv	99
45) t-1,3-Dichloropropene	9.42	75	915584	10.088	ppbv	100
46) cis-1,3-Dichloropropene	8.83	75	1029980	9.718	ppbv	97
47) 1,1,2-Trichloroethane	9.62	97	693449	8.563	ppbv	97
48) Dibromochloromethane	10.45	129	1254957	9.188	ppbv	100
49) Bromoform	13.20	173	1164613	8.925	ppbv	99
50) 4-Methyl-2-Pentanone	8.87	43	1652537	8.658	ppbv	98
51) 2-Hexanone	10.27	43	1305722	9.170	ppbv #	100
52) Tetrachloroethene	11.37	164	633032	9.216	ppbv	99
53) Toluene	9.95	91	1951355	9.224	ppbv	99
54) 1,2-Dibromoethane	10.76	107	1117265	8.821	ppbv	97
56) 1,1,1,2-Tetrachloroethane	12.28	131	867581	8.236	ppbv #	56
57) Chlorobenzene	12.30	112	1473035	8.253	ppbv	97
58) Ethyl Benzene	12.86	91	2703209	8.807	ppbv	99
59) m/p-Xylene	13.15	91	4651000m	16.964	ppbv	
60) o-Xylene	13.85	91	2288526	8.632	ppbv	99
61) Styrene	13.69	104	1709298	9.394	ppbv	99
62) Isopropylbenzene	14.83	105	3212604	8.557	ppbv	100
63) 1,1,2,2-Tetrachloroethane	13.84	83	1584804	8.118	ppbv	99
64) n-propylbenzene	15.72	120	827536	8.687	ppbv	98
65) tert-Butylbenzene	16.91	119	2909349	8.794	ppbv	98
66) Benzyl Chloride	17.15	91	1658916	10.784	ppbv	99
67) sec-Butylbenzene	17.46	105	4100611	8.716	ppbv	100
69) p-Isopropyltoluene	17.82	119	3421480	8.932	ppbv	99
70) n-Butylbenzene	18.71	91	3534796	8.709	ppbv	99
71) 2-Chlorotoluene	15.61	91	2483906	8.701	ppbv	99
72) 4-Ethyltoluene	16.00	105	2663147	8.936	ppbv	100
73) 1,3,5-Trimethylbenzene	16.15	105	2554857	8.959	ppbv	99
74) 1,2,4-Trimethylbenzene	16.93	105	2656672	8.691	ppbv	96
75) 1,3-Dichlorobenzene	17.16	146	1584098	8.385	ppbv	99
76) 1,4-Dichlorobenzene	17.31	146	1585061	8.611	ppbv	100
77) 1,2-Dichlorobenzene	17.99	146	1555763	8.679	ppbv	100
78) Hexachloro-1,3-Butadiene	23.54	225	912343	7.689	ppbv	99
79) Naphthalene	22.41	128	2623870	10.028	ppbv	98
80) Naphthalene, 2-methyl-	25.11	142	1152179	13.121	ppbv	98
81) 1,2,4-Trichlorobenzene	22.14	180	1390903	9.583	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

Bromochloromethane

Concen: 10.000 ppbv

RT: 5.76 min Scan# 879

Delta R.T. -0.00 min

Lab File: VL033602.D

Acq: 17 Apr 2019 14:59

Instrument :

MSVOA_L

ClientSampleId :

VL0417ABSD01


Tot Ion: 49 Resp: 852486

Ion Ratio Lower Upper

49 100

128 50.0 24.0 72.0




130 64.9 31.2 93.6



Abundance Ion 49.10 (48.80 to 49.80): VL0

Ion 128.00 (127.70 to 128.70): V

Ion 130.00 (129.70 to 130.70): V



#2

Dichlorodifluoromethane

Concen: 7.327 ppbv

RT: 3.08 min Scan# 44

Delta R.T. 0.00 min

Lab File: VL033602.D


Acq: 17 Apr 2019 14:59

Tgt Ion: 85 Resp: 910949

Ion Ratio Lower Upper



85 100

87 32.7 26.7 40.1



Abundance Ion 85.00 (84.70 to 85.70): VL0

Ion 87.00 (86.70 to 87.70): VL0









#3
Chlorodifluoromethane
Concen: 9.096 ppbv
RT: 3.01 min Scan# 24
Delta R.T. -0.00 min
Lab File: VL033602.D
Acq: 17 Apr 2019 14:59




Instrument :
MSVOA_L
ClientSampleId :
VL0417ABSD01


Tot Ion: 51 Resp: 892723
Ion Ratio Lower Upper
51 100
67 18.9 15.7 23.5



#4
Chloromethane
Concen: 9.557 ppbv
RT: 3.17 min Scan# 74
Delta R.T. 0.01 min
Lab File: VL033602.D
Acq: 17 Apr 2019 14:59

Tgt Ion: 50 Resp: 496323
Ion Ratio Lower Upper
50 100
52 28.2 25.0 37.6





#5

Vinyl Chloride

Concen: 9.595 ppbv

RT: 3.29 min Scan# 110

Delta R.T. -0.00 min

Lab File: VL033602.D

Acq: 17 Apr 2019 14:59

Instrument :

MSVOA_L

ClientSampleId :


VL0417ABSD01

Tot Ion: 62 Resp: 489574

Ion Ratio Lower Upper




62 100

64 31.1 24.9 37.3



Abundance Ion 62.10 (61.80 to 62.80): VL0

Ion 64.10 (63.80 to 64.80): VL0



#6

Bromomethane

Concen: 9.643 ppbv

RT: 3.52 min Scan# 180

Delta R.T. 0.00 min

Lab File: VL033602.D


Acq: 17 Apr 2019 14:59

Tgt Ion: 94 Resp: 316223

Ion Ratio Lower Upper



94 100


96 94.7 70.2 105.2



Abundance Ion 94.00 (93.70 to 94.70): VL0

Ion 96.00 (95.70 to 96.70): VL0





#7

Chloroethane

Concen: 9.499 ppbv

RT: 3.60 min Scan# 207

Delta R.T. -0.00 min

Lab File: VL033602.D

Acq: 17 Apr 2019 14:59

Instrument :

MSVOA_L

ClientSampleId :


VL0417ABSD01

Tot Ion: 64 Resp: 180397

Ion Ratio Lower Upper

64 100

66 33.3 26.7 40.1



Abundance Ion 64.10 (63.80 to 64.80): VL0

Ion 66.10 (65.80 to 66.80): VL0


150000

100000


50000

0

Time-->



Time-->



#8

Dichlorotetrafluoroethane

Concen: 9.069 ppbv

RT: 3.22 min Scan# 88

Delta R.T. -0.00 min

Lab File: VL033602.D


Acq: 17 Apr 2019 14:59

Tgt Ion: 85 Resp: 1201434

Ion Ratio Lower Upper

85 100

135 68.7 54.0 81.0



Abundance Ion 85.00 (84.70 to 85.70): VL0

Ion 135.00 (134.70 to 135.70): V

1000000

800000


600000

400000


200000

0

Time-->



Time-->



#9

Propene

Concen: 9.672 ppbv

RT: 3.04 min Scan# 31

Delta R.T. 0.00 min

Lab File: VL033602.D

Acq: 17 Apr 2019 14:59

Instrument :

MSVOA_L

ClientSampleId :





VL0417ABSD01

Tot Ion: 41 Resp: 398626

Ion Ratio Lower Upper

41 100

42 67.9 53.9 80.9



#10

Heptane

Concen: 9.082 ppbv

RT: 8.25 min Scan# 1652

Delta R.T. 0.00 min

Lab File: VL033602.D




Acq: 17 Apr 2019 14:59


Tgt Ion: 43 Resp: 1000522

Ion Ratio Lower Upper

43 100

57 51.1 40.6 60.8









#11
 Trichlorofluoromethane
 Concen: 9.166 ppbv
 RT: 4.00 min Scan# 332
 Delta R.T. -0.00 min
 Lab File: VL033602.D
 Acq: 17 Apr 2019 14:59




Instrument :
 MSVOA_L
 ClientSampleId :
 VL0417ABSD01


Tot Ion:101 Resp: 1404573
 Ion Ratio Lower Upper
 101 100
 103 64.1 51.3 76.9



#12
 1,1,2-Trichlorotrifluoroethane
 Concen: 9.314 ppbv
 RT: 4.55 min Scan# 502
 Delta R.T. -0.00 min
 Lab File: VL033602.D
 Acq: 17 Apr 2019 14:59

Tgt Ion:101 Resp: 958593
 Ion Ratio Lower Upper
 101 100
 151 73.2 57.7 86.5





#13

Ethanol

Concen: 11.093 ppbv

RT: 3.65 min Scan# 221

Delta R.T. -0.00 min

Lab File: VL033602.D

Acq: 17 Apr 2019 14:59

Instrument :

MSVOA_L

ClientSampleId :





VL0417ABSD01

Tot Ion: 45 Resp: 118311

Ion Ratio Lower Upper

45 100

46 37.6 25.4 101.8



#14

Bromoethene

Concen: 9.683 ppbv

RT: 3.79 min Scan# 265

Delta R.T. 0.00 min

Lab File: VL033602.D

Acq: 17 Apr 2019 14:59




Tgt Ion: 108 Resp: 397487


Ion Ratio Lower Upper

108 100

106 107.8 86.4 129.6

79 17.8 15.8 23.8





#15

Acetone

Concen: 8.825 ppbv

RT: 3.90 min Scan# 300

Delta R.T. 0.00 min

Lab File: VL033602.D

Acq: 17 Apr 2019 14:59

Instrument :

MSVOA_L

ClientSampleId :


VL0417ABSD01

Tot Ion: 43 Resp: 904583

Ion Ratio Lower Upper




43 100

58 26.7 20.5 30.7



Abundance Ion 43.10 (42.80 to 43.80): VL0

Ion 58.10 (57.80 to 58.80): VL0



#16

1,3-Butadiene

Concen: 9.818 ppbv

RT: 3.36 min Scan# 133

Delta R.T. 0.00 min

Lab File: VL033602.D


Acq: 17 Apr 2019 14:59

Tgt Ion: 39 Resp: 399185

Ion Ratio Lower Upper



39 100


54 95.9 69.4 104.2



Abundance Ion 39.10 (38.80 to 39.80): VL0

Ion 54.10 (53.80 to 54.80): VL0





#17

tert-Butyl alcohol

Concen: 10.587 ppbv

RT: 4.35 min Scan# 441

Delta R.T. -0.00 min

Lab File: VL033602.D

Acq: 17 Apr 2019 14:59

Instrument :

MSVOA_L

ClientSampleID :





VL0417ABSD01

Tot Ion: 59 Resp: 199534

Ion Ratio Lower Upper

59 100

57 9.5 7.8 11.8



#18

1,1-Dichloroethene

Concen: 9.678 ppbv

RT: 4.35 min Scan# 440

Delta R.T. -0.00 min

Lab File: VL033602.D

Acq: 17 Apr 2019 14:59




Tgt Ion: 96 Resp: 427853


Ion Ratio Lower Upper

96 100

61 186.0 149.8 224.6

98 64.2 48.7 73.1









#19

Isopropyl Alcohol
 Concen: 10.761 ppbv
 RT: 4.02 min Scan# 338
 Delta R.T. 0.00 min
 Lab File: VL033602.D
 Acq: 17 Apr 2019 14:59

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0417ABSD01




Tot Ion: 45 Resp: 697685
 Ion Ratio Lower Upper
 45 100
 58 1.8 1.4 2.2




#20

Methylene Chloride
 Concen: 8.812 ppbv
 RT: 4.41 min Scan# 458
 Delta R.T. 0.00 min
 Lab File: VL033602.D
 Acq: 17 Apr 2019 14:59

Tgt Ion: 84 Resp: 383433
 Ion Ratio Lower Upper
 84 100
 49 138.2 120.0 180.0
 51 42.1 36.7 55.1
 86 64.4 51.9 77.9









#21
Allvl Chloride
Concen: 9.705 ppbv
RT: 4.48 min Scan# 479
Delta R.T. -0.00 min
Lab File: VL033602.D
Acq: 17 Apr 2019 14:59




Instrument :
MSVOA_L
ClientSampleId :
VL0417ABSD01


Tot Ion: 41 Resp: 628850
Ion Ratio Lower Upper
41 100
76 32.6 25.4 38.2
39 80.4 65.3 97.9



#22
trans-1,2-Dichloroethene
Concen: 9.657 ppbv
RT: 4.96 min Scan# 630
Delta R.T. -0.00 min
Lab File: VL033602.D
Acq: 17 Apr 2019 14:59

Tgt Ion: 96 Resp: 445408
Ion Ratio Lower Upper
96 100
61 153.9 128.9 193.3
98 59.7 49.5 74.3





#23

Vinyl Acetate

Concen: 9.825 ppbv

RT: 5.16 min Scan# 690

Delta R.T. 0.00 min

Lab File: VL033602.D

Acq: 17 Apr 2019 14:59

Instrument :

MSVOA_L

ClientSampleId :

VL0417ABSD01

Tot Ion: 43 Resp: 1433449


Ion Ratio Lower Upper

43 100

86 7.4 5.6 8.4

42 9.8 8.0 12.0

44 5.1 3.9 5.9






Abundance Ion 43.10 (42.80 to 43.80): VL0

Ion 86.10 (85.80 to 86.80): VL0

Ion 42.10 (41.80 to 42.80): VL0

Ion 44.10 (43.80 to 44.80): VL0



#24

1,1-Dichloroethane

Concen: 9.016 ppbv

RT: 5.09 min Scan# 670

Delta R.T. -0.00 min

Lab File: VL033602.D

Acq: 17 Apr 2019 14:59


Tgt Ion: 63 Resp: 999150

Ion Ratio Lower Upper

63 100

98 4.4 2.2 6.6



100 3.1 1.5 4.4




Abundance Ion 63.10 (62.80 to 63.80): VL0

Ion 98.00 (97.70 to 98.70): VL0

Ion 100.00 (99.70 to 100.70): VL0





#25

Ethyl Acetate

Concen: 8.779 ppbv

RT: 5.77 min Scan# 880

Delta R.T. -0.00 min


Lab File: VL033602.D

Acq: 17 Apr 2019 14:59

Instrument :
MSVOA_L
ClientSampleId :
VL0417ABSD01

Tot Ion: 43 Resp: 1669523

Ion	Ratio	Lower	Upper
43	100		
45	9.9	7.8	11.6
61	9.5	7.3	10.9
70	7.3	5.6	8.4






Abundance

Ion 43.10 (42.80 to 43.80): VL0

Ion 45.10 (44.80 to 45.80): VL0

Ion 61.10 (60.80 to 61.80): VL0

Ion 70.10 (69.80 to 70.80): VL0



#26

Hexane

Concen: 9.059 ppbv

RT: 5.78 min Scan# 884


Delta R.T. -0.00 min

Lab File: VL033602.D

Acq: 17 Apr 2019 14:59

Tgt Ion: 57 Resp: 765991

Ion	Ratio	Lower	Upper
57	100		
41	86.3	70.7	106.1
43	218.8	182.6	274.0
56	52.3	42.0	63.0





Abundance


Ion 57.10 (56.80 to 57.80): VL0

Ion 41.10 (40.80 to 41.80): VL0

Ion 43.10 (42.80 to 43.80): VL0

Ion 56.10 (55.80 to 56.80): VL0





#27

Carbon Disulfide

Concen: 10.665 ppbv

RT: 4.62 min Scan# 523

Delta R.T. -0.00 min

Lab File: VL033602.D

Acq: 17 Apr 2019 14:59

Instrument :

MSVOA_L

ClientSampleId :

VL0417ABSD01


Tot Ion: 76 Resp: 1072429

Ion Ratio Lower Upper

76 100

44 17.8 14.8 22.2




78 9.6 7.6 11.4



Abundance Ion 76.00 (75.70 to 76.70): VL0

Ion 44.10 (43.80 to 44.80): VL0

Ion 78.00 (77.70 to 78.70): VL0



#28

Methyl tert-Butyl Ether

Concen: 9.984 ppbv

RT: 5.11 min Scan# 677

Delta R.T. 0.00 min

Lab File: VL033602.D


Acq: 17 Apr 2019 14:59

Tgt Ion: 73 Resp: 1216017

Ion Ratio Lower Upper



73 100


57 21.7 17.4 26.2



Abundance Ion 73.10 (72.80 to 73.80): VL0

Ion 57.10 (56.80 to 57.80): VL0





#29

Chloroform

Concen: 8.738 ppbv

RT: 5.85 min Scan# 905

Delta R.T. -0.00 min

Lab File: VL033602.D

Acq: 17 Apr 2019 14:59

Instrument :

MSVOA_L

ClientSampleId :





VL0417ABSD01

Tot Ion: 83 Resp: 1266586

Ion Ratio Lower Upper

83 100

85 63.4 52.6 79.0



#30

Cyclohexane

Concen: 9.954 ppbv

RT: 7.25 min Scan# 1340

Delta R.T. -0.00 min

Lab File: VL033602.D

Acq: 17 Apr 2019 14:59




Tgt Ion: 84 Resp: 665878


Ion Ratio Lower Upper

84 100

56 135.3 109.7 164.5

41 81.8 68.2 102.2









#31

cis-1,2-Dichloroethene
 Concen: 9.245 ppbv
 RT: 5.64 min Scan# 841
 Delta R.T. -0.00 min
 Lab File: VL033602.D
 Acq: 17 Apr 2019 14:59

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0417ABSD01




Tot Ion	Ion	Ratio	Lower	Upper
61	100			
96	66.6	50.6	76.0	
98	42.7	32.2	48.2	




#32

1,1,1-Trichloroethane
 Concen: 9.401 ppbv
 RT: 6.62 min Scan# 1146
 Delta R.T. 0.00 min
 Lab File: VL033602.D
 Acq: 17 Apr 2019 14:59

Tgt Ion	Ion	Ratio	Lower	Upper
97	100			
99	64.6	51.5	77.3	
61	47.5	38.8	58.2	









#33
 1,4-Difluorobenzene
 Concen: 10.000 ppbv
 RT: 7.29 min Scan# 1355
 Delta R.T. -0.00 min
 Lab File: VL033602.D
 Acq: 17 Apr 2019 14:59

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0417ABSD01

Tot Ion:114 Resp: 2169687




Ion	Ratio	Lower	Upper
114	100		
63	26.0	21.5	32.3
88	18.0	14.5	21.7




#34
 2-Butanone
 Concen: 8.543 ppbv
 RT: 5.33 min Scan# 744
 Delta R.T. 0.00 min
 Lab File: VL033602.D
 Acq: 17 Apr 2019 14:59

Tgt Ion: 43 Resp: 1113643

Ion	Ratio	Lower	Upper
43	100		
72	22.1	17.4	26.2





#35

Carbon Tetrachloride

Concen: 9.096 ppbv

RT: 7.14 min Scan# 1306

Delta R.T. 0.00 min

Lab File: VL033602.D

Acq: 17 Apr 2019 14:59

Instrument :

MSVOA_L

ClientSampleId :

VL0417ABSD01


Tot Ion:117 Resp: 1397830

Ion Ratio Lower Upper

117 100

119 98.1 75.8 113.8




121 30.6 24.7 37.1



Abundance Ion 117.00 (116.70 to 117.70): V

Ion 119.00 (118.70 to 119.70): V

Ion 121.00 (120.70 to 121.70): V



#36

Benzene

Concen: 9.094 ppbv

RT: 7.01 min Scan# 1266

Delta R.T. 0.00 min

Lab File: VL033602.D

Acq: 17 Apr 2019 14:59


Tgt Ion: 78 Resp: 1655565

Ion Ratio Lower Upper

78 100

77 22.3 18.1 27.1



50 19.5 16.7 25.1




Abundance Ion 78.10 (77.80 to 78.80): VL0

Ion 77.10 (76.80 to 77.80): VL0

Ion 50.10 (49.80 to 50.80): VL0





#37

1,2-Dichloroethane

Concen: 8.031 ppbv

RT: 6.42 min Scan# 1082

Delta R.T. 0.00 min

Lab File: VL033602.D

Acq: 17 Apr 2019 14:59

Instrument :

MSVOA_L

ClientSampleId :





VL0417ABSD01

Tot Ion: 62 Resp: 980012

Ion Ratio Lower Upper

62 100

98 5.8 4.4 6.6



#38

Trichloroethene

Concen: 9.603 ppbv

RT: 7.96 min Scan# 1563

Delta R.T. -0.00 min

Lab File: VL033602.D

Acq: 17 Apr 2019 14:59

Tgt Ion:130 Resp: 657909




Ion Ratio Lower Upper


130 100

95 111.8 89.8 134.8

132 93.0 76.6 115.0

97 70.7 58.6 88.0









#39
 1,2-Dichloropropane
 Concen: 8.529 ppbv
 RT: 7.74 min Scan# 1493
 Delta R.T. -0.00 min
 Lab File: VL033602.D
 Acq: 17 Apr 2019 14:59




Instrument :
 MSVOA_L
 ClientSampleId :
 VL0417ABSD01


Tot Ion: 63 Resp: 608243
 Ion Ratio Lower Upper
 63 100
 41 72.1 61.2 91.8
 62 68.7 54.0 81.0



#40
 1,4-Dioxane
 Concen: 9.416 ppbv
 RT: 7.95 min Scan# 1560
 Delta R.T. -0.00 min
 Lab File: VL033602.D
 Acq: 17 Apr 2019 14:59

Tgt Ion: 88 Resp: 239496
 Ion Ratio Lower Upper
 88 100
 58 128.9 105.6 158.4
 43 252.1 216.7 325.1
 57 1140.6 975.0 1462.4





#41

Tetrahydrofuran

Concen: 8.915 ppbv

RT: 6.15 min Scan# 998

Delta R.T. -0.00 min

Lab File: VL033602.D

Acq: 17 Apr 2019 14:59

Instrument :

MSVOA_L

ClientSampleId :

VL0417ABSD01


Tot Ion: 42 Resp: 639234

Ion Ratio Lower Upper

42 100

72 39.4 30.2 45.2




71 38.5 29.4 44.2



Abundance Ion 42.10 (41.80 to 42.80): VL0

Ion 72.10 (71.80 to 72.80): VL0

Ion 71.10 (70.80 to 71.80): VL0



#42

Bromodichloromethane

Concen: 8.599 ppbv

RT: 7.92 min Scan# 1549

Delta R.T. -0.00 min

Lab File: VL033602.D

Acq: 17 Apr 2019 14:59


Tgt Ion: 83 Resp: 1429174

Ion Ratio Lower Upper

83 100

85 64.7 50.7 76.1



127 8.2 6.6 9.8




Abundance Ion 83.00 (82.70 to 83.70): VL0

Ion 85.00 (84.70 to 85.70): VL0

Ion 126.90 (126.60 to 127.60): V









#43
Methvl Methacrylate
Concen: 8.880 ppbv
RT: 8.14 min Scan# 1617
Delta R.T. -0.00 min
Lab File: VL033602.D
Acq: 17 Apr 2019 14:59




Instrument :
MSVOA_L
ClientSampleId :
VL0417ABSD01


Tot Ion: 69 Resp: 656427
Ion Ratio Lower Upper
69 100
41 146.0 116.6 174.8
100 31.0 24.4 36.6



#44
2,2,4-Trimethylpentane
Concen: 8.572 ppbv
RT: 7.99 min Scan# 1572
Delta R.T. -0.00 min
Lab File: VL033602.D
Acq: 17 Apr 2019 14:59

Tgt Ion: 57 Resp: 2727097
Ion Ratio Lower Upper
57 100
41 31.4 26.0 39.0
56 30.7 24.6 37.0





#45

t-1,3-Dichloropropene

Concen: 10.088 ppbv

RT: 9.42 min Scan# 2015

Delta R.T. -0.00 min

Lab File: VL033602.D

Acq: 17 Apr 2019 14:59

Instrument :

MSVOA_L

ClientSampleId :


VL0417ABSD01

Tot Ion: 75 Resp: 915584

Ion Ratio Lower Upper




75 100

77 30.7 24.5 36.7



Abundance Ion 75.10 (74.80 to 75.80): VL0

Ion 77.10 (76.80 to 77.80): VL0



#46

cis-1,3-Dichloropropene

Concen: 9.718 ppbv

RT: 8.83 min Scan# 1834

Delta R.T. -0.00 min

Lab File: VL033602.D

Acq: 17 Apr 2019 14:59


Tgt Ion: 75 Resp: 1029980

Ion Ratio Lower Upper

75 100

39 65.2 54.1 81.1



77 31.9 23.9 35.9




Abundance Ion 75.00 (74.70 to 75.70): VL0

Ion 39.00 (38.70 to 39.70): VL0

Ion 77.00 (76.70 to 77.70): VL0





#47

1,1,2-Trichloroethane

Concen: 8.563 ppbv

RT: 9.62 min Scan# 2077

Delta R.T. -0.00 min

Lab File: VL033602.D

Acq: 17 Apr 2019 14:59

Instrument :

MSVOA_L

ClientSampleId :

VL0417ABSD01

Tot Ion: 97 Resp: 693449


Ion Ratio Lower Upper

97 100

83 91.8 70.5 105.7

85 59.3 45.6 68.4

99 62.1 49.4 74.0






Abundance Ion 97.00 (96.70 to 97.70): VL0

Ion 83.00 (82.70 to 83.70): VL0

Ion 85.00 (84.70 to 85.70): VL0

Ion 99.00 (98.70 to 99.70): VL0



#48

Dibromochloromethane

Concen: 9.188 ppbv

RT: 10.45 min Scan# 2337

Delta R.T. -0.00 min

Lab File: VL033602.D


Acq: 17 Apr 2019 14:59

Tgt Ion: 129 Resp: 1254957

Ion Ratio Lower Upper



129 100


127 77.7 39.1 117.2



Abundance Ion 129.00 (128.70 to 129.70): V

Ion 127.00 (126.70 to 127.70): V





#49

Bromoform

Concen: 8.925 ppbv

RT: 13.20 min Scan# 3193

Delta R.T. -0.00 min

Lab File: VL033602.D

Acq: 17 Apr 2019 14:59

Instrument :

MSVOA_L

ClientSampleId :

VL0417ABSD01


Tot Ion:173 Resp: 1164613

Ion Ratio Lower Upper

173 100

175 48.5 39.3 58.9




171 51.0 41.2 61.8



Abundance Ion 172.90 (172.60 to 173.60): V

Ion 174.90 (174.60 to 175.60): V

Ion 170.90 (170.60 to 171.60): V



#50

4-Methyl-2-Pentanone

Concen: 8.658 ppbv

RT: 8.87 min Scan# 1845

Delta R.T. 0.00 min

Lab File: VL033602.D


Acq: 17 Apr 2019 14:59

Tgt Ion: 43 Resp: 1652537

Ion Ratio Lower Upper



43 100


58 36.1 28.0 42.0



Abundance Ion 43.10 (42.80 to 43.80): VL0

Ion 58.10 (57.80 to 58.80): VL0





#51

2-Hexanone

Concen: 9.170 ppbv

RT: 10.27 min Scan# 2281

Delta R.T. -0.00 min

Lab File: VL033602.D

Acq: 17 Apr 2019 14:59

Instrument :

MSVOA_L

ClientSampleId :

VL0417ABSD01


Tot Ion: 43 Resp: 1305722

Ion Ratio Lower Upper




43 100

58 47.8 38.2 57.2

98 0.1 0.2 0.2#



Abundance Ion 43.10 (42.80 to 43.80): VL0
Ion 58.10 (57.80 to 58.80): VL0
Ion 98.20 (97.90 to 98.90): VL0



#52

Tetrachloroethene

Concen: 9.216 ppbv

RT: 11.37 min Scan# 2623

Delta R.T. -0.00 min

Lab File: VL033602.D

Acq: 17 Apr 2019 14:59

Tgt Ion: 164 Resp: 633032


Ion Ratio Lower Upper

164 100



166 124.9 99.5 149.3


129 101.7 81.4 122.2

131 98.0 75.6 113.4



Abundance Ion 163.90 (163.60 to 164.60): V
Ion 165.90 (165.60 to 166.60): V
Ion 129.00 (128.70 to 129.70): V
Ion 131.00 (130.70 to 131.70): V





#53

Toluene

Concen: 9.224 ppbv

RT: 9.95 min Scan# 2180

Delta R.T. -0.00 min

Lab File: VL033602.D

Acq: 17 Apr 2019 14:59

Instrument :

MSVOA_L

ClientSampleId :





VL0417ABSD01

Tot Ion: 91 Resp: 1951355

Ion Ratio Lower Upper

91 100

92 59.7 47.4 71.0



#54

1,2-Dibromoethane

Concen: 8.821 ppbv

RT: 10.76 min Scan# 2433

Delta R.T. -0.00 min

Lab File: VL033602.D




Acq: 17 Apr 2019 14:59


Tgt Ion:107 Resp: 1117265

Ion Ratio Lower Upper

107 100

109 93.4 72.7 109.1









#55
Chlorobenzene-d5
Concen: 10.000 ppbv m
RT: 12.24 min Scan# 2894
Delta R.T. -0.00 min
Lab File: VL033602.D
Acq: 17 Apr 2019 14:59




Instrument :
MSVOA_L
ClientSampleId :
VL0417ABSD01


Tot Ion:117 Resp: 2298507
Ion Ratio Lower Upper
117 100
82 61.9 51.5 77.3
119 53.0 25.5 38.3#



#56
1,1,1,2-Tetrachloroethane
Concen: 8.236 ppbv
RT: 12.28 min Scan# 2905
Delta R.T. -0.01 min
Lab File: VL033602.D
Acq: 17 Apr 2019 14:59

Tgt Ion:131 Resp: 867581
Ion Ratio Lower Upper
131 100
133 96.5 76.2 114.4
119 140.6 45.0 67.6#





#57

Chlorobenzene

Concen: 8.253 ppbv

RT: 12.30 min Scan# 2913

Delta R.T. -0.00 min

Lab File: VL033602.D

Acq: 17 Apr 2019 14:59

Instrument :

MSVOA_L

ClientSampleId :

VL0417ABSD01


Tot Ion:112 Resp: 1473035

Ion Ratio Lower Upper

112 100

77 67.6 57.3 85.9




114 32.8 29.8 36.4



Abundance Ion 112.10 (111.80 to 112.80): V

Ion 77.10 (76.80 to 77.80): VL0

Ion 114.10 (113.80 to 114.80): V



#58

Ethyl Benzene

Concen: 8.807 ppbv

RT: 12.86 min Scan# 3087

Delta R.T. -0.00 min

Lab File: VL033602.D


Acq: 17 Apr 2019 14:59

Tgt Ion: 91 Resp: 2703209

Ion Ratio Lower Upper



91 100


106 30.2 23.8 35.6



Abundance Ion 91.10 (90.80 to 91.80): VL0

Ion 106.10 (105.80 to 106.80): V









#59
m/p-Xylene
Concen: 16.964 ppbv m
RT: 13.15 min Scan# 3176
Delta R.T. -0.00 min
Lab File: VL033602.D
Acq: 17 Apr 2019 14:59




Instrument :
MSVOA_L
ClientSampleId :
VL0417ABSD01


Tot Ion: 91 Resp: 4651000
Ion Ratio Lower Upper
91 100
106 45.2 0.0 0.0#



#60
o-Xylene
Concen: 8.632 ppbv
RT: 13.85 min Scan# 3394
Delta R.T. -0.00 min
Lab File: VL033602.D
Acq: 17 Apr 2019 14:59

Tgt Ion: 91 Resp: 2288526
Ion Ratio Lower Upper
91 100
106 43.5 21.4 64.3





#61

Stvrene

Concen: 9.394 ppbv

RT: 13.69 min Scan# 3343

Delta R.T. -0.00 min

Lab File: VL033602.D

Acq: 17 Apr 2019 14:59

Instrument :

MSVOA_L

ClientSampleId :

VL0417ABSD01


Tot Ion:104 Resp: 1709298

Ion Ratio Lower Upper

104 100

78 53.2 43.2 64.8




103 47.6 37.8 56.8



Abundance Ion 104.10 (103.80 to 104.80): V

Ion 78.10 (77.80 to 78.80): VL0

Ion 103.10 (102.80 to 103.80): V



#62

Isopropylbenzene

Concen: 8.557 ppbv

RT: 14.83 min Scan# 3698

Delta R.T. -0.00 min

Lab File: VL033602.D


Acq: 17 Apr 2019 14:59

Tgt Ion:105 Resp: 3212604

Ion Ratio Lower Upper



105 100


120 25.4 20.2 30.2



Abundance Ion 105.10 (104.80 to 105.80): V

Ion 120.20 (119.90 to 120.90): V









#63
 1,1,2,2-Tetrachloroethane
 Concen: 8.118 ppbv
 RT: 13.84 min Scan# 3390
 Delta R.T. -0.00 min
 Lab File: VL033602.D
 Acq: 17 Apr 2019 14:59




Instrument :
 MSVOA_L
 ClientSampleId :
 VL0417ABSD01


Tot Ion: 83 Resp: 1584804
 Ion Ratio Lower Upper
 83 100
 131 9.5 4.7 14.0
 85 64.9 32.7 98.1



#64
 n-propylbenzene
 Concen: 8.687 ppbv
 RT: 15.72 min Scan# 3976
 Delta R.T. -0.01 min
 Lab File: VL033602.D
 Acq: 17 Apr 2019 14:59

Tgt Ion:120 Resp: 827536
 Ion Ratio Lower Upper
 120 100
 91 478.0 387.4 581.2









#65
tert-Butylbenzene
Concen: 8.794 ppbv
RT: 16.91 min Scan# 4345
Delta R.T. -0.00 min
Lab File: VL033602.D
Acq: 17 Apr 2019 14:59




Instrument :
MSVOA_L
ClientSampleId :
VL0417ABSD01


Tot Ion: 119 Resp: 2909349
Ion Ratio Lower Upper
119 100
91 81.4 66.4 99.6
134 19.6 15.5 23.3



#66
Benzyl Chloride
Concen: 10.784 ppbv
RT: 17.15 min Scan# 4422
Delta R.T. -0.00 min
Lab File: VL033602.D
Acq: 17 Apr 2019 14:59

Tgt Ion: 91 Resp: 1658916
Ion Ratio Lower Upper
91 100
126 16.9 13.0 19.6
128 5.4 4.2 6.2





#67

sec-Butylbenzene

Concen: 8.716 ppbv

RT: 17.46 min Scan# 4516

Delta R.T. -0.00 min

Lab File: VL033602.D

Acq: 17 Apr 2019 14:59

Instrument :

MSVOA_L

ClientSampleId :


VL0417ABSD01

Tot Ion:105 Resp: 4100611

Ion Ratio Lower Upper




105 100

134 18.2 14.6 21.8



Abundance Ion 105.10 (104.80 to 105.80): V

Ion 134.20 (133.90 to 134.90): V



#68

1-Bromo-4-Fluorobenzene

Concen: 9.629 ppbv

RT: 14.58 min Scan# 3620

Delta R.T. -0.00 min

Lab File: VL033602.D

Acq: 17 Apr 2019 14:59


Tgt Ion: 95 Resp: 1735160

Ion Ratio Lower Upper

95 100

174 68.0 52.9 79.3



175 4.9 3.8 5.6




Abundance Ion 95.10 (94.80 to 95.80): VL0

Ion 174.00 (173.70 to 174.70): V

Ion 175.00 (174.70 to 175.70): V





#69

p-Isopropyltoluene

Concen: 8.932 ppbv

RT: 17.82 min Scan# 4628

Delta R.T. -0.00 min

Lab File: VL033602.D

Acq: 17 Apr 2019 14:59

Instrument :

MSVOA_L

ClientSampleId :

VL0417ABSD01


Tot Ion:119 Resp: 3421480

Ion Ratio Lower Upper

119 100

134 25.5 19.9 29.9




91 28.6 23.4 35.0



Abundance Ion 119.10 (118.80 to 119.80): V

Ion 134.20 (133.90 to 134.90): V

Ion 91.10 (90.80 to 91.80): V



#70

n-Butylbenzene

Concen: 8.709 ppbv

RT: 18.71 min Scan# 4907

Delta R.T. -0.00 min

Lab File: VL033602.D

Acq: 17 Apr 2019 14:59


Tgt Ion: 91 Resp: 3534796

Ion Ratio Lower Upper

91 100

92 53.4 42.6 64.0



134 22.7 17.5 26.3




Abundance Ion 91.10 (90.80 to 91.80): VL0

Ion 92.10 (91.80 to 92.80): VL0

Ion 134.20 (133.90 to 134.90): V





#71

2-Chlorotoluene

Concen: 8.701 ppbv

RT: 15.61 min Scan# 3943

Delta R.T. -0.01 min

Lab File: VL033602.D

Acq: 17 Apr 2019 14:59

Instrument :

MSVOA_L

ClientSampleId :





VL0417ABSD01

Tot Ion: 91 Resp: 2483906

Ion Ratio Lower Upper

91 100

126 29.9 11.7 46.9



#72

4-Ethyltoluene

Concen: 8.936 ppbv

RT: 16.00 min Scan# 4063

Delta R.T. -0.00 min

Lab File: VL033602.D

Acq: 17 Apr 2019 14:59

Tgt Ion:105 Resp: 2663147




Ion Ratio Lower Upper


105 100

120 28.9 23.2 34.8

91 12.9 10.4 15.6

77 13.8 11.3 16.9





#73

1,3,5-Trimethylbenzene

Concen: 8.959 ppbv

RT: 16.15 min Scan# 4111

Delta R.T. -0.01 min

Lab File: VL033602.D

Acq: 17 Apr 2019 14:59

Instrument :

MSVOA_L

ClientSampleId :


VL0417ABSD01

Tot Ion:105 Resp: 2554857

Ion Ratio Lower Upper




105 100

120 45.3 35.7 53.5



Abundance Ion 105.10 (104.80 to 105.80): V

Ion 120.20 (119.90 to 120.90): V



#74

1,2,4-Trimethylbenzene

Concen: 8.691 ppbv

RT: 16.93 min Scan# 4351

Delta R.T. -0.00 min

Lab File: VL033602.D

Acq: 17 Apr 2019 14:59

Tgt Ion:105 Resp: 2656672


Ion Ratio Lower Upper

105 100

120 46.6 39.0 58.6

91 45.8 40.0 60.0

77 20.8 17.5 26.3






Abundance Ion 105.10 (104.80 to 105.80): V

Ion 120.20 (119.90 to 120.90): V

Ion 91.10 (90.80 to 91.80): V

Ion 77.10 (76.80 to 77.80): V





#75

1,3-Dichlorobenzene

Concen: 8.385 ppbv

RT: 17.16 min Scan# 4425

Delta R.T. -0.01 min

Lab File: VL033602.D

Acq: 17 Apr 2019 14:59

Instrument :

MSVOA_L

ClientSampleId :

VL0417ABSD01


Tot Ion:146 Resp: 1584098

Ion Ratio Lower Upper

146 100

111 45.0 22.9 68.8




148 64.3 32.0 96.2



Abundance Ion 146.00 (145.70 to 146.70): V

Ion 111.10 (110.80 to 111.80): V

Ion 148.00 (147.70 to 148.70): V



#76

1,4-Dichlorobenzene

Concen: 8.611 ppbv

RT: 17.31 min Scan# 4470

Delta R.T. -0.00 min

Lab File: VL033602.D

Acq: 17 Apr 2019 14:59


Tgt Ion:146 Resp: 1585061

Ion Ratio Lower Upper

146 100

111 43.7 21.7 65.1



148 64.4 32.0 96.2




Abundance Ion 146.00 (145.70 to 146.70): V

Ion 111.10 (110.80 to 111.80): V

Ion 148.00 (147.70 to 148.70): V









#77
 1,2-Dichlorobenzene
 Concen: 8.679 ppbv
 RT: 17.99 min Scan# 4681
 Delta R.T. -0.00 min
 Lab File: VL033602.D
 Acq: 17 Apr 2019 14:59

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0417ABSD01

Tot Ion:146 Resp: 1555763
 Ion Ratio Lower Upper
 146 100
 111 46.4 23.5 70.5
 148 64.1 32.0 96.0




Abundance Ion 146.00 (145.70 to 146.70): V
 Ion 111.10 (110.80 to 111.80): V
 Ion 148.00 (147.70 to 148.70): V






#78
 Hexachloro-1,3-Butadiene
 Concen: 7.689 ppbv
 RT: 23.54 min Scan# 6409
 Delta R.T. -0.00 min
 Lab File: VL033602.D
 Acq: 17 Apr 2019 14:59

Tgt Ion:225 Resp: 912343
 Ion Ratio Lower Upper
 225 100
 223 63.9 50.0 75.0
 227 63.9 51.1 76.7



Abundance Ion 224.90 (224.60 to 225.60): V
 Ion 222.90 (222.60 to 223.60): V
 Ion 226.90 (226.60 to 227.60): V





#79

Naphthalene

Concen: 10.028 ppbv

RT: 22.41 min Scan# 6056

Delta R.T. -0.01 min

Lab File: VL033602.D

Acq: 17 Apr 2019 14:59

Instrument :

MSVOA_L

ClientSampleId :

VL0417ABSD01


Tot Ion:128 Resp: 2623870

Ion Ratio Lower Upper

128 100

127 13.0 9.8 14.8




129 11.2 8.6 12.8



Abundance Ion 128.10 (127.80 to 128.80): V

Ion 127.10 (126.80 to 127.80): V

Ion 129.10 (128.80 to 129.80): V



#80

Naphthalene, 2-methyl-

Concen: 13.121 ppbv

RT: 25.11 min Scan# 6896

Delta R.T. -0.00 min

Lab File: VL033602.D

Acq: 17 Apr 2019 14:59


Tgt Ion:142 Resp: 1152179

Ion Ratio Lower Upper

142 100

141 82.3 66.6 99.8



115 41.4 34.4 51.6



Abundance Ion 142.00 (141.70 to 142.70): V

Ion 141.00 (140.70 to 141.70): V

Ion 115.00 (114.70 to 115.70): V





#81
 1,2,4-Trichlorobenzene
 Concen: 9.583 ppbv
 RT: 22.14 min Scan# 5971
 Delta R.T. -0.01 min
 Lab File: VL033602.D
 Acq: 17 Apr 2019 14:59

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0417ABSD01

Tot Ion:180 Resp: 1390903
 Ion Ratio Lower Upper
 180 100
 182 95.5 77.0 115.4
 184 31.1 24.6 36.8

