


Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL041719\
Data File : VL033598.D
Acq On : 17 Apr 2019 11:18
Operator : SY/AP
Sample : BFB
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL032519AIR.M
Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Tue Mar 26 16:00:50 2019
Last Update : Tue Mar 26 16:00:50 2019



AutoFind: Scans 3616, 3617, 3618; Background Corrected with Scan 3594

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	8	40	23.7	166890	PASS
75	95	30	66	57.2	402197	PASS
95	95	100	100	100.0	703146	PASS
96	95	5	9	7.3	51637	PASS
173	174	0.00	2	0.9	4123	PASS
174	95	50	120	66.4	467136	PASS
175	174	4	9	7.3	33888	PASS
176	174	93	101	95.9	448128	PASS
177	176	5	9	6.7	30149	PASS