

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL041825\
 Data File : VL042391.D
 Acq On : 18 Apr 2025 14:45
 Operator : SY/MD
 Sample : Q1726-05DL2 100X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 IA-B2-1-03312025DL2

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 04/21/2025
 Supervised By :Mahesh Dadoda 04/21/2025

Quant Time: Apr 18 22:59:44 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL041725AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Fri Apr 18 02:14:00 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.787	49	125703	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.959	114	390416	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.882	117	327579	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.377	95	273896	10.175	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.700%
Target Compounds						
13) Ethanol	1.722	45	3027	1.977	ppbv #	24
15) Acetone	1.835	43	104075	5.058	ppbv	99
19) Isopropyl Alcohol	1.890	45	3377	0.295	ppbv #	1
53) Toluene	6.936	91	9781m	0.194	ppbv	
59) m/p-Xylene	9.529	91	6487m	0.124	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL041825\
Data File : VL042391.D
Acq On : 18 Apr 2025 14:45
Operator : SY/MD
Sample : Q1726-05DL2 100X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
IA-B2-1-03312025DL2

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 04/21/2025
Supervised By :Mahesh Dadoda 04/21/2025

Quant Time: Apr 18 22:59:44 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL041725AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Fri Apr 18 02:14:00 2025
Response via : Initial Calibration

