

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL041825\
 Data File : VL042393.D
 Acq On : 18 Apr 2025 15:47
 Operator : SY/MD
 Sample : Q1726-06DL2DUP 100X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 IA-B2-3-03312025DL2DUP

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 04/21/2025
 Supervised By :Mahesh Dadoda 04/21/2025

Quant Time: Apr 18 23:00:28 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL041725AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Fri Apr 18 02:14:00 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.790	49	123948	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.965	114	373006	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.888	117	318921	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.380	95	266367	10.164	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.600%
Target Compounds						
					Qvalue	
13) Ethanol	1.722	45	3385	2.243	ppbv #	24
15) Acetone	1.839	43	96169	4.740	ppbv	100
19) Isopropyl Alcohol	1.891	45	3389	0.301	ppbv #	1
53) Toluene	6.940	91	6512m	0.135	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL041825\
Data File : VL042393.D
Acq On : 18 Apr 2025 15:47
Operator : SY/MD
Sample : Q1726-06DL2DUP 100X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
IA-B2-3-03312025DL2DUP

Quant Time: Apr 18 23:00:28 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL041725AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Fri Apr 18 02:14:00 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 04/21/2025
Supervised By : Mahesh Dadoda 04/21/2025

