

Data Path : W:\HPCHEM1\MSVOA L\DATA\VL041918\
Data File : VL031883.D
Acq On : 19 Apr 2018 18:30
Operator : SY/AP
Sample : OC041118 CAN 10261
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC041118 CAN 10261

Quant Time: Apr 20 08:19:37 2018
Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL041218AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Apr 13 16:50:02 2018
QLast Update : Fri Apr 13 16:50:02 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.88	49	1255468	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.42	114	2707524	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.39	117	2548944	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.73	95	2036786	9.94	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.40%

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA L\DATA\VL041918\
Data File : VL031883.D
Acq On : 19 Apr 2018 18:30
Operator : SY/AP
Sample : OC041118 CAN 10261
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC041118 CAN 10261

Quant Time: Apr 20 08:19:37 2018
Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL041218AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Apr 13 16:50:02 2018
QLast Update : Fri Apr 13 16:50:02 2018
Response via : Initial Calibration

