

Data Path : W:\HPCHEM1\MSVOA L\DATA\VL041918\
 Data File : VL031887.D
 Acq On : 19 Apr 2018 20:58
 Operator : SY/AP
 Sample : J2133-13 0.03 PPBV
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 LOD-MDL-AIR-01-QT2-2018

Manual Integrations
 APPROVED

sam
 4/23/2018 12:50:39 PM

Quant Time: Apr 20 09:11:47 2018
 Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL041218AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Apr 13 16:50:02 2018
 QLast Update : Fri Apr 13 16:50:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.88	49	1088314	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.42	114	2498666	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.39	117	2487064	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.72	95	2010539	10.05	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.50%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
5) Vinyl Chloride	3.38	62	3537m	0.05	ppbv	
32) 1,1,1-Trichloroethane	6.75	97	8873m	0.04	ppbv	
35) Carbon Tetrachloride	7.27	117	9134m	0.04	ppbv	
38) Trichloroethene	8.10	130	3245m	0.04	ppbv	
52) Tetrachloroethene	11.51	164	3632m	0.04	ppbv	
63) 1,1,2,2-Tetrachloroethane	13.99	83	10197m	0.04	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA L\DATA\VL041918\
Data File : VL031887.D
Acq On : 19 Apr 2018 20:58
Operator : SY/AP
Sample : J2133-13 0.03 PPBV
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
LOD-MDL-AIR-01-QT2-2018

Manual Integrations
APPROVED

sam
4/23/2018 12:50:39 PM

Quant Time: Apr 20 09:11:47 2018
Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL041218AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Apr
QLast Update : Fri Apr 13 16:50:02 2018
Response via : Initial Calibration

