

Data Path : W:\HPCHEM1\MSVOA L\DATA\VL041918\
Data File : VL031895.D
Acq On : 20 Apr 2018 11:11
Operator : SY/AP
Sample : QC041118 CAN 10261
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC041118 CAN 10261

Quant Time: Apr 21 02:20:04 2018
Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL041218AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Apr 13 16:50:02 2018
QLast Update : Fri Apr 13 16:50:02 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.88	49	1080781	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.42	114	2466980	10.00	ppbv	-0.01
55) Chlorobenzene-d5	12.38	117	2355517	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.72	95	1925486	10.17	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.70%

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA L\DATA\VL041918\
Data File : VL031895.D
Acq On : 20 Apr 2018 11:11
Operator : SY/AP
Sample : OC041118 CAN 10261
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC041118 CAN 10261

Quant Time: Apr 21 02:20:04 2018
Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL041218AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Apr 13 16:50:02 2018
QLast Update : Fri Apr 13 16:50:02 2018
Response via : Initial Calibration

