

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL041923\
 Data File : VL040389.D
 Acq On : 19 Apr 2023 21:06
 Operator : SY/MD
 Sample : 02368-05 20X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VP-1A

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 04/20/2023
 Supervised By :Mahesh Dadoda 04/22/2023

Quant Time: Apr 20 05:14:31 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL041323AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2023
 QLast Update : Thu Apr 13 14:37:02 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.179	49	1005850	10.000	ppbv	-0.01
33) 1,4-Difluorobenzene	6.642	114	2729808	10.000	ppbv	-0.01
55) Chlorobenzene-d5	11.458	117	2437759	10.000	ppbv	-0.02
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.770	95	1929875	9.573	ppbv	-0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	95.700%
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	2.710	85	11238	0.062	ppbv	88
9) Propene	2.687	41	9467	0.182	ppbv	84
13) Ethanol	3.237	45	52733	12.165	ppbv	96
15) Acetone	3.459	43	331758	2.329	ppbv	92
20) Methylene Chloride	3.915	84	15448	0.272	ppbv	98
29) Chloroform	5.259	83	10037m	0.056	ppbv	
34) 2-Butanone	4.787	43	37347m	0.254	ppbv	
52) Tetrachloroethene	10.600	164	6099m	0.088	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL041923\
Data File : VL040389.D
Acq On : 19 Apr 2023 21:06
Operator : SY/MD
Sample : 02368-05 20X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VP-1A

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 04/20/2023
Supervised By : Mahesh Dadoda 04/22/2023

Quant Time: Apr 20 05:14:31 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL041323AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Thu Apr 13 14:37:02 2023
Response via : Initial Calibration

