

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL042021\
Data File : VL036834.D
Acq On : 20 Apr 2021 23:04
Operator : SY/AP
Sample : QC041921 CAN 10281
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC041921 CAN 10281

Quant Time: Apr 21 12:45:07 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL032521AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Mar 25 09:42:40 2021
QLast Update : Thu Mar 25 09:42:40 2021
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.68	49	1147509	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.18	114	2799533	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.10	117	2372706	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.43	95	1846702	9.77	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.70%

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL042021\
Data File : VL036834.D
Acq On : 20 Apr 2021 23:04
Operator : SY/AP
Sample : OC041921 CAN 10281
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC041921 CAN 10281

Quant Time: Apr 21 12:45:07 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL032521AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Mar 25 09:42:40 2021
QLast Update : Thu Mar 25 09:42:40 2021
Response via : Initial Calibration

