

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL042021\
 Data File : VL036836.D
 Acq On : 21 Apr 2021 00:19
 Operator : SY/AP
 Sample : VIBLK
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VIBLK

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 04/21/2021
 Supervised By :Mahesh Dadoda 04/21/2021

Quant Time: Apr 21 12:47:25 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL032521A
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
 QLast Update : Thu Mar 25 09:42:40 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.67	49	1095781	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.19	114	2810938	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.10	117	2456241	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.43	95	1782141	9.10	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	91.00%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
15) Acetone	3.89	43	66058m	0.612	ppbv	
20) Methylene Chloride	4.36	84	11408m	0.226	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL042021\
Data File : VL036836.D
Acq On : 21 Apr 2021 00:19
Operator : SY/AP
Sample : VIBLK
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Manual Integrations
APPROVED

Quant Time: Apr 21 12:47:25 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL032521
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
QLast Update : Thu Mar 25 09:42:40 2021
Response via : Initial Calibration

Reviewed By :Semsettin Yesilyurt 04/21/2021
Supervised By :Mahesh Dadoda 04/21/2021

