

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL042125\
 Data File : VL042397.D
 Acq On : 21 Apr 2025 10:14
 Operator : SY/MD
 Sample : VL0421ABS01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0421ABS01

Quant Time: Apr 22 02:51:20 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL041725AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Fri Apr 18 02:14:00 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 04/22/2025
 Supervised By : Mahesh Dadoda 04/22/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.781	49	118401	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.949	114	337921	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.875	117	293537	10.000	ppbv	0.00

System Monitoring Compounds
 68) 1-Bromo-4-Fluorobenzene 10.371 95 260048 10.781 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 107.800%

Target Compounds	Qvalue					
2) Dichlorodifluoromethane	1.496	85	212503	10.158	ppbv	98
3) Chlorodifluoromethane	1.473	51	175535	8.812	ppbv	95
4) Chloromethane	1.531	50	79791	9.638	ppbv	94
5) Vinyl Chloride	1.577	62	77905	8.645	ppbv	97
6) Bromomethane	1.664	94	35087	9.064	ppbv	87
7) Chloroethane	1.700	64	30814	9.455	ppbv	96
8) Dichlorotetrafluoroethane	1.551	85	182258	10.147	ppbv	94
9) Propene	1.480	41	69080	7.615	ppbv	97
10) Heptane	4.969	43	187072	7.340	ppbv	99
11) Trichlorofluoromethane	1.875	101	205641	10.337	ppbv	100
12) 1,1,2-Trichlorotrifluo...	2.133	101	156363	9.613	ppbv	98
13) Ethanol	1.716	45	4602	3.192	ppbv	92
14) Bromoethene	1.777	108	58938	9.613	ppbv	98
15) Acetone	1.832	43	168065	8.671	ppbv	98
16) 1,3-Butadiene	1.606	39	83577	9.015	ppbv	92
17) tert-Butyl alcohol	2.036	59	216440	9.011	ppbv	100
18) 1,1-Dichloroethene	2.030	96	71097	9.356	ppbv	95
19) Isopropyl Alcohol	1.881	45	96437	8.956	ppbv	93
20) Methylene Chloride	2.056	84	60417	7.959	ppbv	89
21) Allyl Chloride	2.091	41	122313	9.140	ppbv	97
22) trans-1,2-Dichloroethene	2.334	96	78976	9.128	ppbv	98
23) Vinyl Acetate	2.457	43	54610	6.961	ppbv #	96
24) 1,1-Dichloroethane	2.402	63	157093	9.559	ppbv	99
25) Ethyl Acetate	2.829	43	393203	8.228	ppbv	99
26) Hexane	2.820	57	151103	7.813	ppbv	99
27) Carbon Disulfide	2.146	76	194719	9.404	ppbv	98
28) Methyl tert-Butyl Ether	2.434	73	108007	8.524	ppbv	93
29) Chloroform	2.842	83	262639	9.608	ppbv	98
30) Cyclohexane	3.846	84	133153	8.517	ppbv	98
31) cis-1,2-Dichloroethene	2.713	61	166447	9.052	ppbv	95
32) 1,1,1-Trichloroethane	3.360	97	285156	9.266	ppbv	98
34) 2-Butanone	2.554	43	220966	8.268	ppbv	98
35) Carbon Tetrachloride	3.755	117	291941	9.546	ppbv	95
36) Benzene	3.648	78	312290	8.568	ppbv	97
37) 1,2-Dichloroethane	3.211	62	225346	10.301	ppbv #	96
38) Trichloroethene	4.542	130	123977	7.783	ppbv	98
39) 1,2-Dichloropropane	4.305	63	112058	8.250	ppbv #	90
40) 1,4-Dioxane	4.571	88	55511	7.437	ppbv #	93
41) Tetrahydrofuran	3.046	42	121497	7.797	ppbv	99
42) Bromodichloromethane	4.480	83	294127	9.661	ppbv	98
43) Methyl Methacrylate	4.872	69	125779	7.657	ppbv	93
44) 2,2,4-Trimethylpentane	4.629	57	508984	8.085	ppbv	95
45) t-1,3-Dichloropropene	6.399	75	142114	8.469	ppbv	93

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL042125\
 Data File : VL042397.D
 Acq On : 21 Apr 2025 10:14
 Operator : SY/MD
 Sample : VL0421ABS01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0421ABS01

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 04/22/2025
 Supervised By :Mahesh Dadoda 04/22/2025

Quant Time: Apr 22 02:51:20 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL041725AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri Apr 18 02:14:00 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.590	75	171445	8.348	ppbv #	84
47) 1,1,2-Trichloroethane	6.574	97	124092	8.967	ppbv	95
48) Dibromochloromethane	7.377	129	225553	9.226	ppbv	99
49) Bromoform	9.474	173	188764	8.830	ppbv	100
50) 4-Methyl-2-Pentanone	5.739	43	318982	7.182	ppbv	99
51) 2-Hexanone	7.429	43	266951	6.702	ppbv #	95
52) Tetrachloroethene	8.222	164	104899	7.446	ppbv	97
53) Toluene	6.920	91	361443	8.264	ppbv	99
54) 1,2-Dibromoethane	7.645	107	175799	8.877	ppbv	94
56) 1,1,1,2-Tetrachloroethane	8.917	131	169656	9.550	ppbv	97
57) Chlorobenzene	8.914	112	270733	8.868	ppbv	98
58) Ethyl Benzene	9.348	91	524070	9.011	ppbv	100
59) m/p-Xylene	9.545	91	852193m	18.189	ppbv	
60) o-Xylene	9.956	91	433062	9.179	ppbv	97
61) Styrene	9.866	104	194756	8.648	ppbv	98
62) Isopropylbenzene	10.533	105	641535	9.262	ppbv	97
63) 1,1,2,2-Tetrachloroethane	9.956	83	257428	8.420	ppbv	99
64) n-propylbenzene	10.982	120	155343	8.922	ppbv	96
65) tert-Butylbenzene	11.526	119	573929	9.015	ppbv	96
66) Benzyl Chloride	11.610	91	51025	7.294	ppbv	97
67) sec-Butylbenzene	11.762	105	802355	9.061	ppbv	99
69) p-Isopropyltoluene	11.921	119	685449	8.910	ppbv	97
70) n-Butylbenzene	12.277	91	723138	9.025	ppbv	98
71) 2-Chlorotoluene	10.895	91	500409	9.211	ppbv	97
72) 4-Ethyltoluene	11.122	105	519236	9.002	ppbv	98
73) 1,3,5-Trimethylbenzene	11.199	105	456404	8.963	ppbv	96
74) 1,2,4-Trimethylbenzene	11.533	105	509154	8.561	ppbv	91
75) 1,3-Dichlorobenzene	11.604	146	267483	8.652	ppbv	98
76) 1,4-Dichlorobenzene	11.669	146	267283	8.602	ppbv	98
77) 1,2-Dichlorobenzene	11.944	146	261316	8.693	ppbv	97
78) Hexachloro-1,3-Butadiene	13.879	225	259140	8.537	ppbv	99
79) Naphthalene	13.510	128	584987	8.225	ppbv	99
80) Naphthalene,2-methyl-	14.455	142	300148	7.405	ppbv	96
81) 1,2,4-Trichlorobenzene	13.436	180	258107	8.452	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

