

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL042821\
 Data File : VL036878.D
 Acq On : 28 Apr 2021 14:01
 Operator : SY/AP
 Sample : M2172-01DL 10X
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SVE-INFLUENTDL

Manual Integrations
 APPROVED

MMDadoda
 4/29/2021 5:53:46 PM

Quant Time: Apr 29 03:24:02 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL042721AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LTue Apr 2
 QLast Update : Tue Apr 27 16:56:35 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.68	49	1548607	10.00	ppbv	0.02
33) 1,4-Difluorobenzene	7.19	114	3579288	10.00	ppbv	0.02
55) Chlorobenzene-d5	12.11	117	3099747	10.00	ppbv	0.03

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.44	95	2486402	9.99	ppbv	0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.90%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	3.03	41	12453m	0.145	ppbv	
13) Ethanol	3.65	45	38070m	3.473	ppbv	
15) Acetone	3.87	43	104533	0.716	ppbv #	65
19) Isopropyl Alcohol	4.03	45	19576m	0.271	ppbv	
20) Methylene Chloride	4.36	84	321000	4.599	ppbv	98
26) Hexane	5.71	57	163687	0.981	ppbv #	41
52) Tetrachloroethene	11.24	164	8256m	0.080	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL042821\
Data File : VL036878.D
Acq On : 28 Apr 2021 14:01
Operator : SY/AP
Sample : M2172-01DL 10X
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SVE-INFLUENTDL

Manual Integrations
APPROVED

MMDadoda
4/29/2021 5:53:46 PM

Quant Time: Apr 29 03:24:02 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL042721AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LTue Apr
QLast Update : Tue Apr 27 16:56:35 2021
Response via : Initial Calibration

