

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL042821\
 Data File : VL036881.D
 Acq On : 28 Apr 2021 16:01
 Operator : SY/AP
 Sample : M2201-01DL 10X
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 IA1DL

Manual Integrations
 APPROVED

MMDadoda
 4/29/2021 5:53:40 PM

Quant Time: Apr 29 03:31:13 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL042721AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Apr 28 16:00:41 2021
 QLast Update : Tue Apr 27 16:56:35 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.69	49	1499017	10.00	ppbv	0.03
33) 1,4-Difluorobenzene	7.20	114	3457573	10.00	ppbv	0.03
55) Chlorobenzene-d5	12.11	117	3062602	10.00	ppbv	0.03

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.44	95	2439200	9.92	ppbv	0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.20%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	3.03	41	382449m	4.597	ppbv	
13) Ethanol	3.64	45	17628650	1661.386	ppbv	99
15) Acetone	3.87	43	421650	2.984	ppbv	94
17) tert-Butyl alcohol	4.34	59	41400m	0.361	ppbv	
20) Methylene Chloride	4.36	84	159637	2.363	ppbv	96
26) Hexane	5.71	57	101863	0.631	ppbv #	40

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL042821\
Data File : VL036881.D
Acq On : 28 Apr 2021 16:01
Operator : SY/AP
Sample : M2201-01DL 10X
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
IA1DL

Manual Integrations
APPROVED

MMDadoda
4/29/2021 5:53:40 PM

Quant Time: Apr 29 03:31:13 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL042721AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Apr
QLast Update : Tue Apr 27 16:56:35 2021
Response via : Initial Calibration

