

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL042821\
 Data File : VL036885.D
 Acq On : 28 Apr 2021 18:40
 Operator : SY/AP
 Sample : M2201-03DL 10X
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 OA1DL

Manual Integrations
 APPROVED

MMDadoda
 4/29/2021 5:53:32 PM

Quant Time: Apr 29 12:27:13 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL042721AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_LTue Apr 2
 QLast Update : Tue Apr 27 16:56:35 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.69	49	1318057	10.00	ppbv	0.03
33) 1,4-Difluorobenzene	7.19	114	3255019	10.00	ppbv	0.03
55) Chlorobenzene-d5	12.11	117	2854349	10.00	ppbv	0.03

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.44	95	2297114	10.03	ppbv	0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.30%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Ethanol	3.65	45	35929m	3.851	ppbv	
15) Acetone	3.89	43	85426m	0.688	ppbv	
19) Isopropyl Alcohol	4.03	45	19079m	0.310	ppbv	
20) Methylene Chloride	4.36	84	53744	0.905	ppbv	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL042821\
Data File : VL036885.D
Acq On : 28 Apr 2021 18:40
Operator : SY/AP
Sample : M2201-03DL 10X
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QA1DL

Manual Integrations
APPROVED

MMDadoda
4/29/2021 5:53:32 PM

Quant Time: Apr 29 12:27:13 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL042721AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Apr
QLast Update : Tue Apr 27 16:56:35 2021
Response via : Initial Calibration

