

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL042821\
 Data File : VL036891.D
 Acq On : 28 Apr 2021 22:28
 Operator : SY/AP
 Sample : QC042721 CAN 10300
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC042721 CAN 10300

Manual Integrations
 APPROVED

MMDadoda
 4/29/2021 5:53:23 PM

Quant Time: Apr 29 03:57:43 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL042721AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Apr 2
 QLast Update : Tue Apr 27 16:56:35 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.68	49	1350291	10.00	ppbv	0.02
33) 1,4-Difluorobenzene	7.19	114	3164650	10.00	ppbv	0.02
55) Chlorobenzene-d5	12.10	117	2842766	10.00	ppbv	0.02

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.43	95	2232238	9.78	ppbv	0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.80%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
15) Acetone	3.89	43	34309m	0.270	ppbv	
20) Methylene Chloride	4.36	84	17217	0.283	ppbv #	82
30) Cyclohexane	7.17	84	35482m	0.308	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL042821\
Data File : VL036891.D
Acq On : 28 Apr 2021 22:28
Operator : SY/AP
Sample : QC042721 CAN 10300
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC042721 CAN 10300

Manual Integrations
APPROVED

MMDadoda
4/29/2021 5:53:23 PM

Quant Time: Apr 29 03:57:43 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL042721AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Apr
QLast Update : Tue Apr 27 16:56:35 2021
Response via : Initial Calibration

