

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
 Data File : VL042459.D
 Acq On : 01 May 2025 09:37
 Operator : SY/MD
 Sample : VSTDICC002
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC002

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 05/02/2025
 Supervised By : Mahesh Dadoda 05/05/2025

Quant Time: May 02 01:40:35 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2025

QLast Update : Fri May 02 01:37:48 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.787	49	73220	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.955	114	201951	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.881	117	174624	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.377 95 159777 9.869 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 98.700%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.499	85	23859	1.969	ppbv	100
3) Chlorodifluoromethane	1.476	51	26655	2.081	ppbv	95
4) Chloromethane	1.531	50	10751	2.123	ppbv	95
5) Vinyl Chloride	1.580	62	9122	1.789	ppbv	92
6) Bromomethane	1.667	94	4793	1.990	ppbv #	62
7) Chloroethane	1.703	64	3972	2.110	ppbv	94
8) Dichlorotetrafluoroethane	1.554	85	20529	2.055	ppbv	100
9) Propene	1.482	41	9234	2.204	ppbv	95
10) Heptane	4.978	43	23329m	2.146	ppbv	
11) Trichlorofluoromethane	1.877	101	22375	2.070	ppbv	96
12) 1,1,2-Trichlorotrifluo...	2.140	101	18230	2.042	ppbv	97
13) Ethanol	1.722	45	2099	2.468	ppbv	94
14) Bromoethene	1.780	108	6686	2.029	ppbv #	96
15) Acetone	1.835	43	29607	2.215	ppbv	99
16) 1,3-Butadiene	1.609	39	12245	2.077	ppbv	92
17) tert-Butyl alcohol	2.042	59	23722	2.014	ppbv #	93
18) 1,1-Dichloroethene	2.033	96	8303	1.992	ppbv	92
19) Isopropyl Alcohol	1.884	45	14288	1.931	ppbv	99
20) Methylene Chloride	2.062	84	7590	1.889	ppbv	96
21) Allyl Chloride	2.094	41	17050	1.983	ppbv	96
22) trans-1,2-Dichloroethene	2.340	96	9444	1.940	ppbv	97
23) Vinyl Acetate	2.463	43	26502	2.128	ppbv	98
24) 1,1-Dichloroethane	2.408	63	20027	2.016	ppbv	98
25) Ethyl Acetate	2.839	43	43122	2.015	ppbv	97
26) Hexane	2.822	57	19691	2.112	ppbv	95
27) Carbon Disulfide	2.149	76	21945	2.017	ppbv #	43
28) Methyl tert-Butyl Ether	2.441	73	11119	1.985	ppbv	94
29) Chloroform	2.848	83	35127	2.010	ppbv	86
30) Cyclohexane	3.858	84	16056	2.060	ppbv	95
31) cis-1,2-Dichloroethene	2.719	61	22619	2.079	ppbv	95
32) 1,1,1-Trichloroethane	3.369	97	36768	1.940	ppbv	99
34) 2-Butanone	2.557	43	31242	2.139	ppbv	95
35) Carbon Tetrachloride	3.764	117	35326	2.011	ppbv	97
36) Benzene	3.661	78	41848	2.196	ppbv #	88
37) 1,2-Dichloroethane	3.217	62	31049	2.114	ppbv	99
38) Trichloroethene	4.544	130	16311m	2.244	ppbv	
39) 1,2-Dichloropropane	4.311	63	14611	2.047	ppbv	91
40) 1,4-Dioxane	4.593	88	6957m	2.096	ppbv	
41) Tetrahydrofuran	3.059	42	15641m	2.258	ppbv	
42) Bromodichloromethane	4.489	83	37615	2.096	ppbv #	99
43) Methyl Methacrylate	4.884	69	14610m	2.059	ppbv	
44) 2,2,4-Trimethylpentane	4.638	57	65967	2.142	ppbv	99
45) t-1,3-Dichloropropene	6.409	75	18560m	1.987	ppbv	

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050125\
 Data File : VL042459.D
 Acq On : 01 May 2025 09:37
 Operator : SY/MD
 Sample : VSTDICC002
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC002

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/02/2025
 Supervised By :Mahesh Dadoda 05/05/2025

Quant Time: May 02 01:40:35 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri May 02 01:37:48 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.593	75	22184	2.000	ppbv	93
47) 1,1,2-Trichloroethane	6.587	97	17307m	2.212	ppbv	
48) Dibromochloromethane	7.386	129	27174	2.057	ppbv	99
49) Bromoform	9.483	173	23126	2.088	ppbv	98
50) 4-Methyl-2-Pentanone	5.758	43	39798	2.155	ppbv	99
51) 2-Hexanone	7.441	43	32196	2.177	ppbv #	97
52) Tetrachloroethene	8.224	164	13670	1.885	ppbv	93
53) Toluene	6.930	91	41392	2.052	ppbv	97
54) 1,2-Dibromoethane	7.652	107	24650	2.049	ppbv	99
56) 1,1,1,2-Tetrachloroethane	8.927	131	20822	2.019	ppbv	95
57) Chlorobenzene	8.920	112	36187	2.062	ppbv #	88
58) Ethyl Benzene	9.354	91	61177	2.096	ppbv	91
59) m/p-Xylene	9.551	91	105185m	4.299	ppbv	
60) o-Xylene	9.963	91	53051	2.169	ppbv	99
61) Styrene	9.872	104	18776	2.054	ppbv	97
62) Isopropylbenzene	10.539	105	75626	2.140	ppbv	99
63) 1,1,2,2-Tetrachloroethane	9.963	83	34562	2.178	ppbv	98
64) n-propylbenzene	10.989	120	19160	2.226	ppbv	94
65) tert-Butylbenzene	11.532	119	70813	2.246	ppbv	97
66) Benzyl Chloride	11.613	91	6794	1.864	ppbv	97
67) sec-Butylbenzene	11.769	105	101458	2.264	ppbv	99
69) p-Isopropyltoluene	11.924	119	84026	2.226	ppbv	98
70) n-Butylbenzene	12.280	91	89400	2.291	ppbv	97
71) 2-Chlorotoluene	10.901	91	59271	2.139	ppbv	97
72) 4-Ethyltoluene	11.124	105	63833	2.210	ppbv	98
73) 1,3,5-Trimethylbenzene	11.205	105	56516	2.225	ppbv	94
74) 1,2,4-Trimethylbenzene	11.536	105	72485	2.359	ppbv	94
75) 1,3-Dichlorobenzene	11.607	146	36885	2.156	ppbv	97
76) 1,4-Dichlorobenzene	11.671	146	33728	2.083	ppbv	94
77) 1,2-Dichlorobenzene	11.947	146	36078	2.154	ppbv	99
78) Hexachloro-1,3-Butadiene	13.882	225	37360	2.234	ppbv	99
79) Naphthalene	13.513	128	75248	2.920	ppbv	96
80) Naphthalene,2-methyl-	14.458	142	54922	2.857	ppbv	96
81) 1,2,4-Trichlorobenzene	13.442	180	33063	2.396	ppbv	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

