

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL050219\
 Data File : VL033633.D
 Acq On : 2 May 2019 18:24
 Operator : SY/AP
 Sample : K2628-04DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 A-4DL

Manual Integrations
 APPROVED

MMDadoda
 5/4/2019 7:18:41 AM

Quant Time: May 03 04:27:46 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL050119AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu May 03 04:27:46 2019
 QLast Update : Thu May 02 06:29:02 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.76	49	968685	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.29	114	2307039	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.23	117	2572012	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.57	95	2101105	10.73	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	107.30%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
2) Dichlorodifluoromethane	3.07	85	13800	0.095	ppbv	98
13) Ethanol	3.66	45	76064m	5.059	ppbv	
15) Acetone	3.91	43	68286	0.579	ppbv	99
20) Methylene Chloride	4.41	84	10809m	0.235	ppbv	
53) Toluene	9.94	91	101728	0.410	ppbv	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL050219\
Data File : VL033633.D
Acq On : 2 May 2019 18:24
Operator : SY/AP
Sample : K2628-04DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
A-4DL

Manual Integrations
APPROVED

MMDadoda
5/4/2019 7:18:41 AM

Quant Time: May 03 04:27:46 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL050119AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu May
QLast Update : Thu May 02 06:29:02 2019
Response via : Initial Calibration

