

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL050219\
 Data File : VL033641.D
 Acq On : 2 May 2019 23:44
 Operator : SY/AP
 Sample : K2628-08DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 A-8DL

Quant Time: May 03 04:40:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL050119AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu May 02 06:29:02 2019
 QLast Update : Thu May 02 06:29:02 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.76	49	1027765	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.29	114	2275926	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.23	117	2385017	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.57	95	1914654	10.54	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	105.40%

Target Compounds					Ovalue
13) Ethanol	3.65	45	86238	5.406 ppbv	96
15) Acetone	3.91	43	46448	0.371 ppbv	97
53) Toluene	9.95	91	86008	0.351 ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL050219\
Data File : VL033641.D
Acq On : 2 May 2019 23:44
Operator : SY/AP
Sample : K2628-08DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
A-8DL

Quant Time: May 03 04:40:37 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL050119AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu May 02 06:29:02 2019
QLast Update : Thu May 02 06:29:02 2019
Response via : Initial Calibration

