

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL050219\
 Data File : VL033649.D
 Acq On : 3 May 2019 5:01
 Operator : SY/AP
 Sample : K2628-12DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 A-12DL

Manual Integrations
 APPROVED

MMDadoda
 5/4/2019 7:19:05 AM

Quant Time: May 03 08:26:48 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL050119AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu May 03 08:26:48 2019
 QLast Update : Thu May 02 06:29:02 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.76	49	932805	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.29	114	2267741	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.23	117	2333428	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.57	95	1918317	10.80	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	108.00%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Ethanol	3.65	45	66956m	4.624	ppbv	
15) Acetone	3.91	43	67727	0.596	ppbv	97
53) Toluene	9.94	91	73947	0.303	ppbv	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL050219\
Data File : VL033649.D
Acq On : 3 May 2019 5:01
Operator : SY/AP
Sample : K2628-12DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
A-12DL

Manual Integrations
APPROVED

MMDadoda
5/4/2019 7:19:05 AM

Quant Time: May 03 08:26:48 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL050119AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu May
QLast Update : Thu May 02 06:29:02 2019
Response via : Initial Calibration

