

Data Path : W:\HPCHEM1\MSVOA L\DATA\VL050318\
 Data File : VL031972.D
 Acq On : 3 May 2018 21:19
 Operator : SY/AP
 Sample : J2629-04DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 IA2DL

Manual Integrations
 APPROVED

MMDadoda
 5/4/2018 5:01:07 PM

Quant Time: May 04 04:03:54 2018
 Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL050318AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu May 03 2018
 QLast Update : Thu May 03 15:52:13 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.88	49	1229818	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.42	114	2527866	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.39	117	2521891	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.72	95	2011711	10.29	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.90%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Ethanol	3.75	45	83138	5.25	ppbv #	100
15) Acetone	4.01	43	92228	0.53	ppbv	96
19) Isopropyl Alcohol	4.14	45	47793	0.51	ppbv #	89
20) Methylene Chloride	4.51	84	25967m	0.44	ppbv	
26) Hexane	5.91	57	19181m	0.14	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA L\DATA\VL050318\
Data File : VL031972.D
Acq On : 3 May 2018 21:19
Operator : SY/AP
Sample : J2629-04DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
IA2DL

Manual Integrations
APPROVED

MMDadoda
5/4/2018 5:01:07 PM

Quant Time: May 04 04:03:54 2018
Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL050318AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu May
QLast Update : Thu May 03 15:52:13 2018
Response via : Initial Calibration

