

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050525\
 Data File : VL042473.D
 Acq On : 05 May 2025 11:22
 Operator : SY/MD
 Sample : Q1927-04DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 EBR-P03-SC-SS01-20250428DL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/06/2025
 Supervised By :Mahesh Dadoda 05/06/2025

Quant Time: May 06 01:40:40 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Fri May 02 03:16:31 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.790	49	80347	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.965	114	224542	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.885	117	185866	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.377	95	169008	9.808	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.100%
Target Compounds						
					Qvalue	
9) Propene	1.479	41	5215m	1.134	ppbv	
10) Heptane	4.994	43	5436m	0.456	ppbv	
13) Ethanol	1.722	45	8323	8.920	ppbv #	37
15) Acetone	1.839	43	61972	4.225	ppbv	98
26) Hexane	2.832	57	13946	1.363	ppbv #	34
59) m/p-Xylene	9.532	91	3492m	0.134	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050525\
Data File : VL042473.D
Acq On : 05 May 2025 11:22
Operator : SY/MD
Sample : Q1927-04DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
EBR-P03-SC-SS01-20250428DL

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025

Quant Time: May 06 01:40:40 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug
QLast Update : Fri May 02 03:16:31 2025
Response via : Initial Calibration

