

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL050619\
 Data File : VL033672.D
 Acq On : 6 May 2019 23:48
 Operator : SY/AP
 Sample : QC042919 CAN 10607
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC042919 CAN 10607

Quant Time: May 07 05:38:01 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL050619AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Mon May 06 15:19:01 2019
 QLast Update : Mon May 06 15:19:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.75	49	808511	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.28	114	1901253	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.23	117	1894076	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.57	95	1674470	10.78	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	107.80%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL050619\
Data File : VL033672.D
Acq On : 6 May 2019 23:48
Operator : SY/AP
Sample : OC042919 CAN 10607
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC042919 CAN 10607

Quant Time: May 07 05:38:01 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL050619AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon May 06 15:19:01 2019
QLast Update : Mon May 06 15:19:01 2019
Response via : Initial Calibration

