

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL050620\
 Data File : VL034947.D
 Acq On : 6 May 2020 18:16
 Operator : SY/AP
 Sample : QC040320 CAN 10599
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC040320 CAN 10599

Quant Time: May 07 10:43:49 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL050620AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_LWed May 06 14:49:39 2020
 QLast Update : Wed May 06 14:49:39 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.66	49	1125565	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.18	114	2557167	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.12	117	2402663	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.46	95	1984549	10.20	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL050620\
Data File : VL034947.D
Acq On : 6 May 2020 18:16
Operator : SY/AP
Sample : OC040320 CAN 10599
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC040320 CAN 10599

Quant Time: May 07 10:43:49 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL050620AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed May 06 14:49:39 2020
QLast Update : Wed May 06 14:49:39 2020
Response via : Initial Calibration

