

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050624\
 Data File : VL041232.D
 Acq On : 7 May 2024 8:46
 Operator : SY/MD
 Sample : VSTDC0010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDC0010EC

Quant Time: May 07 17:02:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL042924AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Tue May 07 05:19:44 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.201	49	879547	10.000	ppbv	0.01
33) 1,4-Difluorobenzene	6.664	114	2312774	10.000	ppbv	0.01
55) Chlorobenzene-d5	11.474	117	2771624	10.000	ppbv	# 0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.783	95	1718807	8.214	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	82.100%

Target Compounds						Qvalue
2) Dichlorodifluoromethane	2.742	85	1382839	11.623	ppbv	98
3) Chlorodifluoromethane	2.687	51	1354547	10.140	ppbv	98
4) Chloromethane	2.825	50	518018	9.883	ppbv	94
5) Vinyl Chloride	2.931	62	520221	9.879	ppbv	99
6) Bromomethane	3.131	94	186220	9.693	ppbv	98
7) Chloroethane	3.208	64	188585	10.144	ppbv	100
8) Dichlorotetrafluoroethane	2.870	85	1277037	10.370	ppbv	99
9) Propene	2.706	41	499980	9.564	ppbv	99
10) Heptane	7.587	43	1217983	9.616	ppbv	99
11) Trichlorofluoromethane	3.574	101	1307086	10.810	ppbv	98
12) 1,1,2-Trichlorotrifluo...	4.076	101	944956	10.568	ppbv	98
13) Ethanol	3.253	45	110289	9.411	ppbv	99
14) Bromoethene	3.375	108	377316	10.225	ppbv	99
15) Acetone	3.478	43	960351	9.252	ppbv	98
16) 1,3-Butadiene	2.996	39	540182	10.309	ppbv	97
17) tert-Butyl alcohol	3.886	59	969433	8.658	ppbv	98
18) 1,1-Dichloroethene	3.893	96	416547	10.448	ppbv	95
19) Isopropyl Alcohol	3.587	45	690777	10.141	ppbv #	45
20) Methylene Chloride	3.944	84	337586	9.352	ppbv	96
21) Allyl Chloride	4.005	41	669161	9.887	ppbv	97
22) trans-1,2-Dichloroethene	4.455	96	432965	10.443	ppbv	97
23) Vinyl Acetate	4.636	43	1125317	9.845	ppbv	97
24) 1,1-Dichloroethane	4.574	63	921130	10.364	ppbv	99
25) Ethyl Acetate	5.214	43	2009461	9.699	ppbv	100
26) Hexane	5.221	57	880257	9.520	ppbv	99
27) Carbon Disulfide	4.134	76	1066743	9.747	ppbv	97
28) Methyl tert-Butyl Ether	4.597	73	670671	9.110	ppbv	99
29) Chloroform	5.285	83	1363984	10.693	ppbv	99
30) Cyclohexane	6.610	84	742340	10.143	ppbv	98
31) cis-1,2-Dichloroethene	5.089	61	882525	10.196	ppbv	97
32) 1,1,1-Trichloroethane	6.012	97	1328517	10.439	ppbv	99
34) 2-Butanone	4.796	43	1031576	8.763	ppbv	99
35) Carbon Tetrachloride	6.504	117	1348275	10.388	ppbv	99
36) Benzene	6.378	78	1863458	9.601	ppbv	99
37) 1,2-Dichloroethane	5.816	62	1075664	10.622	ppbv	100
38) Trichloroethene	7.301	130	717420	9.919	ppbv	99
39) 1,2-Dichloropropane	7.082	63	728612	9.741	ppbv	96
40) 1,4-Dioxane	7.295	88	282802	10.583	ppbv	90
41) Tetrahydrofuran	5.565	42	781472	9.170	ppbv	99
42) Bromodichloromethane	7.256	83	1523943	10.493	ppbv	100
43) Methyl Methacrylate	7.484	69	756578	9.638	ppbv	99
44) 2,2,4-Trimethylpentane	7.336	57	3149046	9.410	ppbv	100
45) t-1,3-Dichloropropene	8.719	75	533341	9.376	ppbv	99


Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050624\
 Data File : VL041232.D
 Acq On : 7 May 2024 8:46
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDCCC010EC

Quant Time: May 07 17:02:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL042924AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Tue May 07 05:19:44 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	8.150	75	649939	9.333	ppbv	98
47) 1,1,2-Trichloroethane	8.912	97	759245	10.141	ppbv	97
48) Dibromochloromethane	9.719	129	1266098	10.614	ppbv	100
49) Bromoform	12.417	173	1068477	10.529	ppbv	99
50) 4-Methyl-2-Pentanone	8.185	43	1946905	9.216	ppbv	100
51) 2-Hexanone	9.558	43	1417630	9.047	ppbv #	98
52) Tetrachloroethene	10.622	164	667613	9.787	ppbv	99
53) Toluene	9.230	91	2177514	9.676	ppbv	99
54) 1,2-Dibromoethane	10.021	107	1034955	10.206	ppbv	100
56) 1,1,1,2-Tetrachloroethane	11.519	131	926243	8.242	ppbv	98
57) Chlorobenzene	11.536	112	1618121	7.972	ppbv	99
58) Ethyl Benzene	12.095	91	2931204	7.713	ppbv	97
59) m/p-Xylene	12.378	91	2428156	7.602	ppbv #	28
60) o-Xylene	13.069	91	2319989	7.749	ppbv	99
61) Styrene	12.905	104	1258314	7.956	ppbv	99
62) Isopropylbenzene	14.037	105	3470665	7.750	ppbv	99
63) 1,1,2,2-Tetrachloroethane	13.056	83	1635529	8.014	ppbv	100
64) n-propylbenzene	14.921	120	917539	8.212	ppbv	96
65) tert-Butylbenzene	16.098	119	3059465	7.627	ppbv	100
66) Benzyl Chloride	16.333	91	171648	3.780	ppbv	99
67) sec-Butylbenzene	16.641	105	4347341	7.699	ppbv	99
69) p-Isopropyltoluene	17.005	119	3612515	7.763	ppbv	100
70) n-Butylbenzene	17.892	91	3806194	7.832	ppbv	99
71) 2-Chlorotoluene	14.809	91	2669084	7.828	ppbv	98
72) 4-Ethyltoluene	15.198	105	2812688	8.294	ppbv	98
73) 1,3,5-Trimethylbenzene	15.359	105	2445649	7.970	ppbv	99
74) 1,2,4-Trimethylbenzene	16.117	105	2647748	7.901	ppbv	94
75) 1,3-Dichlorobenzene	16.339	146	1574516	8.194	ppbv	100
76) 1,4-Dichlorobenzene	16.481	146	1544144	8.356	ppbv	100
77) 1,2-Dichlorobenzene	17.149	146	1523432	7.931	ppbv	100
78) Hexachloro-1,3-Butadiene	22.638	225	1111923	6.513	ppbv	99
79) Naphthalene	21.394	128	2634247	8.254	ppbv	99
80) Naphthalene,2-methyl-	24.377	142	811231	6.319	ppbv	98
81) 1,2,4-Trichlorobenzene	21.146	180	1269024	7.531	ppbv	100





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Bromochloromethane
Concen: 10.000 ppbv
RT: 5.201 min Scan# 80
Delta R.T. 0.012 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46




Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010EC


Tgt Ion: 49 Resp: 879547
Ion Ratio Lower Upper
49 100
128 44.9 21.5 64.5
130 56.8 27.2 81.6



#2
Dichlorodifluoromethane
Concen: 11.623 ppbv
RT: 2.742 min Scan# 43
Delta R.T. 0.016 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46

Tgt Ion: 85 Resp: 1382839
Ion Ratio Lower Upper
85 100
87 33.0 25.7 38.5









#3
Chlorodifluoromethane
Concen: 10.140 ppbv
RT: 2.687 min Scan# 20
Delta R.T. 0.016 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46




Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010EC


Tgt Ion: 51 Resp: 1354547
Ion Ratio Lower Upper
51 100
67 18.5 14.0 21.0



#4
Chloromethane
Concen: 9.883 ppbv
RT: 2.825 min Scan# 69
Delta R.T. 0.016 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46

Tgt Ion: 50 Resp: 518018
Ion Ratio Lower Upper
50 100
52 31.9 28.3 42.5









#5
 Vinyl Chloride
 Concen: 9.879 ppbv
 RT: 2.931 min Scan# 10
 Delta R.T. 0.013 min
 Lab File: VL041232.D
 Acq: 7 May 2024 8:46




Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDCCC010EC


Tgt Ion: 62 Resp: 520221
 Ion Ratio Lower Upper
 62 100
 64 32.4 25.7 38.5



#6
 Bromomethane
 Concen: 9.693 ppbv
 RT: 3.131 min Scan# 164
 Delta R.T. 0.013 min
 Lab File: VL041232.D
 Acq: 7 May 2024 8:46

Tgt Ion: 94 Resp: 186220
 Ion Ratio Lower Upper
 94 100
 96 89.2 70.1 105.1









#7
Chloroethane
Concen: 10.144 ppbv
RT: 3.208 min Scan# 11
Delta R.T. 0.013 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46




Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010EC


Tgt Ion: 64 Resp: 188585
Ion Ratio Lower Upper
64 100
66 31.8 25.4 38.2



#8
Dichlorotetrafluoroethane
Concen: 10.370 ppbv
RT: 2.870 min Scan# 83
Delta R.T. 0.016 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46

Tgt Ion: 85 Resp: 1277037
Ion Ratio Lower Upper
85 100
135 65.2 51.5 77.3





#9

Propene

Concen: 9.564 ppbv

RT: 2.706 min Scan# 31

Delta R.T. 0.016 min

Lab File: VL041232.D

Acq: 7 May 2024 8:46

Instrument :

MSVOA_L

ClientSampleId :





VSTDCCC010EC

Tgt Ion: 41 Resp: 499980

Ion Ratio Lower Upper

41 100

42 68.2 55.3 82.9



#10

Heptane

Concen: 9.616 ppbv

RT: 7.587 min Scan# 1550

Delta R.T. 0.006 min

Lab File: VL041232.D




Acq: 7 May 2024 8:46


Tgt Ion: 43 Resp: 1217983

Ion Ratio Lower Upper

43 100

57 49.0 38.7 58.1









#11
Trichlorofluoromethane
Concen: 10.810 ppbv
RT: 3.574 min Scan# 302
Delta R.T. 0.016 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46




Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010EC


Tgt Ion:101 Resp: 1307086
Ion Ratio Lower Upper
101 100
103 63.3 51.6 77.4



#12
1,1,2-Trichlorotrifluoroethane
Concen: 10.568 ppbv
RT: 4.076 min Scan# 458
Delta R.T. 0.013 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46

Tgt Ion:101 Resp: 944956
Ion Ratio Lower Upper
101 100
151 69.4 56.8 85.2









#13
Ethanol
Concen: 9.411 ppbv
RT: 3.253 min Scan# 20
Delta R.T. 0.016 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46




Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010EC


Tgt Ion: 45 Resp: 110289
Ion Ratio Lower Upper
45 100
46 37.3 14.8 59.0



#14
Bromoethene
Concen: 10.225 ppbv
RT: 3.375 min Scan# 240
Delta R.T. 0.013 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46

Tgt Ion: 108 Resp: 377316
Ion Ratio Lower Upper
108 100
106 110.0 88.2 132.4
79 23.4 17.4 26.0









#15
Acetone
Concen: 9.252 ppbv
RT: 3.478 min Scan# 21
Delta R.T. 0.016 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46




Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010EC


Tgt Ion: 43 Resp: 960351
Ion Ratio Lower Upper
43 100
58 25.7 21.4 32.0



#16
1,3-Butadiene
Concen: 10.309 ppbv
RT: 2.996 min Scan# 122
Delta R.T. 0.013 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46

Tgt Ion: 39 Resp: 540182
Ion Ratio Lower Upper
39 100
54 75.3 62.2 93.4









#17
tert-Butyl alcohol
Concen: 8.658 ppbv
RT: 3.886 min Scan# 395
Delta R.T. 0.013 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46




Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010EC


Tgt Ion: 59 Resp: 969433
Ion Ratio Lower Upper
59 100
57 11.2 8.4 12.6



#18
1,1-Dichloroethene
Concen: 10.448 ppbv
RT: 3.893 min Scan# 401
Delta R.T. 0.016 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46

Tgt Ion: 96 Resp: 416547
Ion Ratio Lower Upper
96 100
61 198.1 164.4 246.6
98 60.0 51.8 77.8









#19
Isopropyl Alcohol
Concen: 10.141 ppbv
RT: 3.587 min Scan# 306
Delta R.T. 0.016 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46




Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010EC


Tgt Ion: 45 Resp: 690777
Ion Ratio Lower Upper
45 100
58 2.0 26.8 40.2#



#20
Methylene Chloride
Concen: 9.352 ppbv
RT: 3.944 min Scan# 417
Delta R.T. 0.016 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46

Tgt Ion: 84 Resp: 337586
Ion Ratio Lower Upper
84 100
49 167.2 130.4 195.6
51 52.9 38.0 57.0
86 66.8 50.3 75.5









#21
 Allyl Chloride
 Concen: 9.887 ppbv
 RT: 4.005 min Scan# 41
 Delta R.T. 0.013 min
 Lab File: VL041232.D
 Acq: 7 May 2024 8:46




Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDCCC010EC

Tgt Ion: 41 Resp: 669161
 Ion Ratio Lower Upper
 41 100
 76 28.8 23.7 35.5
 39 83.8 64.7 97.1









#22
 trans-1,2-Dichloroethene
 Concen: 10.443 ppbv
 RT: 4.455 min Scan# 576
 Delta R.T. 0.016 min
 Lab File: VL041232.D
 Acq: 7 May 2024 8:46

Tgt Ion: 96 Resp: 432965
 Ion Ratio Lower Upper
 96 100
 61 177.3 137.6 206.4
 98 63.7 50.6 76.0




Abundance Scan 628 (4.623 min): VL041179.D\data.ms (-62



#23
Vinyl Acetate
Concen: 9.845 ppbv
RT: 4.636 min Scan# 613
Delta R.T. 0.013 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46

Tgt Ion: 43 Resp: 1125317


Ion	Ratio	Lower	Upper
43	100		
86	6.0	4.9	7.3
42	11.8	8.0	12.0
44	5.2	3.7	5.5




#24
1,1-Dichloroethane
Concen: 10.364 ppbv
RT: 4.574 min Scan# 613
Delta R.T. 0.016 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46

Tgt Ion: 63 Resp: 921130

Ion	Ratio	Lower	Upper
63	100		
98	3.9	1.6	4.8
100	2.4	1.1	3.5









#25
Ethyl Acetate
Concen: 9.699 ppbv
RT: 5.214 min Scan# 812
Delta R.T. 0.013 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46

Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010EC

Tgt Ion: 43 Resp: 2009461




Ion	Ratio	Lower	Upper
43	100		
45	10.1	8.2	12.2
61	9.7	7.7	11.5
70	7.1	5.8	8.6




#26
Hexane
Concen: 9.520 ppbv
RT: 5.221 min Scan# 814
Delta R.T. 0.013 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46

Tgt Ion: 57 Resp: 880257

Ion	Ratio	Lower	Upper
57	100		
41	96.2	75.8	113.8
43	231.5	187.0	280.4
56	53.0	42.7	64.1









#27
Carbon Disulfide
Concen: 9.747 ppbv
RT: 4.134 min Scan# 41
Delta R.T. 0.013 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46




Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010EC


Tgt Ion: 76 Resp: 1066743
Ion Ratio Lower Upper
76 100
44 21.9 16.3 24.5
78 9.8 7.3 10.9



#28
Methyl tert-Butyl Ether
Concen: 9.110 ppbv
RT: 4.597 min Scan# 620
Delta R.T. 0.016 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46

Tgt Ion: 73 Resp: 670671
Ion Ratio Lower Upper
73 100
57 24.0 18.8 28.2





#29

Chloroform

Concen: 10.693 ppbv

RT: 5.285 min Scan# 81

Delta R.T. 0.013 min

Lab File: VL041232.D

Acq: 7 May 2024 8:46

Instrument :

MSVOA_L

ClientSampleId :





VSTDCCC010EC

Tgt Ion: 83 Resp: 1363984

Ion Ratio Lower Upper

83 100

85 65.3 51.7 77.5



#30

Cyclohexane

Concen: 10.143 ppbv

RT: 6.610 min Scan# 1246

Delta R.T. 0.013 min

Lab File: VL041232.D

Acq: 7 May 2024 8:46




Tgt Ion: 84 Resp: 742340


Ion Ratio Lower Upper

84 100

56 142.6 116.8 175.2

41 92.5 73.8 110.8









#31
cis-1,2-Dichloroethene
Concen: 10.196 ppbv
RT: 5.089 min Scan# 71
Delta R.T. 0.016 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46




Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010EC


Tgt Ion: 61 Resp: 882525
Ion Ratio Lower Upper
61 100
96 63.5 49.8 74.6
98 40.2 30.0 45.0



#32
1,1,1-Trichloroethane
Concen: 10.439 ppbv
RT: 6.012 min Scan# 1060
Delta R.T. 0.010 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46

Tgt Ion: 97 Resp: 1328517
Ion Ratio Lower Upper
97 100
99 64.7 51.7 77.5
61 53.5 41.1 61.7





#33

1,4-Difluorobenzene

Concen: 10.000 ppbv

RT: 6.664 min Scan# 11

Delta R.T. 0.013 min

Lab File: VL041232.D

Acq: 7 May 2024 8:46

Instrument :

MSVOA_L

ClientSampleId :

VSTDCCC010EC





Tgt Ion:114 Resp: 2312774

Ion Ratio Lower Upper

114 100

63 26.0 21.8 32.6

88 18.4 15.2 22.8



#34

2-Butanone

Concen: 8.763 ppbv

RT: 4.796 min Scan# 682

Delta R.T. 0.016 min

Lab File: VL041232.D




Acq: 7 May 2024 8:46


Tgt Ion: 43 Resp: 1031576

Ion Ratio Lower Upper

43 100

72 21.3 17.3 25.9









#35
Carbon Tetrachloride
Concen: 10.388 ppbv
RT: 6.504 min Scan# 1174
Delta R.T. 0.013 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46




Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010EC


Tgt Ion:117 Resp: 1348275
Ion Ratio Lower Upper
117 100
119 93.0 74.3 111.5
121 30.8 23.8 35.8



#36
Benzene
Concen: 9.601 ppbv
RT: 6.378 min Scan# 1174
Delta R.T. 0.010 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46

Tgt Ion: 78 Resp: 1863458
Ion Ratio Lower Upper
78 100
77 23.4 18.3 27.5
50 19.9 15.9 23.9









#37
1,2-Dichloroethane
Concen: 10.622 ppbv
RT: 5.816 min Scan# 996
Delta R.T. 0.010 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46




Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010EC


Tgt Ion: 62 Resp: 1075664
Ion Ratio Lower Upper
62 100
98 6.1 5.0 7.4



#38
Trichloroethene
Concen: 9.919 ppbv
RT: 7.301 min Scan# 1461
Delta R.T. 0.010 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46

Tgt Ion:130 Resp: 717420
Ion Ratio Lower Upper
130 100
95 116.7 95.0 142.6
132 94.5 77.0 115.4
97 73.5 59.1 88.7





#39

1,2-Dichloropropane

Concen: 9.741 ppbv

RT: 7.082 min Scan# 11

Delta R.T. 0.010 min

Lab File: VL041232.D

Acq: 7 May 2024 8:46

Instrument :

MSVOA_L

ClientSampleId :

VSTDCCC010EC





Tgt Ion: 63 Resp: 728612

Ion Ratio Lower Upper

63 100

41 70.4 59.4 89.0

62 67.7 56.8 85.2



#40

1,4-Dioxane

Concen: 10.583 ppbv

RT: 7.295 min Scan# 1459

Delta R.T. 0.010 min

Lab File: VL041232.D

Acq: 7 May 2024 8:46

Tgt Ion: 88 Resp: 282802




Ion Ratio Lower Upper


88 100

58 126.0 106.1 159.1

43 259.1 219.8 329.6

57 1116.7 936.0 1404.0









#41
Tetrahydrofuran
Concen: 9.170 ppbv
RT: 5.565 min Scan# 91
Delta R.T. 0.010 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46




Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010EC


Tgt Ion: 42 Resp: 781472
Ion Ratio Lower Upper
42 100
72 38.1 29.4 44.0
71 36.3 29.1 43.7



#42
Bromodichloromethane
Concen: 10.493 ppbv
RT: 7.256 min Scan# 1447
Delta R.T. 0.010 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46

Tgt Ion: 83 Resp: 1523943
Ion Ratio Lower Upper
83 100
85 63.0 50.6 75.8
127 7.6 6.2 9.4









#43
Methyl Methacrylate
Concen: 9.638 ppbv
RT: 7.484 min Scan# 11
Delta R.T. 0.013 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46




Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010EC


Tgt Ion: 69 Resp: 756578
Ion Ratio Lower Upper
69 100
41 158.5 125.2 187.8
100 33.4 26.7 40.1



#44
2,2,4-Trimethylpentane
Concen: 9.410 ppbv
RT: 7.336 min Scan# 1472
Delta R.T. 0.013 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46

Tgt Ion: 57 Resp: 3149046
Ion Ratio Lower Upper
57 100
41 36.8 29.5 44.3
56 31.4 25.3 37.9









#45
t-1,3-Dichloropropene
Concen: 9.376 ppbv
RT: 8.719 min Scan# 1902
Delta R.T. 0.013 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46




Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010EC


Tgt Ion: 75 Resp: 533341
Ion Ratio Lower Upper
75 100
77 31.2 25.6 38.4



#46
cis-1,3-Dichloropropene
Concen: 9.333 ppbv
RT: 8.150 min Scan# 1725
Delta R.T. 0.010 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46

Tgt Ion: 75 Resp: 649939
Ion Ratio Lower Upper
75 100
39 72.0 59.4 89.2
77 31.8 25.9 38.9









#47
1,1,2-Trichloroethane
Concen: 10.141 ppbv
RT: 8.912 min Scan# 1958
Delta R.T. 0.013 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46

Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010EC

Tgt Ion: 97 Resp: 759245




Ion	Ratio	Lower	Upper
97	100		
83	86.1	71.2	106.8
85	55.6	46.0	69.0
99	60.7	49.7	74.5




#48
Dibromochloromethane
Concen: 10.614 ppbv
RT: 9.719 min Scan# 2213
Delta R.T. 0.006 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46

Tgt Ion: 129 Resp: 1266098

Ion	Ratio	Lower	Upper
129	100		
127	78.1	39.1	117.5









#49
Bromoform
Concen: 10.529 ppbv
RT: 12.417 min Scan# 3049
Delta R.T. 0.010 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46

Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010EC

Tgt Ion:173 Resp: 1068477




Ion	Ratio	Lower	Upper
173	100		
175	48.7	39.8	59.8
171	52.9	41.8	62.8




#50
4-Methyl-2-Pentanone
Concen: 9.216 ppbv
RT: 8.185 min Scan# 1736
Delta R.T. 0.006 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46

Tgt Ion: 43 Resp: 1946905

Ion	Ratio	Lower	Upper
43	100		
58	34.7	27.7	41.5









#51
2-Hexanone
Concen: 9.047 ppbv
RT: 9.558 min Scan# 2160
Delta R.T. 0.010 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46




Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010EC


Tgt Ion: 43 Resp: 1417630
Ion Ratio Lower Upper
43 100
58 47.0 38.4 57.6
98 0.2 0.0 0.0#



#52
Tetrachloroethene
Concen: 9.787 ppbv
RT: 10.622 min Scan# 2494
Delta R.T. 0.003 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46

Tgt Ion:164 Resp: 667613
Ion Ratio Lower Upper
164 100
166 123.5 101.1 151.7
129 102.5 81.7 122.5
131 101.4 80.4 120.6









#53
Toluene
Concen: 9.676 ppbv
RT: 9.230 min Scan# 2058
Delta R.T. 0.010 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46




Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010EC


Tgt Ion: 91 Resp: 2177514
Ion Ratio Lower Upper
91 100
92 60.8 48.2 72.4



#54
1,2-Dibromoethane
Concen: 10.206 ppbv
RT: 10.021 min Scan# 2307
Delta R.T. 0.013 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46

Tgt Ion: 107 Resp: 1034955
Ion Ratio Lower Upper
107 100
109 94.4 75.2 112.8









#55
Chlorobenzene-d5
Concen: 10.000 ppbv
RT: 11.474 min Scan# 21
Delta R.T. 0.010 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46




Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010EC


Tgt Ion:117 Resp: 2771624
Ion Ratio Lower Upper
117 100
82 52.9 53.6 80.4#
119 25.3 27.2 40.8#



#56
1,1,1,2-Tetrachloroethane
Concen: 8.242 ppbv
RT: 11.519 min Scan# 2773
Delta R.T. 0.010 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46

Tgt Ion:131 Resp: 926243
Ion Ratio Lower Upper
131 100
133 95.7 76.3 114.5
119 58.4 49.4 74.2









#57
Chlorobenzene
Concen: 7.972 ppbv
RT: 11.536 min Scan# 21
Delta R.T. 0.010 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46




Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010EC


Tgt Ion:112 Resp: 1618121
Ion Ratio Lower Upper
112 100
77 68.3 54.9 82.3
114 31.1 28.7 35.1



#58
Ethyl Benzene
Concen: 7.713 ppbv
RT: 12.095 min Scan# 2952
Delta R.T. 0.003 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46

Tgt Ion: 91 Resp: 2931204
Ion Ratio Lower Upper
91 100
106 32.2 24.3 36.5









#59
m/p-Xylene
Concen: 7.602 ppbv
RT: 12.378 min Scan# 3039
Delta R.T. 0.003 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46




Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010EC


Tgt Ion: 91 Resp: 2428156
Ion Ratio Lower Upper
91 100
106 91.2 35.7 53.5#



#60
o-Xylene
Concen: 7.749 ppbv
RT: 13.069 min Scan# 3255
Delta R.T. 0.010 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46

Tgt Ion: 91 Resp: 2319989
Ion Ratio Lower Upper
91 100
106 45.7 22.4 67.3









#61
Styrene
Concen: 7.956 ppbv
RT: 12.905 min Scan# 31
Delta R.T. 0.006 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46




Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010EC


Tgt Ion:104 Resp: 1258314
Ion Ratio Lower Upper
104 100
78 52.4 42.6 63.8
103 47.6 38.4 57.6



#62
Isopropylbenzene
Concen: 7.750 ppbv
RT: 14.037 min Scan# 3556
Delta R.T. 0.006 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46

Tgt Ion:105 Resp: 3470665
Ion Ratio Lower Upper
105 100
120 25.7 20.2 30.2









#63
1,1,2,2-Tetrachloroethane
Concen: 8.014 ppbv
RT: 13.056 min Scan# 31
Delta R.T. 0.010 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46




Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010EC


Tgt Ion: 83 Resp: 1635529
Ion Ratio Lower Upper
83 100
131 9.5 4.5 13.5
85 65.3 32.7 98.1



#64
n-propylbenzene
Concen: 8.212 ppbv
RT: 14.921 min Scan# 3831
Delta R.T. 0.003 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46

Tgt Ion: 120 Resp: 917539
Ion Ratio Lower Upper
120 100
91 469.9 384.8 577.2









#65
tert-Butylbenzene
Concen: 7.627 ppbv
RT: 16.098 min Scan# 4195
Delta R.T. 0.006 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46




Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010EC


Tgt Ion:119 Resp: 3059465
Ion Ratio Lower Upper
119 100
91 83.8 67.0 100.4
134 19.4 15.2 22.8



#66
Benzyl Chloride
Concen: 3.780 ppbv
RT: 16.333 min Scan# 4270
Delta R.T. 0.010 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46

Tgt Ion: 91 Resp: 171648
Ion Ratio Lower Upper
91 100
126 18.6 14.6 21.8
128 5.9 4.7 7.1









#67
 sec-Butylbenzene
 Concen: 7.699 ppbv
 RT: 16.641 min Scan# 41
 Delta R.T. 0.003 min
 Lab File: VL041232.D
 Acq: 7 May 2024 8:46




Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDCCC010EC


Tgt Ion:105 Resp: 4347341
 Ion Ratio Lower Upper
 105 100
 134 17.7 13.8 20.8



#68
 1-Bromo-4-Fluorobenzene
 Concen: 8.214 ppbv
 RT: 13.783 min Scan# 3477
 Delta R.T. 0.006 min
 Lab File: VL041232.D
 Acq: 7 May 2024 8:46

Tgt Ion: 95 Resp: 1718807
 Ion Ratio Lower Upper
 95 100
 174 64.2 53.6 80.4
 175 5.0 4.2 6.4









#69
p-Isopropyltoluene
Concen: 7.763 ppbv
RT: 17.005 min Scan# 4477
Delta R.T. 0.006 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46




Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010EC


Tgt Ion:119 Resp: 3612515
Ion Ratio Lower Upper
119 100
134 25.4 20.1 30.1
91 30.2 24.2 36.2



#70
n-Butylbenzene
Concen: 7.832 ppbv
RT: 17.892 min Scan# 4755
Delta R.T. 0.006 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46

Tgt Ion: 91 Resp: 3806194
Ion Ratio Lower Upper
91 100
92 53.3 43.3 64.9
134 22.6 17.8 26.8









#71
2-Chlorotoluene
Concen: 7.828 ppbv
RT: 14.809 min Scan# 31
Delta R.T. 0.006 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46




Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010EC


Tgt Ion: 91 Resp: 2669084
Ion Ratio Lower Upper
91 100
126 31.7 12.2 48.8



#72
4-Ethyltoluene
Concen: 8.294 ppbv
RT: 15.198 min Scan# 3917
Delta R.T. 0.006 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46

Tgt Ion: 105 Resp: 2812688
Ion Ratio Lower Upper
105 100
120 29.5 24.7 37.1
91 13.0 10.3 15.5
77 14.0 11.4 17.0









#73
1,3,5-Trimethylbenzene
Concen: 7.970 ppbv
RT: 15.359 min Scan# 3965
Delta R.T. 0.006 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46




Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010EC


Tgt Ion:105 Resp: 2445649
Ion Ratio Lower Upper
105 100
120 47.0 37.0 55.4



#74
1,2,4-Trimethylbenzene
Concen: 7.901 ppbv
RT: 16.117 min Scan# 4203
Delta R.T. 0.010 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46


Tgt Ion:105 Resp: 2647748
Ion Ratio Lower Upper
105 100
120 46.4 37.8 56.6
91 47.6 44.3 66.5
77 21.8 19.0 28.4








#75
1,3-Dichlorobenzene
Concen: 8.194 ppbv
RT: 16.339 min Scan# 4316
Delta R.T. 0.006 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46

Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010EC







Tgt Ion:146 Resp: 1574516
Ion Ratio Lower Upper
146 100
111 47.3 23.7 71.1
148 65.0 32.5 97.4



#76
1,4-Dichlorobenzene
Concen: 8.356 ppbv
RT: 16.481 min Scan# 4316
Delta R.T. 0.010 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46

Tgt Ion:146 Resp: 1544144
Ion Ratio Lower Upper
146 100
111 46.6 23.3 69.8
148 65.6 32.6 97.7









#77
1,2-Dichlorobenzene
Concen: 7.931 ppbv
RT: 17.149 min Scan# 4522
Delta R.T. 0.006 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46




Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010EC


Tgt Ion:146 Resp: 1523432
Ion Ratio Lower Upper
146 100
111 48.9 24.5 73.5
148 65.1 32.6 98.0



#78
Hexachloro-1,3-Butadiene
Concen: 6.513 ppbv
RT: 22.638 min Scan# 6231
Delta R.T. 0.006 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46

Tgt Ion:225 Resp: 1111923
Ion Ratio Lower Upper
225 100
223 63.4 50.9 76.3
227 63.8 52.0 78.0









#79
Naphthalene
Concen: 8.254 ppbv
RT: 21.394 min Scan# 5844
Delta R.T. 0.010 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46




Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010EC


Tgt Ion:128 Resp: 2634247
Ion Ratio Lower Upper
128 100
127 13.0 10.8 16.2
129 10.6 8.9 13.3



#80
Naphthalene,2-methyl-
Concen: 6.319 ppbv
RT: 24.377 min Scan# 6772
Delta R.T. 0.003 min
Lab File: VL041232.D
Acq: 7 May 2024 8:46

Tgt Ion:142 Resp: 811231
Ion Ratio Lower Upper
142 100
141 83.8 67.9 101.9
115 37.3 31.4 47.0





#81
 1,2,4-Trichlorobenzene
 Concen: 7.531 ppbv
 RT: 21.146 min Scan# 51
 Delta R.T. 0.010 min
 Lab File: VL041232.D
 Acq: 7 May 2024 8:46

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDCCC010EC

Tgt Ion:180 Resp: 1269024

Ion	Ratio	Lower	Upper
180	100		
182	96.2	77.2	115.8
184	30.9	24.6	36.8

