

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050624\
Data File : VL041213.D
Acq On : 6 May 2024 11:39
Operator : SY/MD
Sample : VIBLK
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Quant Time: May 07 05:22:30 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL042924AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Tue May 07 05:19:44 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.192	49	885143	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.648	114	2314600	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.462	117	2093677	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	13.770	95	1590665	10.064	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.600%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050624\
Data File : VL041213.D
Acq On : 6 May 2024 11:39
Operator : SY/MD
Sample : VIBLK
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Quant Time: May 07 05:22:30 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL042924AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Tue May 07 05:19:44 2024
Response via : Initial Calibration

