

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050625\
 Data File : VL042500.D
 Acq On : 06 May 2025 17:04
 Operator : SY/MD
 Sample : MDL3 0.03
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 MDL3 0.03

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/07/2025
 Supervised By :Mahesh Dadoda 05/09/2025

Quant Time: May 07 02:47:14 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Fri May 02 03:16:31 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.790	49	65653	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.965	114	166790	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.891	117	142443	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.387	95	129628	9.815	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.200%
Target Compounds						
					Qvalue	
5) Vinyl Chloride	1.583	62	147m	0.032	ppbv	
32) 1,1,1-Trichloroethane	3.376	97	653m	0.038	ppbv	
35) Carbon Tetrachloride	3.765	117	656m	0.043	ppbv	
38) Trichloroethene	4.567	130	311m	0.052	ppbv	
63) 1,1,2,2-Tetrachloroethane	9.966	83	422m	0.031	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050625\
Data File : VL042500.D
Acq On : 06 May 2025 17:04
Operator : SY/MD
Sample : MDL3 0.03
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
MDL3 0.03

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/07/2025
Supervised By :Mahesh Dadoda 05/09/2025

Quant Time: May 07 02:47:14 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Fri May 02 03:16:31 2025
Response via : Initial Calibration

